

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD
Time of Inspection: 21:35

Date of Inspection: 06/29/2020

Serial Number: 80-006384
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.044 / 0.044			
0.000	0.045 / 0.045			
0.000	0.045 / 0.045			

Number of Simulators Used: 4


Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. .05 OUTSIDE TOL.

- Relocated to new room, verified connections, seals, and hoses on simulators.
- Conducted a new inspection which was in compliance.
- Contacted Brett Kirkland at FDLE ATP and was advised to keep instrument in service since subsequent inspection was in compliance after verifying the issue was not with the instrument.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AARON M STEMEN

Signature and Printed Name

06/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD
Time of Inspection: 23:29

Date of Inspection: 06/29/2020

Serial Number: 80-006384
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24818080A2 Exp: 10/05/2020
0.000	0.050	0.081	0.205	0.078
0.000	0.050	0.081	0.206	0.078
0.000	0.050	0.081	0.206	0.078

Number of Simulators Used: 4

Remarks:
IN COMPLIANCE

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AARON M. STEMEN

Signature and Printed Name

06/29/2020
Date

21:47 06/29/2020

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006384
06/29/2020
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	21:59
Control Test	0.048	22:00
Air Blank	0.000	22:00
Control Test	0.049	22:01
Air Blank	0.000	22:01
Control Test Stats		
Average	0.0485	
Std Dev	0.0007	
Rel Std Dev(%)	1.4579	

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006384
06/29/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Control Test Stats		
Average	0.0795	
Std Dev	0.0007	
Rel Std Dev(%)	0.8894	

GAINESVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006384
06/29/2020
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	22:36
Control Test	0.203	22:36
Air Blank	0.000	22:37
Control Test	0.204	22:38
Air Blank	0.000	22:38
Control Test Stats		
Average	0.2035	
Std Dev	0.0007	
Rel Std Dev(%)	0.3475	



Operator's Signature



Operator's Signature



Operator's Signature