

AMENDED  
03/19/2021

# Florida Department of Law Enforcement Alcohol Testing Program

---

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006632

Time of Inspection: 14:01

Date of Inspection: 10/28/2020

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000	0.044 / 0.045			
0.000	0.044 / 0.044			
0.000	0.044 / 0.045			

Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .  
INSTRUMENT WAS SENT TO FDLE DUE TO INSTRUMENT ISSUE(S). A.H.

The above instrument complies ( ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*ALAIN HERNANDEZ-GINARTE*

ALAIN HERNANDEZ

Signature and Printed Name

10/28/2020

Date

*Amended*  
**Florida Department of Law Enforcement  
 Alcohol Testing Program**

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: FL HIGHWAY PATROL  
 Time of Inspection: 14:46

Date of Inspection: 11/24/2020

Serial Number: 80-006632  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:972301 Exp: 06/15/2021
0.000	0.045	0.074 / 0.075	0.194	0.080
0.000	0.046	0.074 / 0.075	0.196	0.080
0.000	0.046	0.075 / 0.076	0.196	0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

*/ 0.08 Simulator warmed up further on  
 Second Attempt 11/24/2020 (A.H.)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALAIN HERNANDEZ  
 Signature and Printed Name

11/24/2020  
 Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Florida Highway Patrol

Instrument Serial Number: 80-006632

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Alain Hernandez
Date of Inspection: 10/28/2020
Time of Inspection: 14:01:01
Agency Inspection Discrepancy: [ ] Incomplete [ ] Untimely/Not Received [ ] Erroneous Information [ ] Procedural [x] Other (Missing Required Information)
[ ] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[ ] Lot Number [ ] Expiration Date for \_\_\_g/ 210L [ ] Alcohol Reference Solution [ ] Dry Gas Standard is [ ] Incorrect [ ] Expired.
[x] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [ ] REASON for repeating the following test(s); OR the [ ] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[ ] Alcohol Free Subject Test [ ] Mouth Alcohol Test [ ] Alcohol Free Test [ ] Interferent Detect Test
[x] 0.05 g/210L Test [ ] 0.08 g/210L Test [ ] 0.20 g/210L Test [ ] 0.08 g/210L Dry Gas Standard Test
[ ] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[ ] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[ ] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[ ] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[ ] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[ ] Other: \_\_\_\_\_

OTHER ELECTRONIC DATA REVIEW
[ ] Login Records Date: \_\_\_\_\_
[ ] Cylinder Change Records Date: \_\_\_\_\_
[ ] Control Test Records Date: \_\_\_\_\_
[ ] Diagnostic Check Records Date: \_\_\_\_\_
Comments:
The 0.05 g/210 L Test was repeated. The reason the test was repeated, as well as any corrective action taken prior to repeating the test, must be included. Please see below for corrective action.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 03/17/2021 (Date).
[ ] Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).
[ ] Upload the Agency Inspection(s).
[ ] Remove the instrument from evidentiary use until otherwise directed by the Department.
[ ] No action required
[ ] Other: \_\_\_\_\_

Signature of Alcohol Testing Program Staff Member

2/12/2021
Date



Florida Department of Law Enforcement

# AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Florida Highway Patrol

Instrument Serial Number: 80-006632

AGENCY INSPECTION DATA REVIEW		
Agency Inspector: <u>Alain Hernandez</u>	Date of Inspection: <u>05/24/2020</u> <u>06/30/2020</u> <u>07/28/2020</u>	Time of Inspection: <u>16:11:07</u> <u>05:26:07</u> <u>15:56:53</u>
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input checked="" type="checkbox"/> Other ( <u>Required Information Missing</u> )		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for ____g/ 210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input checked="" type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input checked="" type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: <u>During the Agency Inspections dated 05/24/2020, 06/30/2020, and 07/28/2020, the 0.08 g/210 L Test was repeated. The reason for repeating the test, as well as the corrective action taken prior to repeating the test, should be included on each inspection.</u>  <u>See below for corrective action. Please send to the Department Inspector by 09/20/2020.</u>
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION	
<input checked="" type="checkbox"/>	Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>09/20/2020</u> (Date).
<input type="checkbox"/>	Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/>	Upload the Agency Inspection(s).
<input type="checkbox"/>	Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/>	No action required
<input type="checkbox"/>	Other: _____

Signature of Alcohol Testing Program Staff Member

8/20/2020

Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 05:26

Date of Inspection: 06/30/2020

Serial Number: 80-006632  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:972301 Exp: 06/15/2021
0.000	0.046	0.074 / 0.075	0.194	0.081
0.000	0.045	0.075 / 0.075	0.196	0.081
0.000	0.046	0.075 / 0.076	0.197	0.081

Number of Simulators Used: 5


**Remarks:**

08: Control Outside Tolerance.

*Warmed up sim, repeat, pass*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 ALAIN HERNANDEZ  
Signature and Printed Name

06/30/2020  
Date