

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 11:28

Date of Inspection: 11/18/2020

Serial Number: 80-007007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) <i>11/21</i> Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1298240 Exp: 03/19/2023
0.000	0.047	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.077
0.000	0.048	0.079	0.197	0.078

Number of Simulators Used: 5

Remarks:

A FRAZIER PRESENT AND OBSERVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

K. Keating KAYLA C KEATING
Signature and Printed Name

11/18/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT LAUDERDALE PD
Time of Inspection: 12:46

Date of Inspection: 07/02/2020

Serial Number: 80-007007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:1298240 Exp: 03/19/2023	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1298240 Exp: 03/19/2023
0.000	0.048	0.076	0.194	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.078

Number of Simulators Used: 5

Remarks:

A FRAZIER PRESENT AND OBSERVED

AMENDED *ll* 8/13/2020

Lot Number /Expiration for .05g/210L Alcohol Reference Solution Incorrect. Correct Lot Number is 201811C and Expiration is 11/13/2020.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Kayla Keating

KAYLA C KEATING

Signature and Printed Name

07/02/2020
Date