

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 14:41

Date of Inspection: 03/29/2020

Serial Number: 80-007160
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:181632 Exp: 03/05/2021
0.000	0.044 / 0.045	0.073 / 0.075	0.192	0.079
0.000	0.044 / 0.045	0.076 / 0.076	0.193	0.080
0.000	0.044 / 0.045	0.075 / 0.076	0.193	0.079

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

Checked all connections and repeat both test

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W. CARTWRIGHT JR

Signature and Printed Name

03/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 18:59

Date of Inspection: 05/30/2020

Serial Number: 80-007160
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:181632 Exp: 03/05/2021
0.000	0.044 / 0.045	0.075	0.192	0.080
0.000	0.044 / 0.045	0.076	0.194	0.080
0.000	0.045 / 0.045	0.076	0.194	0.080


Number of Simulators Used: 5

Remarks:

Time-Date changed. 05: Control Outside Tolerance. Time was off by 15 mins
Control outside Tolerance 2 test did not reach 0.045

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

GARY W CARTWRIGHT JR

05/30/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 15:02

Date of Inspection: 06/30/2020

Serial Number: 80-007160
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:181632 Exp: 03/05/2021
0.000	0.044 / 0.046	0.075	0.191	0.079
0.000	0.044 / 0.045	0.075	0.192	0.079
0.000	0.045 / 0.045	0.076	0.192	0.079

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. ~~05~~ Checks all Connection and repeated test

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W. CARTWRIGHT JR

Signature and Printed Name

06/30/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 11:47

Date of Inspection: 07/04/2020

Serial Number: 80-007160
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:181632 Exp: 03/05/2021
0.000	0.044 / 0.045	0.074 / 0.076	0.194	0.079
0.000	0.045 / 0.045	0.074 / 0.076	0.195	0.079
0.000	0.045 / 0.045	0.075 / 0.076	0.194	0.079

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance. *Checked all connections and repeated test*

June inspection had to conduct a second inspection due to 5 day period expiring

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W CARTWRIGHT JR

Signature and Printed Name

07/04/2020
Date