



INSTRUMENT PROCESSING SHEET

Agency FDLES/N 80-000224

Florida Department of Law Enforcement

Date In 10-13-2020 DI Completion Date 10-16-2020 Ship P/U H/D CMI EE

Intake Performed By <u>IS</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>IS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>222</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.10.20 09:34:43 -04'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Maintenance Performed By <u>IS</u> <input type="checkbox"/> Battery Replacement <input checked="" type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>IS</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.39</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5090</u>																								
Simulator	Serial #	Lot #/Exp																																							
0.050	MP5088	201905A 05-14-2021																																							
0.080	MP5089	201905B 05-14-2021																																							
0.200	MP5090	201904D 04-30-2021																																							
0.080 DGS	N/A	AG931603 11-12-2021																																							
Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge <u>1015</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03519080A4</td> <td>04-05-2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	03519080A4	04-05-2021	Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1016</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																						
0.000	MP5091	N/A	N/A																																						
0.040	MP5082	20060	02-10-2022																																						
0.100	MP5083	20190	04-06-2022																																						
0.200	MP5084	20160	03-18-2022																																						
0.300	MP5085	20030	01-21-2022																																						
0.080 DGS	N/A	03519080A4	04-05-2021																																						
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
Notes/Suggested Service: _____ _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use 2020.10.19 2020.10.2 <i>Michael D. Haughney</i> 13:10:49 -04'00' <i>SR</i> 09:34:12 Tech Review / Date Admin Review / Date																																								

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 10:44

Date of Inspection: 10/16/2020

Serial Number: 80-000224

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.050	0.078	0.206	0.079
0.000	0.050	0.079	0.206	0.079
0.000	0.050	0.078	0.205	0.079
0.000	0.050	0.079	0.205	0.080
0.000	0.050	0.079	0.206	0.079
0.000	0.051	0.079	0.206	0.080
0.000	0.050	0.079	0.206	0.079
0.000	0.050	0.079	0.205	0.079
0.000	0.051	0.080	0.207	0.080
0.000	0.051	0.079	0.205	0.078

Standard Deviations	0.0004	0.0005	0.0006	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

10/16/2020
Date

MX

2020.1
0.20
09:33:4
9
-04'00'

Stability Checks

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
10/13/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:47
Control Test	0.045	13:48
Air Blank	0.000	13:48
Control Test	0.045	13:49
Air Blank	0.000	13:50
Control Test	0.045	13:50
Air Blank	0.000	13:51
Control Test Stats		
Average	0.0497	
Std Dev	0.0012	
Rel Std Dev(%)	2.3249	

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
10/13/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:52
Control Test	0.077	13:53
Air Blank	0.000	13:53
Control Test	0.076	13:54
Air Blank	0.000	13:54
Control Test	0.076	13:55
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0763	
Std Dev	0.0006	
Rel Std Dev(%)	0.7564	

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
10/13/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:56
Control Test	0.205	13:57
Air Blank	0.000	13:58
Control Test	0.201	13:58
Air Blank	0.000	13:59
Control Test	0.200	14:00
Air Blank	0.000	14:00
Control Test Stats		
Average	0.2020	
Std Dev	0.0026	
Rel Std Dev(%)	1.3098	

Wet

Operator's Signature

Operator's Signature

Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
10/13/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:10
Control Test	0.081	14:11
Air Blank	0.000	14:11
Control Test	0.080	14:12
Air Blank	0.000	14:12
Control Test	0.080	14:12
Air Blank	0.000	14:13
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry

Operator's Signature

2020.10.
20
09:33:28
-04'00"



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000224, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000224</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/16/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>10:44</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date 10/16/2020
Israel Soto
ISRAEL SOTO,
Department Inspector

Optical Bench Cal. Adj.

FDLE
 Analyzer - Alcohol Analyzer
 Model 8000 SN 80-000224
 10/14/2020 09:02:12

Auto Calibration
 Max Power Res Value = 29
 Auto Range Res Value = 17

Sol Value = 0.000 g/210L ***
 Fit Value = 0.000 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum To = 12582, Sum To = 13566

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	-0.0010	(-0.0220)
Sample #2	0.0200	(-0.0270)
Sample #3	0.0230	(-0.0030)
Sample #4	0.0280	(0.0440)
Aug % Abs	0.0237	(0.0047)
STD DEU	0.0040	(0.0361)
REL STD DEU	17.077	(773.905)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.0670	(-0.0140)
Sample #2	0.1030	(-0.0380)
Sample #3	0.0960	(-0.0140)
Sample #4	0.0840	(0.0000)
Aug % Abs	0.0943	(-0.0173)
STD DEU	0.0096	(0.0192)
REL STD DEU	10.186	(110.873)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum To = 12669, Sum To = 13661

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	0.7250	(-0.0450)
Sample #2	0.6850	(-0.0150)
Sample #3	0.6830	(-0.0050)
Sample #4	0.6990	(0.0350)
Aug % Abs	0.6890	(0.0050)
STD DEU	0.0087	(0.0265)
REL STD DEU	1.265	(529.150)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5020	(-0.0220)
Sample #2	1.4470	(0.0280)
Sample #3	1.4610	(0.0140)
Sample #4	1.4270	(0.0360)
Aug % Abs	1.4450	(0.0253)
STD DEU	0.0171	(0.0110)
REL STD DEU	1.183	(43.481)

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum To = 12659, Sum To = 13655

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.6660	(-0.0140)
Sample #2	1.6760	(0.0380)
Sample #3	1.6720	(0.0540)
Sample #4	1.6860	(0.0350)
Aug % Abs	1.6780	(0.0423)
STD DEU	0.0072	(0.0102)
REL STD DEU	0.430	(24.128)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.4810	(0.0040)
Sample #2	3.4430	(0.0790)
Sample #3	3.4620	(0.0670)
Sample #4	3.4120	(0.1120)
Aug % Abs	3.4390	(0.0860)
STD DEU	0.0252	(0.0233)
REL STD DEU	0.734	(27.036)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum To = 12640, Sum To = 13647

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.3180	(-0.0160)
Sample #2	3.2640	(0.0220)
Sample #3	3.2580	(0.0670)
Sample #4	3.2520	(0.0490)
Aug % Abs	3.2580	(0.0460)
STD DEU	0.0068	(0.0226)
REL STD DEU	0.184	(49.238)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	6.7120	(0.0070)
Sample #2	6.5970	(0.1310)
Sample #3	6.5590	(0.1620)
Sample #4	6.5530	(0.1560)
Aug % Abs	6.5697	(0.1497)
STD DEU	0.0239	(0.0164)
REL STD DEU	0.363	(110.986)

Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %
 Samples Taken = 4, Discarded = 1
 Sum To = 12640, Sum To = 13640

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	4.8510	(-0.0260)
Sample #2	4.7580	(0.0750)
Sample #3	4.7420	(0.0990)
Sample #4	4.7380	(0.0600)
Aug % Abs	4.7460	(0.0847)
STD DEU	0.0106	(0.0127)
REL STD DEU	0.223	(14.955)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	9.6640	(0.0000)
Sample #2	9.5220	(0.1620)
Sample #3	9.4910	(0.2000)
Sample #4	9.4670	(0.2050)
Aug % Abs	9.4933	(0.1890)
STD DEU	0.0276	(0.0235)
REL STD DEU	0.290	(12.442)

***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.024
 Std Dev = 0.00 Rel Std Dev = 17.06
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.689
 Std Dev = 0.01 Rel Std Dev = 1.27
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.678
 Std Dev = 0.01 Rel Std Dev = 0.43
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.258
 Std Dev = 0.01 Rel Std Dev = 0.18
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 4.746
 Std Dev = 0.01 Rel Std Dev = 0.22
 Zero Order Coef = -51.77
 First Order Coef = 2779.74
 Second Order Coef = 58.52
 Standard Deviation = 14.873244

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0010 mg/l or 0.000 g/210L
 % Abs = 0.194
 Std Dev = 0.01 Rel Std Dev = 10.19
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.445
 Std Dev = 0.02 Rel Std Dev = 1.18
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.439
 Std Dev = 0.03 Rel Std Dev = 0.73
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.570
 Std Dev = 0.02 Rel Std Dev = 0.36
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.493
 Std Dev = 0.03 Rel Std Dev = 0.29
 Zero Order Coef = -116.03
 First Order Coef = 1362.35
 Second Order Coef = 16.23
 Standard Deviation = 12.703492

Optical Bench Cal. Adj.

Post Stability Checks

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0004
0.100	0.100	0.0001
0.200	0.200	-0.0004
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0004
0.100	0.100	0.0000
0.200	0.200	-0.0002
0.300	0.300	0.0001

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/L %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3403.00
 Sample #2 = 3403.00
 Sample #3 = 3510.00
 Sample #4 = 3358.00
 Average Result = 3423.6667
 STD DEV = 78.0790
 REL STD DEV = 2.281

 ***** CHANNEL 2
 Sample #1 = 3330.00
 Sample #2 = 3340.00
 Sample #3 = 3389.00
 Sample #4 = 3350.00
 Average Result = 3359.5667
 STD DEV = 25.8908
 REL STD DEV = 0.771

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 10.15
 3 um H2O Adjust (mg/l*10,000) = 386
 9 um H2O Adjust (mg/l*10,000) = 450
 **** AUTO CAL PASS

FDLE
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-000224
 10/16/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:31
Control Test	0.049	08:32
Air Blank	0.000	08:32
Control Test	0.049	08:33
Air Blank	0.000	08:34
Control Test	0.048	08:34
Air Blank	0.000	08:35
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

FDLE
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-000224
 10/16/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.079	08:37
Air Blank	0.000	08:37
Control Test	0.077	08:38
Air Blank	0.000	08:39
Control Test	0.077	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0777	
Std Dev	0.0012	
Rel Std Dev(%)	1.4867	

wet



Operator's Signature



Operator's Signature

FDLE
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-000224
 10/16/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:43
Control Test	0.203	08:44
Air Blank	0.000	08:45
Control Test	0.201	08:45
Air Blank	0.000	08:46
Control Test	0.201	08:47
Air Blank	0.000	08:47
Control Test Stats		
Average	0.2017	
Std Dev	0.0012	
Rel Std Dev(%)	0.5726	

FDLE
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-000224
 10/16/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:49
Control Test	0.080	08:50
Air Blank	0.000	08:50
Control Test	0.080	08:51
Air Blank	0.000	08:51
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Dry



Operator's Signature



Operator's Signature

MX

2020.10
 .20
 09:32:1
 1 -04'00'



INSTRUMENT PROCESSING SHEET

Agency FDLES/N 80-000224

Florida Department of Law Enforcement

Date In 1/16/2020DI Completion Date 1/22/20 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JA</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.480</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
Final Release Date <p style="text-align: center;">FDLE</p> <p style="text-align: center;">JAN 27 2020</p> <p style="text-align: center;">Alcohol Testing Program</p>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201905A 05-17-2021</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>SD1013</td> <td>201904D 07-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG016501 06-17-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 05-17-2021	0.080	DR1279	201905B 05-14-2021	0.200	SD1013	201904D 07-30-2021	0.080 DGS	N/A	AG016501 06-17-2021	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.2</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	SD1012	201905A 05-17-2021																																																												
0.080	DR1279	201905B 05-14-2021																																																												
0.200	SD1013	201904D 07-30-2021																																																												
0.080 DGS	N/A	AG016501 06-17-2021																																																												
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	MP5086																																																													
Interferent	MP5087																																																													
0.050	MP5088																																																													
0.080	MP5089																																																													
0.200	MP5090																																																													
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"> <tr> <td style="text-align: center;"><u>DP</u> 1/23/20</td> <td style="text-align: center;"><u>Britt Kirkland</u> 1/27/2020</td> </tr> <tr> <td style="text-align: center;">Tech Review / Date</td> <td style="text-align: center;">Admin Review / Date</td> </tr> </table>		<u>DP</u> 1/23/20	<u>Britt Kirkland</u> 1/27/2020	Tech Review / Date	Admin Review / Date																																																								
<u>DP</u> 1/23/20	<u>Britt Kirkland</u> 1/27/2020																																																													
Tech Review / Date	Admin Review / Date																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000224, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000224</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/22/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

01/22/2020

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

PGM
BSK 1/27/2020

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:13

Date of Inspection: 01/22/2020

Serial Number: 80-000224

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.050	0.082	0.204	0.081
0.000	0.051	0.082	0.205	0.079
0.000	0.050	0.082	0.205	0.080
0.000	0.051	0.082	0.205	0.080
0.000	0.050	0.083	0.204	0.079
0.000	0.050	0.083	0.205	0.079
0.000	0.050	0.082	0.205	0.079
0.000	0.051	0.082	0.205	0.079
0.000	0.051	0.082	0.205	0.080
0.000	0.051	0.083	0.205	0.080
Standard Deviations	0.0005	0.0004	0.0004	0.0006

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Room
BK
1/27/2020

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

01/22/2020
Date

Stability Checks

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
01/22/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.050	10:46
Air Blank	0.000	10:46
Control Test	0.050	10:47
Air Blank	0.000	10:48
Control Test	0.050	10:48
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
01/22/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:52
Control Test	0.080	10:53
Air Blank	0.000	10:53
Control Test	0.080	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet


Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8100 SN: 80-000224
01/22/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.204	10:57
Air Blank	0.000	10:58
Control Test	0.203	10:58
Air Blank	0.000	10:59
Control Test	0.203	11:00
Air Blank	0.000	11:00
Control Test Stats		
Average	0.2033	
Std Dev	0.0006	
Rel Std Dev(%)	0.2839	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 80-000224
01/22/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:05
Control Test	0.082	11:06
Air Blank	0.000	11:06
Control Test	0.082	11:07
Air Blank	0.000	11:07
Control Test Stats		
Average	0.0813	
Std Dev	0.0012	
Rel Std Dev(%)	1.4197	

Dry


Operator's Signature


1314
1/27/2020