




### INSTRUMENT PROCESSING SHEET

Agency Marion County Sheriff's Office SP S/N 80-000813

Florida Department of Law Enforcement

Date In 04/02/2020 DI Completion Date 4/29/20  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ <b>Final Release Date</b> _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>146</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.4</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>																																																											
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>																																																											
0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>																																																											
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																																											
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1013</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>03519080A4</u></td> <td><u>4-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5086</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1008</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	<u>MP5091</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>MP5086</u>	N/A	N/A																																																										
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																																										
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																																										
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																																										
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																																										
0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																																										
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																																										
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																																										
0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>MP5091</u>																																																												
Interferent	<u>MP5087</u>																																																												
0.050	<u>MP5088</u>																																																												
0.080	<u>MP5089</u>																																																												
0.200	<u>MP5090</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> <b>Instrument Complies with Chapter 11D-8, FAC</b> <input type="checkbox"/> <b>Instrument Does Not Comply with Chapter 11D-8, FAC</b> <input checked="" type="checkbox"/> <b>Return to/Place into Evidentiary Use</b> <input type="checkbox"/> <b>Remain Out of Evidentiary Use</b> <input checked="" type="checkbox"/> <b>Conduct an Agency Inspection Before Evidentiary Use</b> <hr/> Tech Review / Date _____ Admin Review / Date _____																																																												



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000813, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000813</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/09/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>10:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Shayla Platt*

04/09/2020

Date

SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO  
Time of Inspection: 10:43

Date of Inspection: 04/09/2020

Serial Number: 80-000813  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079

Standard Deviations	0.0005	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

04/09/2020  
Date

# Stability Checks

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000813  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:27
Control Test	0.047	07:28
Air Blank	0.000	07:28
Control Test	0.046	07:29
Air Blank	0.000	07:29
Control Test	0.047	07:30
Air Blank	0.000	07:30
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000813  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:34
Control Test	0.077	07:35
Air Blank	0.000	07:35
Control Test	0.076	07:36
Air Blank	0.000	07:37
Control Test	0.077	07:37
Air Blank	0.000	07:38
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000813  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:42
Control Test	0.196	07:43
Air Blank	0.000	07:44
Control Test	0.196	07:44
Air Blank	0.000	07:45
Control Test	0.197	07:45
Air Blank	0.000	07:46
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

wet



Operator's Signature



Operator's Signature



Operator's Signature

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000813  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:48
Control Test	0.078	07:48
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:49
Control Test	0.079	07:50
Air Blank	0.000	07:50
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry



Operator's Signature

CHANN 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 1.4700 (-0.0080)
Sample #2 = 1.4940 (-0.0170)
Sample #3 = 1.4940 (-0.0170)
Sample #4 = 1.4630 (-0.0090)
Aug % Abs = 1.4913 (-0.0083)
STD DEV = 0.0064 (0.0150)
REL STD DEV = 0.426 (180.133)

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/08/2020
SN 80-000813
08:17:44

Auto Calibration
Max Power Res Value = 83
Auto Range Res Value = 64

Sol Value = 0.000 g/210L \*\*\*
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12611, Sum Io = 13342

CHANN 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0830 (-0.0120)
Sample #2 = 0.0830 (-0.0280)
Sample #3 = 0.0990 (0.0090)
Sample #4 = 0.0870 (0.0160)
Aug % Abs = 0.0897 (-0.0010)
STD DEV = 0.0083 (0.0236)
REL STD DEV = 9.286 (2354.318)

CHANN 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.0650 (-0.0130)
Sample #2 = 0.0500 (-0.0190)
Sample #3 = 0.0560 (0.0000)
Sample #4 = 0.0510 (0.0030)
Aug % Abs = 0.0523 (-0.0053)
STD DEV = 0.0032 (0.0119)
REL STD DEV = 6.142 (223.694)

Sol Value = 0.040 g/210L \*\*\*
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12598, Sum Io = 13334

CHANN 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8820 (-0.0170)
Sample #2 = 0.8780 (0.0110)
Sample #3 = 0.9200 (0.0000)
Sample #4 = 0.9020 (0.0340)
Aug % Abs = 0.9000 (0.0150)
STD DEV = 0.0211 (0.0173)
REL STD DEV = 2.341 (115.662)

CHANN 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 6.8880 (-0.0120)
Sample #2 = 6.9010 (-0.0130)
Sample #3 = 6.8930 (-0.0150)
Sample #4 = 6.8880 (0.0000)
Aug % Abs = 6.8940 (-0.0093)
STD DEV = 0.0066 (0.0081)
REL STD DEV = 0.095 (87.263)

Sol Value = 0.300 g/210L \*\*\*
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12568, Sum Io = 13317

CHANN 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.5960 (-0.0190)
Sample #2 = 5.5920 (0.0170)
Sample #3 = 5.5740 (0.0480)
Sample #4 = 5.5760 (0.0330)
Aug % Abs = 5.5807 (0.0327)
STD DEV = 0.0099 (0.0155)
REL STD DEV = 0.177 (47.457)

CHANN 2 >>>>

Sample % Abs (% Abs Ref)
Sample #1 = 10.0110 (-0.0090)
Sample #2 = 9.9900 (0.0210)
Sample #3 = 9.9840 (0.0300)
Sample #4 = 9.9840 (0.0200)
Aug % Abs = 9.9860 (0.0237)
STD DEV = 0.0035 (0.0055)
REL STD DEV = 0.035 (23.271)

Sol Value = 0.200 g/210L \*\*\*
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12574, Sum Io = 13320

CHANN 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6550 (-0.0250)
Sample #2 = 3.6680 (-0.0270)
Sample #3 = 3.8450 (-0.0120)
Sample #4 = 3.8430 (-0.0020)
Aug % Abs = 3.6520 (-0.0137)
STD DEV = 0.0139 (0.0126)
REL STD DEV = 0.361 (92.071)

AUTO CAL DATA \*\*\*\*\*

CHANN 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.090
Std Dev = 0.01 Rel Std Dev = 9.29
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.900
Std Dev = 0.02 Rel Std Dev = 2.34
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.982
Std Dev = 0.00 Rel Std Dev = 0.13
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.852
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.581
Std Dev = 0.01 Rel Std Dev = 0.18
Zero Order Coef = -235.61
First Order Coef = 2420.24
Second Order Coef = 32.20
Standard Deviation = 54.466256

CHANN 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.052
Std Dev = 0.00 Rel Std Dev = 6.14
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.490
Std Dev = 0.01 Rel Std Dev = 0.43
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.531
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.894
Std Dev = 0.01 Rel Std Dev = 0.10
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.986
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -62.45
First Order Coef = 1309.00
Second Order Coef = 12.66
Standard Deviation = 31.143244

Table with columns: Act, g/210L, Fit, Residual. Data for Sol Val = 0.0000 mg/l or 0.000 g/210L.

Table with columns: Act, g/210L, Fit, Residual. Data for Sol Val = 0.1905 mg/l or 0.040 g/210L.

INTOXILYZER 8000
Instrument Initialization
09-07 04/08/2020

Cord dislodged while screwing in DGs. Will repeat cal adjustment. SP

CAL ADJUSTMENT #1 SP

MARLON COUNTY SO  
 Intoxilyzer - Alcotest Analyzer  
 Model 8000  
 04/08/2020  
 SN 80-000813  
 09:19:18  
 Auto Calibration  
 Max Power Res Value = 85  
 Auto Range Res Value = 64

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4620 (-0.0120)  
 Sample #2 = 1.4640 (-0.0140)  
 Sample #3 = 1.4900 (-0.0150)  
 Sample #4 = 1.4820 (-0.0030)  
 Avg % Abs = 1.4787 (-0.0087)  
 STD DEV = 0.0133 (0.0101)  
 REL STD DEV = 0.901 (116.723)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.8780 (-0.0040)  
 Sample #2 = 6.9110 (-0.0040)  
 Sample #3 = 6.8950 (-0.0150)  
 Sample #4 = 6.9200 (-0.0060)  
 Avg % Abs = 6.9087 (-0.0083)  
 STD DEV = 0.0127 (0.0059)  
 REL STD DEV = 0.183 (70.314)

Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4752 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12543, Sum Io = 13302

<<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0740 (-0.0190)  
 Sample #2 = 0.0710 (-0.0240)  
 Sample #3 = 0.0650 (0.0220)  
 Sample #4 = 0.0710 (0.0080)  
 Avg % Abs = 0.0690 (0.0020)  
 STD DEV = 0.0035 (0.0236)  
 REL STD DEV = 5.020 (1178.982)

<<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 2.0000 (-0.0020)  
 Sample #2 = 1.9850 (-0.0080)  
 Sample #3 = 1.9800 (0.0030)  
 Sample #4 = 1.9690 (0.0160)  
 Avg % Abs = 1.9840 (0.0037)  
 STD DEV = 0.0145 (0.0120)  
 REL STD DEV = 0.732 (327.651)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5150 (0.0030)  
 Sample #2 = 3.5310 (-0.0190)  
 Sample #3 = 3.5340 (-0.0090)  
 Sample #4 = 3.5390 (-0.0070)  
 Avg % Abs = 3.5347 (-0.0117)  
 STD DEV = 0.0040 (0.0064)  
 REL STD DEV = 0.114 (55.107)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.2000 g/210L \*\*\*  
 Fit value = 0.9524 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12535, Sum Io = 13299

<<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.8720 (-0.0380)  
 Sample #2 = 3.8650 (-0.0260)  
 Sample #3 = 3.8490 (-0.0200)  
 Sample #4 = 3.8470 (0.0140)  
 Avg % Abs = 3.8537 (-0.0107)  
 STD DEV = 0.0099 (0.0216)  
 REL STD DEV = 0.256 (202.234)

Sol Value = 0.040 g/210L \*\*\*  
 Fit value = 0.1905 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12548, Sum Io = 13314

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.8780 (-0.0260)  
 Sample #2 = 0.8770 (-0.0120)  
 Sample #3 = 0.8760 (-0.0060)  
 Sample #4 = 0.8570 (0.0320)  
 Avg % Abs = 0.8700 (0.0047)  
 STD DEV = 0.0113 (0.0239)  
 REL STD DEV = 1.295 (511.301)

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.069  
 Std Dev = 0.00 Rel Std Dev = 5.02  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.099  
 Std Dev = 0.01 Rel Std Dev = 1.30  
 % Abs = 1.984  
 Std Dev = 0.01 Rel Std Dev = 0.73  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.854  
 Std Dev = 0.01 Rel Std Dev = 0.26  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 5.581  
 Std Dev = 0.01 Rel Std Dev = 0.12  
 Zero Order Coef = -167.90  
 First Order Coef = 2385.92  
 Second Order Coef = 36.12  
 Standard Deviation = 37.247425

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 9.9740 (0.0030)  
 Sample #2 = 9.9980 (-0.0110)  
 Sample #3 = 10.0010 (0.0040)  
 Sample #4 = 9.9960 (-0.0040)  
 Avg % Abs = 9.9983 (-0.0037)  
 STD DEV = 0.0025 (0.0075)  
 REL STD DEV = 0.025 (204.697)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 9.9740 (0.0030)  
 Sample #2 = 9.9980 (-0.0110)  
 Sample #3 = 10.0010 (0.0040)  
 Sample #4 = 9.9960 (-0.0040)  
 Avg % Abs = 9.9983 (-0.0037)  
 STD DEV = 0.0025 (0.0075)  
 REL STD DEV = 0.025 (204.697)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 9.9740 (0.0030)  
 Sample #2 = 9.9980 (-0.0110)  
 Sample #3 = 10.0010 (0.0040)  
 Sample #4 = 9.9960 (-0.0040)  
 Avg % Abs = 9.9983 (-0.0037)  
 STD DEV = 0.0025 (0.0075)  
 REL STD DEV = 0.025 (204.697)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 9.9740 (0.0030)  
 Sample #2 = 9.9980 (-0.0110)  
 Sample #3 = 10.0010 (0.0040)  
 Sample #4 = 9.9960 (-0.0040)  
 Avg % Abs = 9.9983 (-0.0037)  
 STD DEV = 0.0025 (0.0075)  
 REL STD DEV = 0.025 (204.697)

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3094.00  
 Sample #2 = 3050.00  
 Sample #3 = 2987.00  
 Sample #4 = 3122.00  
 Average Result = 3053.0000  
 STD DEV = 67.5500  
 REL STD DEV = 2.213  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3094.00  
 Sample #2 = 3050.00  
 Sample #3 = 2987.00  
 Sample #4 = 3122.00  
 Average Result = 3053.0000  
 STD DEV = 67.5500  
 REL STD DEV = 2.213  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3094.00  
 Sample #2 = 3050.00  
 Sample #3 = 2987.00  
 Sample #4 = 3122.00  
 Average Result = 3053.0000  
 STD DEV = 67.5500  
 REL STD DEV = 2.213  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3094.00  
 Sample #2 = 3050.00  
 Sample #3 = 2987.00  
 Sample #4 = 3122.00  
 Average Result = 3053.0000  
 STD DEV = 67.5500  
 REL STD DEV = 2.213  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.000 g/210L \*\*\*  
 Fit value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3094.00  
 Sample #2 = 3050.00  
 Sample #3 = 2987.00  
 Sample #4 = 3122.00  
 Average Result = 3053.0000  
 STD DEV = 67.5500  
 REL STD DEV = 2.213  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3474.00  
 Sample #2 = 3469.00  
 Sample #3 = 3461.00  
 Sample #4 = 3502.00  
 Average Result = 3477.3333  
 STD DEV = 21.7332  
 REL STD DEV = 0.625

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10.000) = 756  
 9 um H2O Adjust (mg/l\*10.000) = 332  
 \*\*\*\* AUTO CAL PASS

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10.000) = 756  
 9 um H2O Adjust (mg/l\*10.000) = 332  
 \*\*\*\* AUTO CAL PASS

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10.000) = 756  
 9 um H2O Adjust (mg/l\*10.000) = 332  
 \*\*\*\* AUTO CAL PASS

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10.000) = 756  
 9 um H2O Adjust (mg/l\*10.000) = 332  
 \*\*\*\* AUTO CAL PASS

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10.000) = 756  
 9 um H2O Adjust (mg/l\*10.000) = 332  
 \*\*\*\* AUTO CAL PASS

CAL ADJUSTMENT #2  
 # 80-000813 SP

# Post Cal Adjust Stability Checks #80-000813

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000813  
04/09/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.049	08:14
Air Blank	0.000	08:15
Control Test	0.049	08:15
Air Blank	0.000	08:16
Control Test	0.049	08:17
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP  
-----  
Operator's Signature

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000813  
04/09/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:02
Control Test	0.080	08:03
Air Blank	0.000	08:03
Control Test	0.079	08:04
Air Blank	0.000	08:05
Control Test	0.079	08:05
Air Blank	0.000	08:06
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP  
-----  
Operator's Signature

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000813  
04/09/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:08
Control Test	0.200	08:09
Air Blank	0.000	08:09
Control Test	0.199	08:10
Air Blank	0.000	08:11
Control Test	0.200	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SP  
-----  
Operator's Signature

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000813  
04/09/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.079	08:21
Air Blank	0.000	08:21
Control Test	0.079	08:22
Air Blank	0.000	08:22
Control Test	0.080	08:22
Air Blank	0.000	08:23
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

765

SP  
-----  
Operator's Signature