

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Ashley Geoghegan on 10-23-2020

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-000815

| | |
|--|--|
| Bill To Address: <u>Sumter County Sheriff Office</u> _____ _____ _____ _____ | Ship to Address: <u>FDLE</u> _____ _____ _____ _____ |
|--|--|

Reason for Return:
Instrument giving low and variable readings; after two optical bench calibration adjusts
instrument still does not give nominal values.

Contact Ashley Geoghegan with estimate:
ageoghegan@sumtercountysheriff.org

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$ _____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

 Please contact: Name: Ashley Geoghegan
 Phone #: _____ Email: ageoghegan@sumtercountysheriff.c

ATP Contact Name: Israel Soto ATP Email: israelsoto@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Sumter County Sheriff's Office

S/N 80-000815

Florida Department of Law Enforcement

Date In 10/06/2020

DI Completion Date

Ship P/U H/D CMI EE

Intake Performed By RAW, Quality Checks Performed By IS, Flow Calibration Performed By, Maintenance Performed By, Temperature Checks Performed By IS. Includes tables for flow column and simulator data.

Calibration Adjustment Performed By IS, Barometric Pressure Gauge 1015 ID # 26932. Includes tables for simulator and post calibration adjustment stability checks.

Department Inspection Performed By IS, Barometric Pressure ID# 30793, Gauge 1016, Instrument 1016. Includes table for simulator and serial number.

Notes/Suggested Service: Second Opt Cal Adjust Serial/Lot # and expiration dates the same as first Opt Cal Adjust. Second Cal Adjust Baro Pressure: 1019. Could not bring stability checks after second cal adjust to nominal values, sending to repair.

Attachments: Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, Other. Compliance checkboxes for Chapter 11D-8, FAC.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 10:58

Date of Inspection: 10/16/2020

Serial Number: 80-000815
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021 | 0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021 | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|---|---|---|--|
| 0.000 | 0.048 | 0.075 / 0.076 | | |
| 0.000 | 0.048 | 0.074 / 0.074 | | |
| 0.000 | 0.049 | 0.074 / 0.074 | | |
| 0.000 | 0.048 | 0.075 / 0.073 | | |
| 0.000 | 0.048 | 0.073 / 0.073 | | |
| 0.000 | 0.048 | 0.075 / 0.075 | | |
| 0.000 | 0.048 | 0.075 / 0.075 | | |
| 0.000 | 0.048 | 0.075 / 0.074 | | |
| 0.000 | 0.048 | 0.075 / 0.075 | | |
| 0.000 | 0.048 | 0.075 / 0.075 | | |
| 0.000 | 0.049 | 0.075 / 0.077 | | |

| | | | | |
|---------------------|--------|-----------------|--|--|
| Standard Deviations | 0.0004 | 0.0006 / 0.0012 | | |
|---------------------|--------|-----------------|--|--|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance CHECKED SEALS TEMP ORINGS, C . Non-compliance:WILL PERFORM OPT CAL ADJ UST REPEAT INSP.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

10/16/2020
Date

Stability Checks

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/07/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:34 |
| Control Test | 0.048 | 09:35 |
| Air Blank | 0.000 | 09:35 |
| Control Test | 0.048 | 09:36 |
| Air Blank | 0.000 | 09:37 |
| Control Test | 0.045 | 09:37 |
| Air Blank | 0.000 | 09:38 |
| Control Test Stats | | |
| Average | 0.0470 | |
| Std Dev | 0.0017 | |
| Rel Std Dev(%) | 3.6852 | |



Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/07/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:45 |
| Control Test | 0.078 | 09:45 |
| Air Blank | 0.000 | 09:46 |
| Control Test | 0.078 | 09:47 |
| Air Blank | 0.000 | 09:47 |
| Control Test | 0.078 | 09:48 |
| Air Blank | 0.000 | 09:48 |
| Control Test Stats | | |
| Average | 0.0780 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

wet



Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/07/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:54 |
| Control Test | 0.196 | 09:55 |
| Air Blank | 0.000 | 09:55 |
| Control Test | 0.195 | 09:56 |
| Air Blank | 0.000 | 09:56 |
| Control Test | 0.196 | 09:57 |
| Air Blank | 0.000 | 09:58 |
| Control Test Stats | | |
| Average | 0.1957 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2951 | |



Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/07/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:00 |
| Control Test | 0.080 | 10:01 |
| Air Blank | 0.000 | 10:01 |
| Control Test | 0.080 | 10:02 |
| Air Blank | 0.000 | 10:02 |
| Control Test | 0.080 | 10:02 |
| Air Blank | 0.000 | 10:03 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Dry



Operator's Signature

Optical Bench Cal. Adj.

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/14/2020 09:01:12

Auto Calibration
 Max Power Res Value = 22
 Auto Range Res Value = 8

Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12682, Sum Io = 13085

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.0830 | (0.0030) |
| Sample #2 = | 0.0860 | (0.0550) |
| Sample #3 = | 0.0710 | (0.1280) |
| Sample #4 = | 0.0760 | (0.1260) |
| Avg % Abs = | 0.0777 | (0.1030) |
| STD DEU = | 0.0076 | (0.0416) |
| REL STD DEU = | 9.834 | (40.370) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.1430 | (-0.0060) |
| Sample #2 = | 0.1230 | (-0.0010) |
| Sample #3 = | 0.1300 | (0.0220) |
| Sample #4 = | 0.1110 | (0.0230) |
| Avg % Abs = | 0.1213 | (0.0147) |
| STD DEU = | 0.0096 | (0.0136) |
| REL STD DEU = | 7.920 | (92.570) |

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12657, Sum Io = 13077

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.7990 | (-0.0250) |
| Sample #2 = | 0.7770 | (0.0300) |
| Sample #3 = | 0.7660 | (0.0630) |
| Sample #4 = | 0.7720 | (0.0730) |
| Avg % Abs = | 0.7717 | (0.0553) |
| STD DEU = | 0.0055 | (0.0225) |
| REL STD DEU = | 0.714 | (40.666) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 1.4650 | (0.0030) |
| Sample #2 = | 1.4460 | (0.0240) |
| Sample #3 = | 1.4690 | (0.0220) |
| Sample #4 = | 1.4520 | (0.0280) |
| Avg % Abs = | 1.4557 | (0.0247) |
| STD DEU = | 0.0119 | (0.0031) |
| REL STD DEU = | 0.820 | (12.385) |

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12642, Sum Io = 13069

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 1.8280 | (-0.0040) |
| Sample #2 = | 1.8430 | (0.0000) |
| Sample #3 = | 1.7960 | (0.0170) |
| Sample #4 = | 1.8230 | (0.0510) |
| Avg % Abs = | 1.8207 | (0.0227) |
| STD DEU = | 0.0236 | (0.0260) |
| REL STD DEU = | 1.296 | (114.564) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 3.4330 | (-0.0090) |
| Sample #2 = | 3.4150 | (0.0050) |
| Sample #3 = | 3.4220 | (-0.0020) |
| Sample #4 = | 3.4100 | (0.0190) |
| Avg % Abs = | 3.4157 | (0.0073) |
| STD DEU = | 0.0060 | (0.0107) |
| REL STD DEU = | 0.176 | (145.809) |

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12636, Sum Io = 13067

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 3.5570 | (-0.0120) |
| Sample #2 = | 3.5240 | (0.0530) |
| Sample #3 = | 3.5210 | (0.0420) |
| Sample #4 = | 3.5270 | (0.0780) |
| Avg % Abs = | 3.5240 | (0.0577) |
| STD DEU = | 0.0030 | (0.0184) |
| REL STD DEU = | 0.085 | (31.991) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 6.5810 | (-0.0130) |
| Sample #2 = | 6.5800 | (0.0240) |
| Sample #3 = | 6.5700 | (0.0190) |
| Sample #4 = | 6.5650 | (0.0330) |
| Avg % Abs = | 6.5717 | (0.0253) |
| STD DEU = | 0.0076 | (0.0071) |
| REL STD DEU = | 0.116 | (28.005) |

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12627, Sum Io = 13062

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 5.1710 | (-0.0160) |
| Sample #2 = | 5.1350 | (0.0040) |
| Sample #3 = | 5.1600 | (0.0190) |
| Sample #4 = | 5.1520 | (0.0350) |
| Avg % Abs = | 5.1490 | (0.0193) |
| STD DEU = | 0.0128 | (0.0155) |
| REL STD DEU = | 0.248 | (80.186) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 9.5300 | (-0.0270) |
| Sample #2 = | 9.4920 | (0.0010) |
| Sample #3 = | 9.4440 | (0.0410) |
| Sample #4 = | 9.4790 | (0.0140) |
| Avg % Abs = | 9.4717 | (0.0187) |
| STD DEU = | 0.0248 | (0.0204) |
| REL STD DEU = | 0.262 | (109.308) |

***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.078
 Std Dev = 0.01 Rel Std Dev = 9.83
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.772
 Std Dev = 0.01 Rel Std Dev = 0.71
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.821
 Std Dev = 0.02 Rel Std Dev = 1.30
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.524
 Std Dev = 0.00 Rel Std Dev = 0.09
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.149
 Std Dev = 0.01 Rel Std Dev = 0.25
 Zero Order Coef = -188.03
 First Order Coef = 2660.63
 Second Order Coef = 28.86
 Standard Deviation = 19.713961

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.121
 Std Dev = 0.01 Rel Std Dev = 7.92
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.456
 Std Dev = 0.01 Rel Std Dev = 0.82
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.416
 Std Dev = 0.01 Rel Std Dev = 0.18
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.572
 Std Dev = 0.01 Rel Std Dev = 0.12
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.472
 Std Dev = 0.02 Rel Std Dev = 0.26
 Zero Order Coef = -147.47
 First Order Coef = 1376.48
 Second Order Coef = 15.39
 Standard Deviation = 28.266783

Optical Bench Cal. Adj.

Post stability checks

```
-----
: Solution Stats Quadratic Fit Chan 1
: Act      Fit      Residual
: g/210L   g/210L   g/210L
: 0.000    0.000    -0.0004
: 0.040    0.040    0.0005
: 0.100    0.100    0.0002
: 0.200    0.200    -0.0005
: 0.300    0.300    0.0002
-----
```

```
-----
: Solution Stats Quadratic Fit Chan 2
: Act      Fit      Residual
: g/210L   g/210L   g/210L
: 0.000    0.000    -0.0004
: 0.040    0.040    0.0003
: 0.100    0.099    0.0005
: 0.200    0.201    -0.0008
: 0.300    0.300    0.0003
-----
```

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3239.00
 Sample #2 = 3219.00
 Sample #3 = 3371.00
 Sample #4 = 3206.00
 Average Result = 3265.3333
 STD DEV = 91.7406
 REL STD DEV = 2.810

***** CHANNEL 2

Sample #1 = 3307.00
 Sample #2 = 3306.00
 Sample #3 = 3371.00
 Sample #4 = 3293.00
 Average Result = 3323.3333
 STD DEV = 41.7892
 REL STD DEV = 1.257

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1016
 3 um H2O Adjust (mg/l*10,000) = 544
 9 um H2O Adjust (mg/l*10,000) = 486
 **** AUTO CAL PASS

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/16/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:35 |
| Control Test | 0.049 | 08:36 |
| Air Blank | 0.000 | 08:37 |
| Control Test | 0.048 | 08:37 |
| Air Blank | 0.000 | 08:38 |
| Control Test | 0.047 | 08:39 |
| Air Blank | 0.000 | 08:39 |
| Control Test Stats | | |
| Average | 0.0480 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 2.0833 | |



Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/16/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:47 |
| Control Test | 0.199 | 08:48 |
| Air Blank | 0.000 | 08:49 |
| Control Test | 0.198 | 08:49 |
| Air Blank | 0.000 | 08:50 |
| Control Test | 0.199 | 08:51 |
| Air Blank | 0.000 | 08:51 |
| Control Test Stats | | |
| Average | 0.1987 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2906 | |




Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/16/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:43 |
| Control Test | 0.078 | 08:43 |
| Air Blank | 0.000 | 08:44 |
| Control Test | 0.077 | 08:45 |
| Air Blank | 0.000 | 08:45 |
| Control Test | 0.078 | 08:46 |
| Air Blank | 0.000 | 08:46 |
| Control Test Stats | | |
| Average | 0.0777 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7434 | |

wet



Operator's Signature

SUMTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/16/2020
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 08:52 |
| Control Test | 0.080 | 08:52 |
| Air Blank | 0.000 | 08:53 |
| Control Test | 0.079 | 08:53 |
| Air Blank | 0.000 | 08:54 |
| Control Test | 0.081 | 08:54 |
| Air Blank | 0.000 | 08:54 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2500 | |

Dry



Operator's Signature

second Opt. Bench Cal Adj

SUNTER COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000815
 10/19/2020 09:04:24

Auto Calibration
 Max Power Res Value = 22
 Auto Range Res Value = 8

Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12648, Sum Io = 13073

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.0820 | (-0.0090) |
| Sample #2 = | 0.0700 | (0.0480) |
| Sample #3 = | 0.0470 | (0.0630) |
| Sample #4 = | 0.0860 | (0.0950) |
| Aug % Abs = | 0.0577 | (0.0687) |
| STD DEV = | 0.0196 | (0.0240) |
| REL STD DEV = | 28.972 | (34.962) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.1000 | (0.0090) |
| Sample #2 = | 0.1240 | (0.0280) |
| Sample #3 = | 0.1260 | (-0.0040) |
| Sample #4 = | 0.1430 | (0.0220) |
| Aug % Abs = | 0.1310 | (0.0153) |
| STD DEV = | 0.0104 | (0.0170) |
| REL STD DEV = | 7.970 | (110.933) |

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12628, Sum Io = 13063

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 0.7650 | (-0.0190) |
| Sample #2 = | 0.7770 | (0.0060) |
| Sample #3 = | 0.7630 | (0.0200) |
| Sample #4 = | 0.7710 | (0.0440) |
| Aug % Abs = | 0.7703 | (0.0233) |
| STD DEV = | 0.0070 | (0.0192) |
| REL STD DEV = | 0.912 | (82.363) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 1.4640 | (-0.0280) |
| Sample #2 = | 1.4740 | (-0.0310) |
| Sample #3 = | 1.4550 | (-0.0220) |
| Sample #4 = | 1.4640 | (-0.0280) |
| Aug % Abs = | 1.4643 | (-0.0270) |
| STD DEV = | 0.0095 | (0.0046) |
| REL STD DEV = | 0.649 | (16.973) |

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12613, Sum Io = 13059

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 1.8490 | (-0.0340) |
| Sample #2 = | 1.8170 | (-0.0020) |
| Sample #3 = | 1.8070 | (0.0260) |
| Sample #4 = | 1.8400 | (0.0150) |
| Aug % Abs = | 1.8213 | (0.0130) |
| STD DEV = | 0.0169 | (0.0141) |
| REL STD DEV = | 0.929 | (108.513) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 3.4680 | (-0.0310) |
| Sample #2 = | 3.4110 | (0.0040) |
| Sample #3 = | 3.4410 | (-0.0060) |
| Sample #4 = | 3.4490 | (-0.0070) |
| Aug % Abs = | 3.4337 | (-0.0030) |
| STD DEV = | 0.0200 | (0.0061) |
| REL STD DEV = | 0.583 | (202.759) |

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12609, Sum Io = 13059

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 3.5510 | (-0.0220) |
| Sample #2 = | 3.5440 | (0.0400) |
| Sample #3 = | 3.5390 | (0.0500) |
| Sample #4 = | 3.5460 | (0.0540) |
| Aug % Abs = | 3.5430 | (0.0480) |
| STD DEV = | 0.0036 | (0.0072) |
| REL STD DEV = | 0.102 | (15.023) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 6.5790 | (-0.0150) |
| Sample #2 = | 6.5730 | (0.0430) |
| Sample #3 = | 6.5610 | (0.0470) |
| Sample #4 = | 6.5700 | (0.0380) |
| Aug % Abs = | 6.5680 | (0.0427) |
| STD DEV = | 0.0062 | (0.0045) |
| REL STD DEV = | 0.095 | (10.569) |

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12603, Sum Io = 13053

<<<< CHANNEL 1 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 5.1680 | (-0.0190) |
| Sample #2 = | 5.1770 | (0.0200) |
| Sample #3 = | 5.1370 | (0.0300) |
| Sample #4 = | 5.1300 | (0.0350) |
| Aug % Abs = | 5.1480 | (0.0283) |
| STD DEV = | 0.0254 | (0.0076) |
| REL STD DEV = | 0.493 | (26.956) |

<<<< CHANNEL 2 >>>>

| Sample | % Abs | (% Abs Ref) |
|---------------|--------|-------------|
| Sample #1 = | 9.5220 | (-0.0230) |
| Sample #2 = | 9.4920 | (0.0230) |
| Sample #3 = | 9.4660 | (0.0200) |
| Sample #4 = | 9.4690 | (0.0070) |
| Aug % Abs = | 9.4757 | (0.0167) |
| STD DEV = | 0.0142 | (0.0085) |
| REL STD DEV = | 0.150 | (51.029) |

***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.068
 Std Dev = 0.02 Rel Std Dev = 28.97
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.770
 Std Dev = 0.01 Rel Std Dev = 0.91
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.821
 Std Dev = 0.02 Rel Std Dev = 0.93
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.543
 Std Dev = 0.00 Rel Std Dev = 0.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.148
 Std Dev = 0.03 Rel Std Dev = 0.49
 Zero Order Coef = -152.66
 First Order Coef = 2615.67
 Second Order Coef = 36.08
 Standard Deviation = 32.370148

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.131
 Std Dev = 0.01 Rel Std Dev = 7.97
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.464
 Std Dev = 0.01 Rel Std Dev = 0.65
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.434
 Std Dev = 0.02 Rel Std Dev = 0.58
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.568
 Std Dev = 0.01 Rel Std Dev = 0.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.476
 Std Dev = 0.01 Rel Std Dev = 0.15
 Zero Order Coef = -163.65
 First Order Coef = 1377.25
 Second Order Coef = 15.46
 Standard Deviation = 19.941599

Second Opt. Bench
Cal. Adj.

Second Post stability Checks

Solution Stats Quadratic Fit Chan 1

| Act | Fit | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L |
| 0.000 | 0.001 | -0.0005 |
| 0.040 | 0.040 | 0.0004 |
| 0.100 | 0.099 | 0.0006 |
| 0.200 | 0.201 | -0.0009 |
| 0.300 | 0.300 | 0.0004 |

Solution Stats Quadratic Fit Chan 2

| Act | Fit | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L |
| 0.000 | 0.000 | -0.0004 |
| 0.040 | 0.040 | 0.0004 |
| 0.100 | 0.100 | 0.0003 |
| 0.200 | 0.201 | -0.0005 |
| 0.300 | 0.300 | 0.0002 |

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1


**** CHANNEL 1
Sample #1 = 3263.00
Sample #2 = 3390.00
Sample #3 = 3333.00
Sample #4 = 3362.00
Average Result = 3361.6667
STD DEV = 28.5015
REL STD DEV = 0.848

**** CHANNEL 2
Sample #1 = 3316.00
Sample #2 = 3366.00
Sample #3 = 3358.00
Sample #4 = 3362.00
Average Result = 3362.0000
STD DEV = 4.0000
REL STD DEV = 0.119

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l*10,000) = 448
9 um H2O Adjust (mg/l*10,000) = 447
**** AUTO CAL PASS

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
10/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:59 |
| Control Test | 0.046 | 12:00 |
| Air Blank | 0.000 | 12:00 |
| Control Test | 0.047 | 12:01 |
| Air Blank | 0.000 | 12:02 |
| Control Test | 0.047 | 12:02 |
| Air Blank | 0.000 | 12:03 |
| Control Test Stats | | |
| Average | 0.0467 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 1.2372 | |


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
10/21/2020
Software: 8100.27


| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:19 |
| Control Test | 0.199 | 12:20 |
| Air Blank | 0.000 | 12:20 |
| Control Test | 0.199 | 12:21 |
| Air Blank | 0.000 | 12:22 |
| Control Test | 0.199 | 12:22 |
| Air Blank | 0.000 | 12:23 |
| Control Test Stats | | |
| Average | 0.1990 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
10/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:14 |
| Control Test | 0.078 | 12:14 |
| Air Blank | 0.000 | 12:15 |
| Control Test | 0.078 | 12:16 |
| Air Blank | 0.000 | 12:16 |
| Control Test | 0.074 | 12:17 |
| Air Blank | 0.000 | 12:17 |
| Control Test Stats | | |
| Average | 0.0767 | |
| Std Dev | 0.0023 | |
| Rel Std Dev(%) | 3.0123 | |

wet


Operator's Signature

SUMTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000815
10/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 12:25 |
| Control Test | 0.079 | 12:25 |
| Air Blank | 0.000 | 12:26 |
| Control Test | 0.079 | 12:26 |
| Air Blank | 0.000 | 12:26 |
| Control Test | 0.079 | 12:27 |
| Air Blank | 0.000 | 12:27 |
| Control Test Stats | | |
| Average | 0.0790 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DRY


Operator's Signature