



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-000903

Florida Department of Law Enforcement

Date In 03/16/2020 DI Completion Date 03/24/2020 Ship P/U H/D CMI EE

Intake Performed By <u>MAX</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ Final Release Date <div style="text-align: center; padding: 10px;"> FDLE APR 09 2020 Alcohol Testing Program </div>	Quality Checks Performed By <u>MAX</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>132</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.175</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.261</u> (.228 - .278) 103 mm <u>0.539</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD1017</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11/12/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	201905B 05/14/2021	0.200	SD1017	201904D 04/30/2021	0.080 DGS	N/A	AG931603 11/12/2021	Flow Calibration Performed By <u>MAX</u> Flow Column # <u>ATP 104</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>133</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) Maintenance Performed By <u>MAX</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Replaced Keyboard</u> Temperature Checks Performed By <u>MAX</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.66</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1017</u>																																													
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080		201905B 05/14/2021																																																											
0.200	SD1017	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG931603 11/12/2021																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MAX</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>SD1017</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	SD1017
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	SD1017																																																												
Notes/Suggested Service: <u>E-mailed</u> <u>Tech Review: Corrected DGS Exp Date</u> <div style="text-align: center; padding: 10px;"> <input checked="" type="checkbox"/> APPROVED <u>3/26/2020</u> </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: center; padding: 10px;"> <u>SP 4/3/20</u> <u>Bratt Kirkland 4/7/2020</u> Tech Review / Date Admin Review / Date </div>																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC SOUTH REGION
Time of Inspection: 13:04

Date of Inspection: 03/24/2020

Serial Number: 80-000903
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.197	0.079
Standard Deviations	0.0005	0.0005	0.0009	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

SP
TBK
4/7/2020

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Haughey

MICHAEL D HAUGHEY

Signature and Printed Name

03/24/2020
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-00903	FFWCC	03/24/2020	MX

0.05g/210L
0.047 to 0.053

0.08g/210L
0.077 to 0.083

0.20g/210L
0.194 to 0.206

DGS 0.08g/210L
0.077 to 0.083

FFWCC SOUTH REGION
Intoxilizer - Alcohol Analyzer
Model 8000
03/24/2020
SN 80-00903
Software: 8100.27

FFWCC SOUTH REGION
Intoxilizer - Alcohol Analyzer
Model 8000
03/24/2020
SN 80-00903
Software: 8100.27

FFWCC SOUTH REGION
Intoxilizer - Alcohol Analyzer
Model 8000
03/24/2020
SN 80-00903
Software: 8100.27

FFWCC SOUTH REGION
Intoxilizer - Alcohol Analyzer
Model 8000
03/24/2020
SN 80-00903
Software: 8100.27

DGS

Test 9/21/0L Time

Air Blank 0.000 09:26
Control Test 0.048 09:26
Air Blank 0.000 09:27
Control Test 0.048 09:27
Air Blank 0.000 09:28
Control Test 0.048 09:28
Air Blank 0.000 09:29
Control Test 0.000 09:29

Average 0.0480
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test 9/21/0L Time

Air Blank 0.000 09:32
Control Test 0.079 09:32
Air Blank 0.000 09:33
Control Test 0.079 09:33
Air Blank 0.000 09:34
Control Test 0.079 09:34
Air Blank 0.000 09:35
Control Test 0.078 09:35
Air Blank 0.000 09:35
Control Test 0.000 09:35

Average 0.0787
Std Dev 0.0006
Rel Std Dev(%) 0.7339

Test 9/21/0L Time

Air Blank 0.000 09:38
Control Test 0.197 09:38
Air Blank 0.000 09:39
Control Test 0.197 09:39
Air Blank 0.000 09:40
Control Test 0.197 09:40
Air Blank 0.000 09:41
Control Test 0.197 09:41
Air Blank 0.000 09:41
Control Test 0.000 09:42

Average 0.1970
Std Dev 0.0000
Rel Std Dev(%) 0.0000

Test 9/21/0L Time

Air Blank 0.006 09:44
Control Test 0.078 09:44
Air Blank 0.000 09:45
Control Test 0.079 09:45
Air Blank 0.000 09:45
Control Test 0.079 09:45
Air Blank 0.000 09:46
Control Test 0.079 09:46
Air Blank 0.000 09:46
Control Test 0.000 09:47

Average 0.0787
Std Dev 0.0006
Rel Std Dev(%) 0.7339


Operator's Signature *MX*

Operator's Signature *MX*

Operator's Signature *MX*

Operator's Signature *MX*

SP
BLK
4/9/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000903, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000903</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FWCC SOUTH REGION</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/24/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/24/2020 Date *Michael D Haughey*
MICHAEL D HAUGHEY,
Department Inspector

SB BK 4/4/2020

Flow Calibration

80-000903

03/24/2020

MM

FWCC SOUTH REGION
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN: 88-000903
03/24/2020
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SORT(Diff) = 6.797
2: Rate (Liters/min) = 15
SORT(Diff) = 11.090
3: Rate (Liters/min) = 30
SORT(Diff) = 19.922
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 728
Rounded Intercept = -576734
Correlation = 0.99703

SP
BK
4/7/2020