



INSTRUMENT PROCESSING SHEET

Agency Glades CSOS/N 80-000948Florida Department of
Law EnforcementDate In 11/05/2020 DI Completion Date 11/19/2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Quality Checks	Flow Calibration																																																												
Performed By <u>TDG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Performed By <u>TDG</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>238</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.247</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>TDG</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.82</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5097</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP4863	201905A 05/14/2021																																																												
0.080	MP4864	201905B 05/14/2021																																																												
0.200	MP5097	201904D 04/30/2021																																																												
0.080 DGS	N/A	AG003005 01/30/2022																																																												
Final Release Date FDLE Alcohol Testing Program	Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.23 12:33:24 -05'00'																																																													
Calibration Adjustment Performed By <u>MH</u> Barometric Pressure Gauge <u>1026</u> ID # <u>28199</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>20190</td> <td>04/06/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/20/2022 11/20/20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005</td> <td>01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	20060	02/10/2022	0.100	MP5099	20190	04/06/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/20/2022 11/20/20	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP4863	201905A	05/14/2021	0.080	MP4864	201905B	05/14/2021	0.200	MP5097	201904D	04/30/2021	0.080 DGS	N/A	AG003005	01/30/2022	Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1026</u> Instrument <u>1025</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____		Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	MP5095	N/A	N/A																																																											
0.040	MP5098	20060	02/10/2022																																																											
0.100	MP5099	20190	04/06/2022																																																											
0.200	MP5100	20160	03/18/2022																																																											
0.300	MP5101	20030	01/20/2022 11/20/20																																																											
0.080 DGS	N/A	08819080A1	06/05/2021																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	MP4863	201905A	05/14/2021																																																											
0.080	MP4864	201905B	05/14/2021																																																											
0.200	MP5097	201904D	04/30/2021																																																											
0.080 DGS	N/A	AG003005	01/30/2022																																																											
Simulator	Serial Number																																																													
0.000	SD1014																																																													
Interferent	SD1015																																																													
0.050	MP4863																																																													
0.080	MP4864																																																													
0.200	MP5097																																																													
Notes/Suggested Service: Correct expiration date for the 0.300 g/210L ARS lot number 20030 used in the calibration adjustment is: 01/21/2022 2020.11.20 10:18:51 -05'00'																																																														
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Israel Soto <small>Digitally signed by Israel Soto Date: 2020.11.20 10:42:22 -05'00'</small> 2020.11.23 12:32:44 -05'00' Tech Review / Date Admin Review / Date																																																														

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY S.O.
Time of Inspection: 12:38

Date of Inspection: 11/19/2020

Serial Number: 80-000948
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.198	0.079


Standard Deviations	0.0003	0.0000	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:


Israel
Soto

Digitally signed
by Israel Soto
Date: 2020.11.20
10:39:57 -05'00'

 2020.11.20
3 12:32:11
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY





Signature and Printed Name

11/19/2020
Date


88

2020.11.2
3 12:31:40
-05'00'

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000948	Glades CSO	11/06/2020	TDG <i>MG</i>

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	
<div>✓</div>		<div>✗</div>		<div>✓</div>		<div>✓</div>	
<div>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/06/2020 Software: 8100.27 SN 80-000948</div>		<div>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/06/2020 Software: 8100.27 SN 80-000948</div>		<div>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/06/2020 Software: 8100.27 SN 80-000948</div>		<div>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/06/2020 Software: 8100.27 SN 80-000948</div>	
<div>Test g/210L Time</div> <div>Air Blank 0.000 10:07</div> <div>Control Test 0.047 10:08</div> <div>Air Blank 0.000 10:08</div> <div>Control Test 0.047 10:09</div> <div>Air Blank 0.000 10:09</div> <div>Control Test 0.047 10:10</div> <div>Air Blank 0.000 10:11</div> <div>Control Test Stats</div> <div>Average 0.0470</div> <div>Std Dev 0.0000</div> <div>Rel Std Dev(%) 0.0000</div> <div>Operator's Signature </div>		<div>Test g/210L Time</div> <div>Air Blank 0.000 10:15</div> <div>Control Test 0.077 10:16</div> <div>Air Blank 0.000 10:16</div> <div>Control Test 0.077 10:17</div> <div>Air Blank 0.000 10:17</div> <div>Control Test 0.076 10:18</div> <div>Air Blank 0.000 10:19</div> <div>Control Test Stats</div> <div>Average 0.0767</div> <div>Std Dev 0.0006</div> <div>Rel Std Dev(%) 0.7531</div> <div>Operator's Signature </div>		<div>Test g/210L Time</div> <div>Air Blank 0.000 10:22</div> <div>Control Test 0.197 10:22</div> <div>Air Blank 0.000 10:23</div> <div>Control Test 0.198 10:24</div> <div>Air Blank 0.000 10:24</div> <div>Control Test 0.197 10:25</div> <div>Air Blank 0.000 10:25</div> <div>Control Test Stats</div> <div>Average 0.1973</div> <div>Std Dev 0.0006</div> <div>Rel Std Dev(%) 0.2926</div> <div>Operator's Signature </div>		<div>Test g/210L Time</div> <div>Air Blank 0.000 10:27</div> <div>Control Test 0.077 10:27</div> <div>Air Blank 0.000 10:28</div> <div>Control Test 0.078 10:28</div> <div>Air Blank 0.000 10:29</div> <div>Control Test 0.078 10:29</div> <div>Air Blank 0.000 10:29</div> <div>Control Test Stats</div> <div>Average 0.0777</div> <div>Std Dev 0.0006</div> <div>Rel Std Dev(%) 0.7434</div> <div>Operator's Signature </div>	

Comments: Stabilities failed. Will perform optical calibration. *MG* 11/06/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000948, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000948</u>	UNCERTAINTY* \pm
Owning Agency:	<u>GLADES COUNTY S.O.</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/19/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>12:38</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/19/2020

Date

MICHAEL D'HAUGHEY,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

GLADES COUNTY S.O.
Intoxilyzer - Alcotest Analyzer
Model 8000
11/19/2020
09:36:09
SN 80-000948

Auto Calibration
Max Power Res Value = 29
Auto Range Res Value = 28

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12902, Sum Io = 13745

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1240 (-0.0160)
Sample #2 = 0.1100 (0.0110)
Sample #3 = 0.1100 (-0.0050)
Sample #4 = 0.0950 (0.0000)
Avg % Abs = 0.1050 (0.0020)
STD DEV = 0.0087 (0.0082)
REL STD DEV = 8.248 (409.268)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1400 (-0.0240)
Sample #2 = 0.1140 (0.0050)
Sample #3 = 0.1100 (-0.0070)
Sample #4 = 0.1040 (0.0070)
Avg % Abs = 0.1093 (0.0020)
STD DEV = 0.0050 (0.0078)
REL STD DEV = 4.604 (390.512)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12894, Sum Io = 13739

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8510 (-0.0140)
Sample #2 = 0.8580 (-0.0160)
Sample #3 = 0.8270 (0.0260)
Sample #4 = 0.8190 (0.0340)
Avg % Abs = 0.8347 (0.0147)
STD DEV = 0.0206 (0.0259)
REL STD DEV = 2.468 (183.120)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5630 (-0.0310)
Sample #2 = 1.5580 (-0.0230)
Sample #3 = 1.5410 (0.0020)
Sample #4 = 1.5360 (0.0090)
Avg % Abs = 1.5450 (-0.0040)
STD DEV = 0.0115 (0.0168)
REL STD DEV = 0.746 (420.565)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12890, Sum Io = 13739

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 2.0110 (-0.0020)
Sample #2 = 2.0020 (0.0010)
Sample #3 = 1.9760 (0.0020)
Sample #4 = 1.9550 (0.0210)
Avg % Abs = 1.9790 (0.0080)
STD DEV = 0.0217 (0.0113)
REL STD DEV = 1.094 (140.868)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.6800 (0.0020)
Sample #2 = 3.6720 (0.0100)
Sample #3 = 3.6690 (0.0080)
Sample #4 = 3.6650 (0.0120)
Avg % Abs = 3.6687 (0.0100)
STD DEV = 0.0035 (0.0020)
REL STD DEV = 0.096 (20.000)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12773, Sum Io = 13659

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.8030 (-0.0300)
Sample #2 = 3.7740 (-0.0010)
Sample #3 = 3.8150 (-0.0140)
Sample #4 = 3.7830 (-0.0020)
Avg % Abs = 3.7907 (-0.0057)
STD DEV = 0.0215 (0.0072)
REL STD DEV = 0.568 (127.662)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 7.0160 (0.0070)
Sample #2 = 6.9950 (0.0030)
Sample #3 = 7.0150 (0.0000)
Sample #4 = 6.9990 (0.0040)
Avg % Abs = 7.0030 (0.0023)
STD DEV = 0.0106 (0.0021)
REL STD DEV = 0.151 (89.214)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12782, Sum Io = 13675

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.4850 (-0.0170)
Sample #2 = 5.4950 (0.0140)
Sample #3 = 5.4870 (0.0220)
Sample #4 = 5.5100 (0.0130)
Avg % Abs = 5.4973 (0.0163)
STD DEV = 0.0117 (0.0049)
REL STD DEV = 0.212 (30.201)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 10.0790 (-0.0030)
Sample #2 = 10.0910 (0.0220)
Sample #3 = 10.0790 (0.0340)
Sample #4 = 10.0530 (0.0370)
Avg % Abs = 10.0777 (0.0310)
STD DEV = 0.0140 (0.0075)
REL STD DEV = 0.139 (25.604)

Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.01 Rel Std Dev = 8.25
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.835
Std Dev = 0.02 Rel Std Dev = 2.47
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.979

Std Dev = 0.02 Rel Std Dev = 1.09
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.791
Std Dev = 0.02 Rel Std Dev = 0.57
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.497
Std Dev = 0.01 Rel Std Dev = 0.21
Zero Order Coef = -218.99
First Order Coef = 2449.58
Second Order Coef = 33.92
Standard Deviation = 37.627892

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.109
Std Dev = 0.01 Rel Std Dev = 4.60
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.545
Std Dev = 0.01 Rel Std Dev = 0.75
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.669
Std Dev = 0.00 Rel Std Dev = 0.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.003
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.078
Std Dev = 0.01 Rel Std Dev = 0.14
Zero Order Coef = -120.15
First Order Coef = 1270.96
Second Order Coef = 15.63
Standard Deviation = 20.325151

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0008
0.040 0.039 0.0012
0.100 0.100 0.0000
0.200 0.201 -0.0006
0.300 0.300 0.0003

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.039 0.0005
0.100 0.100 0.0002
0.200 0.200 -0.0005
0.300 0.300 0.0002

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 3072.00
Sample #2 = 3044.00
Sample #3 = 3058.00
Sample #4 = 3087.00
Average Result = 3063.0000
STD DEV = 21.9317
REL STD DEV = 0.716

Channel 2
Sample #1 = 3302.00
Sample #2 = 3272.00
Sample #3 = 3312.00
Sample #4 = 3336.00
Average Result = 3306.6667
STD DEV = 32.3316
REL STD DEV = 0.978
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1025
3 um H2O Adjust (mg/l*10,000) = 746
9 um H2O Adjust (mg/l*10,000) = 503
**** AUTO CAL PASS

Cal Adj.
80-000948
11/19/2020
MJC

Israel
Soto

Digitally signed by
Israel Soto
Date: 2020.11.20
19:14:05-0500

2020.11.2
3 12:29:14
-05'00'

2020.11.2
3 12:28:34
-05'00'

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities <i>POST</i>	80-000948	<i>Glades County SO</i>	<i>11/19/2020</i>	<i>MS</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27</p> <p>SN 80-000948</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0480</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p>Operator's Signature <i>MS</i></p>	Test	g/210L	Time	Air Blank	0.000	10:14	Control Test	0.048	10:15	Air Blank	0.000	10:16	Control Test	0.048	10:16	Air Blank	0.000	10:17	Control Test	0.048	10:17	Air Blank	0.000	10:18	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27</p> <p>SN 80-000948</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr><tr><td>Control Test</td><td>0.077</td><td>10:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr></table> <p>Operator's Signature <i>MS</i></p>	Test	g/210L	Time	Air Blank	0.000	10:19	Control Test	0.078	10:20	Air Blank	0.000	10:21	Control Test	0.077	10:21	Air Blank	0.000	10:22	Control Test	0.078	10:22	Air Blank	0.000	10:23	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27</p> <p>SN 80-000948</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p>Operator's Signature <i>MS</i></p>	Test	g/210L	Time	Air Blank	0.000	10:27	Control Test	0.198	10:28	Air Blank	0.000	10:29	Control Test	0.198	10:29	Air Blank	0.000	10:30	Control Test	0.198	10:30	Air Blank	0.000	10:31	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>GLADES COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27</p> <p>SN 80-000948</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p>Operator's Signature <i>MS</i></p>	Test	g/210L	Time	Air Blank	0.000	10:32	Control Test	0.080	10:33	Air Blank	0.000	10:33	Control Test	0.080	10:33	Air Blank	0.000	10:34	Control Test	0.080	10:34	Air Blank	0.000	10:35	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:14																																																																																																																																																	
Control Test	0.048	10:15																																																																																																																																																	
Air Blank	0.000	10:16																																																																																																																																																	
Control Test	0.048	10:16																																																																																																																																																	
Air Blank	0.000	10:17																																																																																																																																																	
Control Test	0.048	10:17																																																																																																																																																	
Air Blank	0.000	10:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test	0.078	10:20																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.077	10:21																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.078	10:22																																																																																																																																																	
Air Blank	0.000	10:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.198	10:28																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.198	10:29																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.198	10:30																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test	0.080	10:33																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.080	10:33																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test	0.080	10:34																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		