



INSTRUMENT PROCESSING SHEET

Agency Osceola Corrections S/N 80-000968

Florida Department of Law Enforcement

Date In 06/04/2020 DI Completion Date 6/26/20 Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>VS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>164</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.07.02 15:11:55 -04'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.9</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5090</u>																								
Simulator	Serial #	Lot #/Exp																																							
0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>																																							
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>																																							
0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>																																							
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																							
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1017</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>20060</u></td> <td><u>2-10-22</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>20190</u></td> <td><u>4-6-22</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>20160</u></td> <td><u>3-18-22</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>20030</u></td> <td><u>1-21-22</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>08819080A1</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>20060</u>	<u>2-10-22</u>	0.100	<u>MP5083</u>	<u>20190</u>	<u>4-6-22</u>	0.200	<u>MP5084</u>	<u>20160</u>	<u>3-18-22</u>	0.300	<u>MP5085</u>	<u>20030</u>	<u>1-21-22</u>	0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1022</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																						
0.000	<u>MP5091</u>	N/A	N/A																																						
0.040	<u>MP5082</u>	<u>20060</u>	<u>2-10-22</u>																																						
0.100	<u>MP5083</u>	<u>20190</u>	<u>4-6-22</u>																																						
0.200	<u>MP5084</u>	<u>20160</u>	<u>3-18-22</u>																																						
0.300	<u>MP5085</u>	<u>20030</u>	<u>1-21-22</u>																																						
0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>																																						
Simulator	Serial Number																																								
0.000	<u>MP5086</u>																																								
Interferent	<u>MP5087</u>																																								
0.050	<u>MP5088</u>																																								
0.080	<u>MP5089</u>																																								
0.200	<u>MP5090</u>																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 x2 <input type="checkbox"/> Other _____																				
Simulator	Serial Number	Lot Number	Expiration																																						
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																						
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																						
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																						
0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>																																						
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																								
_____ _____	<table style="width:100%;"> <tr> <td style="width:50%; text-align: center;"> 2020.07.02 <u>Michael D. Hough</u> 10:28:21 -04'00' </td> <td style="width:50%; text-align: center;"> 2020.07.02 15:06:57 <u>Brett Kirkland</u> -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review / Date</td> <td style="text-align: center;">Admin Review / Date</td> </tr> </table>	2020.07.02 <u>Michael D. Hough</u> 10:28:21 -04'00'	2020.07.02 15:06:57 <u>Brett Kirkland</u> -04'00'	Tech Review / Date	Admin Review / Date																																				
2020.07.02 <u>Michael D. Hough</u> 10:28:21 -04'00'	2020.07.02 15:06:57 <u>Brett Kirkland</u> -04'00'																																								
Tech Review / Date	Admin Review / Date																																								



Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000968, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000968</u>	UNCERTAINTY* ±	
Owning Agency:	<u>OSCEOLA CORRECTIONS</u>	0.050 g/210 L	0.004
Calibration Date:	<u>06/26/2020</u>	0.080 g/210 L	0.005
Calibration Time:	<u>12:22</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

06/26/2020

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MK
BK
2020.07.02
15:07:23
-04'00"

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 12:22

Date of Inspection: 06/26/2020

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080

Standard Deviations	0.0000	0.0009	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

MH

BK 2020.07.0
2 15:07:49
-04'00'

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/26/2020
Date

Stability Checks

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:12
Control Test	0.050	09:13
Air Blank	0.000	09:14
Control Test	0.050	09:14
Air Blank	0.000	09:15
Control Test	0.050	09:15
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.081	09:18
Air Blank	0.000	09:19
Control Test	0.081	09:19
Air Blank	0.000	09:20
Control Test	0.082	09:21
Air Blank	0.000	09:21
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.203	09:25
Air Blank	0.000	09:25
Control Test	0.204	09:26
Air Blank	0.000	09:27
Control Test	0.204	09:27
Air Blank	0.000	09:28
Control Test Stats		
Average	0.2037	
Std Dev	0.0006	
Rel Std Dev(%)	0.2835	

Wet



Operator's Signature



Operator's Signature



Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.077	09:29
Air Blank	0.000	09:30
Control Test	0.078	09:30
Air Blank	0.000	09:31
Control Test	0.078	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

MA

BK 2020.07.0
 2 15:08:22
 -04'00'

Dry



Operator's Signature

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5030 (-0.0350)
 Sample #2 = 1.4700 (-0.0150)
 Sample #3 = 1.4820 (0.0160)
 Sample #4 = 1.5110 (0.0150)
 Avg % Abs = 1.4877 (0.0053)
 STD DEV = 0.0211 (0.0176)
 REL STD DEV = 1.417 (330.305)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.8940 (-0.0100)
 Sample #2 = 6.8660 (-0.0070)
 Sample #3 = 6.9060 (0.0000)
 Sample #4 = 6.8860 (0.0050)
 Avg % Abs = 6.8860 (-0.0007)
 STD DEV = 0.0200 (0.0060)
 REL STD DEV = 0.290 (904.157)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.097
 Std Dev = 0.01 Rel Std Dev = 12.51
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.841
 Std Dev = 0.01 Rel Std Dev = 0.61
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.969
 Std Dev = 0.01 Rel Std Dev = 0.55
 Sol Val = 0.9524 mg/l or 0.210 g/210L
 % Abs = 3.771
 Std Dev = 0.01 Rel Std Dev = 0.20
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.498
 Std Dev = 0.01 Rel Std Dev = 0.11
 Zero Order Coef = -223.70
 First Order Coef = 2475.53
 Second Order Coef = 29.62
 Standard Deviation = 15.765910

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0003
 0.040 0.039 0.0005
 0.100 0.100 -0.0001
 0.200 0.200 -0.0002
 0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0004
 0.040 0.040 0.0003
 0.100 0.100 0.0004
 0.200 0.200 -0.0006
 0.300 0.300 0.0002

OSEOLA CORRECTIONS
 Intoxilizer - Alcohol Analyzer
 Model 8100
 06/25/2020
 SN 80-001968
 10:34:55

Auto Calibration
 Max Power Res Value = 89
 Auto Range Res Value = 50
 Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12458, Sum Io = 13083

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.9650 (-0.0190)
 Sample #2 = 1.9600 (0.0040)
 Sample #3 = 1.9810 (0.0230)
 Sample #4 = 1.9660 (0.0480)
 Avg % Abs = 1.9690 (0.0250)
 STD DEV = 0.0108 (0.0221)
 REL STD DEV = 0.549 (88.272)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.9710 (0.0010)
 Sample #2 = 9.9550 (0.0170)
 Sample #3 = 9.9780 (0.0200)
 Sample #4 = 9.9490 (0.0330)
 Avg % Abs = 9.9607 (0.0233)
 STD DEV = 0.0153 (0.0085)
 REL STD DEV = 0.154 (36.451)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.076
 Std Dev = 0.03 Rel Std Dev = 39.56
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.488
 Std Dev = 0.02 Rel Std Dev = 1.42
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.566
 Std Dev = 0.00 Rel Std Dev = 0.09
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.886
 Std Dev = 0.02 Rel Std Dev = 0.29
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.961
 Std Dev = 0.02 Rel Std Dev = 0.15
 Zero Order Coef = -82.90
 First Order Coef = 1304.55
 Second Order Coef = 13.74
 Standard Deviation = 21.029495

***** CHANNEL 2 *****
 ***** CHANNEL 2 *****
 Sample #1 = 3477.00
 Sample #2 = 3492.00
 Sample #3 = 3516.00
 Sample #4 = 3477.00
 Average Result = 3495.0000
 STD DEV = 19.6723
 REL STD DEV = 0.563

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 620
 9 um H2O Adjust (mg/l*10,000) = 314
 ***** AUTO CAL PASS *****

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5920 (-0.0300)
 Sample #2 = 3.5700 (0.0000)
 Sample #3 = 3.5650 (0.0170)
 Sample #4 = 3.5640 (0.0160)
 Avg % Abs = 3.5663 (0.0110)
 STD DEV = 0.0032 (0.0095)
 REL STD DEV = 0.090 (86.722)

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.7650 (-0.0160)
 Sample #2 = 3.7660 (-0.0030)
 Sample #3 = 3.7800 (0.0150)
 Sample #4 = 3.7680 (0.0350)
 Avg % Abs = 3.7713 (0.0157)
 STD DEV = 0.0076 (0.0190)
 REL STD DEV = 0.201 (121.333)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12409, Sum Io = 13063

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12421, Sum Io = 13069

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.100 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3208.00
 Sample #2 = 3221.00
 Sample #3 = 3195.00
 Sample #4 = 3151.00
 Average Result = 3189.0000
 STD DEV = 35.3836
 REL STD DEV = 1.110

 ***** CHANNEL 2 *****
 Sample #1 = 3477.00
 Sample #2 = 3492.00
 Sample #3 = 3516.00
 Sample #4 = 3477.00
 Average Result = 3495.0000
 STD DEV = 19.6723
 REL STD DEV = 0.563

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 620
 9 um H2O Adjust (mg/l*10,000) = 314
 ***** AUTO CAL PASS *****

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8340 (-0.0300)
 Sample #2 = 0.8400 (-0.0040)
 Sample #3 = 0.8370 (0.0470)
 Sample #4 = 0.8470 (0.0610)
 Avg % Abs = 0.8413 (0.0347)
 STD DEV = 0.0051 (0.0342)
 REL STD DEV = 0.610 (98.683)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12409, Sum Io = 13063

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12421, Sum Io = 13069

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.100 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3208.00
 Sample #2 = 3221.00
 Sample #3 = 3195.00
 Sample #4 = 3151.00
 Average Result = 3189.0000
 STD DEV = 35.3836
 REL STD DEV = 1.110

 ***** CHANNEL 2 *****
 Sample #1 = 3477.00
 Sample #2 = 3492.00
 Sample #3 = 3516.00
 Sample #4 = 3477.00
 Average Result = 3495.0000
 STD DEV = 19.6723
 REL STD DEV = 0.563

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 620
 9 um H2O Adjust (mg/l*10,000) = 314
 ***** AUTO CAL PASS *****

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12421, Sum Io = 13069

MH
 BK
 2020.07.0
 215:08:46
 -04'00"
 CAL ADJUSTMENT
 #80-000968 SP

Post Cal Adjust Stability Checks # 80-000968

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/26/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:05
Control Test	0.049	09:05
Air Blank	0.000	09:06
Control Test	0.049	09:06
Air Blank	0.000	09:07
Control Test	0.048	09:08
Air Blank	0.600	09:08
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1853	

SP

 Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/26/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.079	09:10
Air Blank	0.000	09:10
Control Test	0.079	09:11
Air Blank	0.000	09:12
Control Test	0.080	09:12
Air Blank	0.000	09:13
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

 Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/26/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.199	09:16
Air Blank	0.000	09:17
Control Test	0.199	09:17
Air Blank	0.000	09:18
Control Test	0.200	09:19
Air Blank	0.000	09:19
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

SP

 Operator's Signature

OSCEOLA CORRECTIONS
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000968
 06/26/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:00
Control Test	0.080	09:00
Air Blank	0.000	09:00
Control Test	0.081	09:01
Air Blank	0.000	09:01
Control Test	0.081	09:02
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DAS

 Operator's Signature

MAH

BK
 2020.07.
 02
 15:09:12
 -04'00"

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 08:09

Date of Inspection: 06/08/2020

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

MSH

BK 2020.07.02
15.0933-0400

MSH
N/A compliance not determined

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/08/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 10:34

Date of Inspection: 06/25/2020

Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

MH

BK 2020.07.02
15:10:03
-04'00'

*n/a compliance
not determined*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/25/2020
Date

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Dan Lyons on 3/24/2020

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000968

Bill To Address:
Osceola Corrections

Ship to Address:
FDLE Tallahassee

Reason for Return:
Instrument will not power on. Please try to upload data prior to repair.

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$_____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Dan Lyons
Phone #: (407) 709-0010 Email: Dlyo@osceola.org

ATP Contact Name: Shayla Platt ATP Email: shaylplatt@fdle.state.fl.us

MH
2020.07.
02
BK 15:10:28
-04'00'



INSTRUMENT PROCESSING SHEET

Agency Osceola Corrections s/N 80-000968

Florida Department of Law Enforcement

Date In 02/28/2020 DI Completion Date _____ Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u>		Quality Checks Performed By _____		Flow Calibration Performed By _____																
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
Final Release Date		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050																				
0.080																				
0.200																				
0.080 DGS	N/A																			
				Temperature Checks Performed By _____ <input type="checkbox"/> Lab Temp °C _____ External Digital Therm. ID#: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____																

Calibration Adjustment Performed By _____				Department Inspection Performed By _____																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# _____																															
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____			
Simulator	Serial Number	Lot Number	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.050</td><td></td></tr> <tr><td>0.080</td><td></td></tr> <tr><td>0.200</td><td></td></tr> </tbody> </table>				Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																	
Simulator	Serial Number																																		
0.000																																			
Interferent																																			
0.050																																			
0.080																																			
0.200																																			
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____											
Simulator	Serial Number	Lot Number	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		

Notes/Suggested Service: _____ _____ _____ _____ _____		<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	
Tech Review / Date _____		Admin Review / Date _____	

MK
2020.07.02
15:10:55
-04'00"
BK