



INSTRUMENT PROCESSING SHEET

Agency Cocoa PD S/N 80-001000
 Florida Department of Law Enforcement Date In 09/25/2020 DI Completion Date 09/28/2020 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>TDG</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE Alcohol Testing Program	Quality Checks Performed By <u>TDG</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>194</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>TDG</u> ☒ Lab Temp °C <u>23.00</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +- .2 Serial #: <u>SD3967</u> ☒ 34°C +- .2 Serial #: <u>SD3968</u> ☒ 34°C +- .2 Serial #: <u>SD3969</u>																																															
Simulator	Serial #	Lot #/Exp																																																														
0.050	SD3967	201905A 05/14/2021																																																														
0.080	SD3968	201905B 05/14/2021																																																														
0.200	SD3969	201904D 04/30/2021																																																														
0.080 DGS	N/A	AG003005 01/30/2022																																																														
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>			Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1011</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																													
0.000		N/A	N/A																																																													
0.040																																																																
0.100																																																																
0.200																																																																
0.300																																																																
0.080 DGS	N/A																																																															
Simulator	Serial Number	Lot Number	Expiration																																																													
0.050																																																																
0.080																																																																
0.200																																																																
0.080 DGS	N/A																																																															
Simulator	Serial Number																																																															
0.000	SD3965																																																															
Interferent	SD3966																																																															
0.050	SD3967																																																															
0.080	SD3968																																																															
0.200	SD3969																																																															
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes Rivera Tech Review / Date Digitally signed by David Eliezer Reyes Rivera Date: 2020.09.29 09:13:35 -04'00' Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA PD

Time of Inspection: 14:27

Date of Inspection: 09/28/2020

Serial Number: 80-001000

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.047	0.077	0.198	0.077
0.000	0.047	0.078	0.200	0.077
0.000	0.048	0.078	0.199	0.076
0.000	0.048	0.078	0.198	0.076
0.000	0.048	0.079	0.198	0.075
0.000	0.048	0.079	0.198	0.075
0.000	0.048	0.078	0.197	0.075
0.000	0.048	0.079	0.198	0.075
0.000	0.047	0.079	0.198	0.076
0.000	0.048	0.078	0.198	0.075


Standard Deviations	0.0004	0.0006	0.0007	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

09/28/2020
Date

DERR
Digitally signed
by DERR
Date: 2020.09.29
09:12:37 -04'00'

 2020.09.29
11:18:05
-04'00'

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1000	Coca PD	09/28/2020	TDG MG

0.05g/210L 0.047 to 0.053		0.08g/210L 0.077 to 0.083		0.20g/210L 0.194 to 0.206		DGS 0.08g/210L 0.077 to 0.083	
✓		✓		✓		✓	
COCOA PD Intoxilyzer - Alconol Analyzer Model 8000 09/28/2020 Software: 8100.27		COCOA PD Intoxilyzer - Alconol Analyzer Model 8000 09/28/2020 Software: 8100.27		COCOA PD Intoxilyzer - Alconol Analyzer Model 8000 09/28/2020 Software: 8100.27		COCOA PD Intoxilyzer - Alconol Analyzer Model 8000 09/28/2020 Software: 8100.27	
DGS							
Test g/210L Time		Test g/210L Time		Test g/210L Time		Test g/210L Time	
Air Blank 0.000 11:43		Air Blank 0.000 11:49		Air Blank 0.000 11:55		Air Blank 0.000 12:00	
Control Test 0.048 11:44		Control Test 0.078 11:49		Control Test 0.197 11:56		Control Test 0.077 12:01	
Air Blank 0.000 11:45		Air Blank 0.000 11:50		Air Blank 0.000 11:56		Air Blank 0.000 12:01	
Control Test 0.047 11:45		Control Test 0.078 11:51		Control Test 0.198 11:57		Control Test 0.077 12:02	
Air Blank 0.000 11:46		Air Blank 0.000 11:51		Air Blank 0.000 11:58		Air Blank 0.000 12:02	
Control Test 0.048 11:47		Control Test 0.078 11:52		Control Test 0.197 11:58		Control Test 0.077 12:02	
Air Blank 0.000 11:47		Air Blank 0.000 11:53		Air Blank 0.000 11:59		Air Blank 0.000 12:03	
Control Test Stats		Control Test Stats		Control Test Stats		Control Test Stats	
Average 0.0477		Average 0.0780		Average 0.1973		Average 0.0770	
Std Dev 0.0005		Std Dev 0.0000		Std Dev 0.0006		Std Dev 0.0000	
Rel Std Dev(%) 1.2112		Rel Std Dev(%) 0.0000		Rel Std Dev(%) 0.2826		Rel Std Dev(%) 0.0000	
Operator's Signature 		Operator's Signature 		Operator's Signature 		Operator's Signature 	

Comments:	DERR	2020.09.29 11:18:37
-----------	------	---------------------



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001000, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001000</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COCOA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/28/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>14:27</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/28/2020

Date


MICHAEL D HAUGHEY,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

DERR

Digitally signed
by DERR
Date: 2020.09.29
09:11:22 -0400

2020.09.29
9 11:19:08
-04'00'





INSTRUMENT PROCESSING SHEET

Agency Cocoa PDS/N 80-001000Florida Department of
Law EnforcementDate In 1/2/2020DI Completion Date 1/9/20☒ Ship☐ P/U☐ H/D☐ CMI☐ EE

Intake Performed By <u>DP</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>DP</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>193</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE JAN 13 2020 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG016501 06-14-2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG016501 06-14-2021	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.5</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +- .2 Serial #: <u>MP5088</u> ☒ 34°C +- .2 Serial #: <u>MP5089</u> ☒ 34°C +- .2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP5088	201905A 05-14-2021																																																													
0.080	MP5089	201905B 05-14-2021																																																													
0.200	MP5090	201904D 04-30-2021																																																													
0.080 DGS	N/A	AG016501 06-14-2021																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1030</u> Instrument <u>1028</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5086																																																														
Interferent	MP5087																																																														
0.050	MP5088																																																														
0.080	MP5089																																																														
0.200	MP5090																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																													
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																													
		<u>SPM 1/9/2020</u> <u>1/10/2020</u> Tech Review / Date Admin Review / Date																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001000, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001000</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COCOA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/09/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/09/2020

Date

Shayla Platt

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Adam BK-8
1/10/2020

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA PD

Time of Inspection: 13:37

Date of Inspection: 01/09/2020

Serial Number: 80-001000

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.080	0.201	0.077

Standard Deviations	0.0000	0.0004	0.0004	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

01/09/2020
Date

13K
1/10/2020

Stability Checks

COCOA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001000
01/09/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:15
Control Test	0.050	11:16
Air Blank	0.000	11:16
Control Test	0.049	11:17
Air Blank	0.000	11:18
Control Test	0.049	11:18
Air Blank	0.000	11:19
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SS

Operator's Signature

COCOA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001000
01/09/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.080	11:22
Air Blank	0.000	11:23
Control Test	0.079	11:23
Air Blank	0.000	11:24
Control Test	0.078	11:25
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

wet

SS

Operator's Signature

COCOA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001000
01/09/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.201	11:27
Air Blank	0.000	11:28
Control Test	0.199	11:28
Air Blank	0.000	11:29
Control Test	0.199	11:29
Air Blank	0.000	11:30
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel Std Dev(%)	0.5783	

SS

Operator's Signature

COCOA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001000
01/09/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	0.079	11:31
Air Blank	0.000	11:32
Control Test	0.079	11:32
Air Blank	0.000	11:33
Control Test	0.079	11:33
Air Blank	0.000	11:34
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

SS

Operator's Signature

SSM
BK
1/10/2020