

INSTRUMENT PROCESSING SHEET

Agency Hardee County Sheriff's OfficeS/N 80-001077Florida Department of
Law EnforcementDate In 03/09/2020 DI Completion Date 03/17/2020 Ship P/U H/D CMI EE

Intake Performed By <u>TG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dry Gas Leak</u>	Quality Checks Performed By <u>MK</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date <p style="text-align: center;">FDLE APR 09 2020 Alcohol Testing Program</p>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501 06/14/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG916501 06/14/2021	Maintenance Performed By <u>MK</u> <input type="checkbox"/> Battery Replacement <input checked="" type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Dry Gas Cal</u> Temperature Checks Performed By <u>MK</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.52</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1017</u>
Simulator	Serial #	Lot #/Exp															
0.050	MP4863	201905A 05/14/2021															
0.080	MP4864	201905B 05/14/2021															
0.200	SD3969	201904D 04/30/2021															
0.080 DGS	N/A	AG916501 06/14/2021															

Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MK</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1022/1023</u> Instrument <u>1021/1022</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>SD1017</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	SD1017
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	SD1017																																																												

Notes/Suggested Service: _____ <u>Repeated DI. Missed sim temp checks prior to first DI.</u> <input checked="" type="checkbox"/> APPROVED <u>03/19/2020</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SRP 3/26/20</u> <u>Britt K. Kuhlmann 3/26/2020</u> Tech Review / Date _____ Admin Review / Date _____
--	--

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 13:21

Date of Inspection: 03/17/2020

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.202	0.078
0.000	0.048	0.080	0.202	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.079	0.200	0.078

Standard Deviations	0.0003	0.0004	0.0007	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

03/17/2020
Date

*SP
RSK
3/26/2020*

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 14:05

Date of Inspection: 03/11/2020

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078

Standard Deviations	0.0005	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

03/11/2020
Date

*SP
BK
3/26/2020*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001077	Howe County SO	3/11/2020	MXC

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

<p>HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 03/11/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.1625</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.050	11:15	Air Blank	0.000	11:15	Control Test	0.049	11:16	Air Blank	0.000	11:16	Control Test	0.050	11:17	Air Blank	0.000	11:18	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	0.1625		<p>HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 03/11/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:22	Control Test	0.079	11:23	Air Blank	0.000	11:24	Control Test	0.080	11:24	Air Blank	0.000	11:25	Control Test	0.080	11:25	Air Blank	0.000	11:26	Control Test	0.080	11:26	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<p>HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 03/11/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:28	Control Test	0.199	11:28	Air Blank	0.000	11:29	Control Test	0.199	11:30	Air Blank	0.000	11:30	Control Test	0.199	11:30	Air Blank	0.000	11:31	Control Test	0.000	11:31	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 03/11/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:33	Control Test	0.079	11:34	Air Blank	0.000	11:34	Control Test	0.079	11:34	Air Blank	0.000	11:35	Control Test	0.080	11:35	Air Blank	0.000	11:36	Control Test	0.000	11:36	Control Test Stats			Average	0.0793		Std Dev	0.0005		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																										
Air Blank	0.000	11:14																																																																																																																																																										
Control Test	0.050	11:15																																																																																																																																																										
Air Blank	0.000	11:15																																																																																																																																																										
Control Test	0.049	11:16																																																																																																																																																										
Air Blank	0.000	11:16																																																																																																																																																										
Control Test	0.050	11:17																																																																																																																																																										
Air Blank	0.000	11:18																																																																																																																																																										
Control Test Stats																																																																																																																																																												
Average	0.0497																																																																																																																																																											
Std Dev	0.0006																																																																																																																																																											
Rel Std Dev(%)	0.1625																																																																																																																																																											
Test	g/210L	Time																																																																																																																																																										
Air Blank	0.000	11:22																																																																																																																																																										
Control Test	0.079	11:23																																																																																																																																																										
Air Blank	0.000	11:24																																																																																																																																																										
Control Test	0.080	11:24																																																																																																																																																										
Air Blank	0.000	11:25																																																																																																																																																										
Control Test	0.080	11:25																																																																																																																																																										
Air Blank	0.000	11:26																																																																																																																																																										
Control Test	0.080	11:26																																																																																																																																																										
Control Test Stats																																																																																																																																																												
Average	0.0797																																																																																																																																																											
Std Dev	0.0006																																																																																																																																																											
Rel Std Dev(%)	0.7247																																																																																																																																																											
Test	g/210L	Time																																																																																																																																																										
Air Blank	0.000	11:28																																																																																																																																																										
Control Test	0.199	11:28																																																																																																																																																										
Air Blank	0.000	11:29																																																																																																																																																										
Control Test	0.199	11:30																																																																																																																																																										
Air Blank	0.000	11:30																																																																																																																																																										
Control Test	0.199	11:30																																																																																																																																																										
Air Blank	0.000	11:31																																																																																																																																																										
Control Test	0.000	11:31																																																																																																																																																										
Control Test Stats																																																																																																																																																												
Average	0.1990																																																																																																																																																											
Std Dev	0.0000																																																																																																																																																											
Rel Std Dev(%)	0.0000																																																																																																																																																											
Test	g/210L	Time																																																																																																																																																										
Air Blank	0.000	11:33																																																																																																																																																										
Control Test	0.079	11:34																																																																																																																																																										
Air Blank	0.000	11:34																																																																																																																																																										
Control Test	0.079	11:34																																																																																																																																																										
Air Blank	0.000	11:35																																																																																																																																																										
Control Test	0.080	11:35																																																																																																																																																										
Air Blank	0.000	11:36																																																																																																																																																										
Control Test	0.000	11:36																																																																																																																																																										
Control Test Stats																																																																																																																																																												
Average	0.0793																																																																																																																																																											
Std Dev	0.0005																																																																																																																																																											
Rel Std Dev(%)	0.7277																																																																																																																																																											

MXC
Operator's Signature


MXC
Operator's Signature

MXC
Operator's Signature

MXC
Operator's Signature

MXC

SR
YDK
3/26/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001077</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HARDEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/17/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2020
Issuing Authority: Alcohol Testing Program

03/17/2020

Date

MICHAEL D HAUGHEY,
Department Inspector

Service • Integrity • Respect • Quality

SR JSK
3/26/2020



INSTRUMENT PROCESSING SHEET

Agency Hardee County SOS/N 80-001077Florida Department of
Law EnforcementDate In 02/06/2020 DI Completion Date 02/06/2020 Ship P/U H/D CMI EE

Intake Performed By <u>mx</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>mx</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>193</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG916501 06/14/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG916501 06/14/2021	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080	MP4864	201905B 05/14/2021																																																											
0.200	MP5097	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG916501 06/14/2021																																																											
Final Release Date <div style="text-align: center; padding: 10px;"> FDLE FEB 24 2020 Alcohol Testing Program </div>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>mx</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.50</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5097</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>mx</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1009</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												
Notes/Suggested Service: <u>E-mailed</u> <div style="margin-top: 20px;"> <input checked="" type="checkbox"/> APPROVED <u>02/07/2020</u> </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="margin-top: 10px;"> <u>SP 2/21/20</u> <u>Scott Kirkland 2/24/2020</u> Tech Review / Date Admin Review / Date </div>																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 13:43

Date of Inspection: 02/06/2020

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.080	0.201	0.078
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.081	0.202	0.077

Standard Deviations	0.0000	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

SP
BK
2/24/2020

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

02/06/2020
Date

TYPE OF TEST Stabilities	SERIAL NUMBER 80-001077	AGENCY Hardee County SO	DATE 02/06/2020	PERFORMED BY MX
------------------------------------	-----------------------------------	-----------------------------------	---------------------------	---------------------------

0.05g/210L		0.047 to 0.053 <input checked="" type="checkbox"/>	
HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001077 02/06/2020 Software: 8100.27			
Test	9/210L	Time	
Air Blank	0.000	11:54	
Control Test	0.049	11:55	
Air Blank	0.000	11:55	
Control Test	0.049	11:56	
Air Blank	0.000	11:57	
Control Test	0.049	11:57	
Air Blank	0.000	11:58	
Control Test	0.000	11:58	
Average	0.0490		
Std Dev	0.0000		
Rel. Std Dev(%)	0.0000		
MX Operator's Signature			



0.08g/210L		0.077 to 0.083 <input checked="" type="checkbox"/>	
HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001077 02/06/2020 Software: 8100.27			
Test	9/210L	Time	
Air Blank	0.000	11:59	
Control Test	0.080	12:00	
Air Blank	0.000	12:00	
Control Test	0.080	12:01	
Air Blank	0.000	12:01	
Control Test	0.080	12:02	
Air Blank	0.000	12:03	
Control Test	0.000	12:03	
Average	0.0800		
Std Dev	0.0000		
Rel. Std Dev(%)	0.0000		
MX Operator's Signature			

0.20g/210L		0.194 to 0.206 <input checked="" type="checkbox"/>	
HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001077 02/06/2020 Software: 8100.27			
Test	9/210L	Time	
Air Blank	0.000	12:04	
Control Test	0.201	12:04	
Air Blank	0.000	12:05	
Control Test	0.200	12:05	
Air Blank	0.000	12:05	
Control Test	0.200	12:06	
Air Blank	0.000	12:07	
Control Test	0.000	12:07	
Average	0.2003		
Std Dev	0.0006		
Rel. Std Dev(%)	0.2862		
MX Operator's Signature			

0.08g/210L		0.077 to 0.083 <input checked="" type="checkbox"/>	
HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001077 02/06/2020 Software: 8100.27			
Test	9/210L	Time	
Air Blank	0.000	12:09	
Control Test	0.079	12:09	
Air Blank	0.000	12:10	
Control Test	0.078	12:10	
Air Blank	0.000	12:11	
Control Test	0.078	12:11	
Air Blank	0.000	12:11	
Control Test	0.000	12:12	
Average	0.0783		
Std Dev	0.0006		
Rel. Std Dev(%)	0.7370		
MX Operator's Signature			

SP

BK
2/24/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001077</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>HARDEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/06/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/06/2020

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality