



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001083, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001083</u>	UNCERTAINTY* ±	
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>08/19/2020</u>	0.080 g/210 L	0.005
Calibration Time:	<u>15:57</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

08/19/2020

Date

SHAYLA D PLATT,
 Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

MX
BK 2020.08.21
 15:15:10
 -0400

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 15:57

Date of Inspection: 08/19/2020

Serial Number: 80-001083
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.198	0.079
Standard Deviations	0.0004	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

MX
BK 2020.08.2
1151538
0400

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

08/19/2020
Date

Stability Checks

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.047	10:13
Air Blank	0.000	10:13
Control Test	0.045	10:14
Air Blank	0.000	10:14
Control Test	0.047	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.078	10:18
Air Blank	0.000	10:19
Control Test	0.078	10:19
Air Blank	0.000	10:20
Control Test	0.078	10:21
Air Blank	0.000	10:21
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.200	10:23
Air Blank	0.000	10:24
Control Test	0.199	10:24
Air Blank	0.000	10:25
Control Test	0.200	10:26
Air Blank	0.000	10:26
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

wet



Operator's Signature



Operator's Signature



Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.077	10:28
Air Blank	0.000	10:29
Control Test	0.078	10:29
Air Blank	0.000	10:29
Control Test	0.079	10:30
Air Blank	0.000	10:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	

Dry



Operator's Signature

MX

JACKSONVILLE SC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001083
08/19/2020 08:58:48

Auto Calibration
Max Power Res Value = 82
Auto Range Res Value = 65

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12620, 9um Io = 13167

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1360 (-0.0100)
Sample #2 = 0.1040 (-0.0020)
Sample #3 = 0.1060 (0.0200)
Sample #4 = 0.1230 (0.0000)
Avg % Abs = 0.1110 (0.0060)
STD DEV = 0.0104 (0.0122)
REL STD DEV = 9.406 (202.759)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1040 (-0.0100)
Sample #2 = 0.1020 (-0.0190)
Sample #3 = 0.0720 (0.0030)
Sample #4 = 0.1100 (-0.0120)
Avg % Abs = 0.0947 (-0.0093)
STD DEV = 0.0200 (0.0112)
REL STD DEV = 21.162 (120.427)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12617, 9um Io = 13169

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8840 (-0.0050)
Sample #2 = 0.9010 (-0.0060)
Sample #3 = 0.8770 (-0.0010)
Sample #4 = 0.8590 (0.0360)
Avg % Abs = 0.8790 (0.0097)
STD DEV = 0.0211 (0.0229)
REL STD DEV = 2.397 (237.331)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5220 (0.0060)
Sample #2 = 1.5030 (0.0200)
Sample #3 = 1.5080 (0.0180)
Sample #4 = 1.5080 (0.0330)
Avg % Abs = 1.5063 (0.0237)
STD DEV = 0.0029 (0.0081)
REL STD DEV = 0.192 (34.414)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12614, 9um Io = 13166

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9700 (-0.0120)
Sample #2 = 1.9790 (-0.0060)
Sample #3 = 1.9700 (0.0030)
Sample #4 = 1.9510 (0.0300)
Avg % Abs = 1.9667 (0.0090)
STD DEV = 0.0143 (0.0187)
REL STD DEV = 0.727 (208.167)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5570 (-0.0030)
Sample #2 = 3.5810 (0.0000)
Sample #3 = 3.5520 (0.0220)
Sample #4 = 3.5450 (0.0240)
Avg % Abs = 3.5593 (0.0153)
STD DEV = 0.0191 (0.0133)
REL STD DEV = 0.536 (86.848)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12608, 9um Io = 13163

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7750 (-0.0240)
Sample #2 = 3.7850 (-0.0340)
Sample #3 = 3.7980 (-0.0150)
Sample #4 = 3.7700 (-0.0190)
Avg % Abs = 3.7843 (-0.0227)
STD DEV = 0.0140 (0.0100)
REL STD DEV = 0.370 (44.191)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.8480 (-0.0040)
Sample #2 = 6.8520 (-0.0030)
Sample #3 = 6.8690 (0.0000)
Sample #4 = 6.8450 (0.0070)
Avg % Abs = 6.8553 (0.0013)
STD DEV = 0.0123 (0.0051)
REL STD DEV = 0.180 (384.870)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12607, 9um Io = 13161

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.4590 (-0.0180)
Sample #2 = 5.4810 (0.0000)
Sample #3 = 5.4940 (-0.0130)
Sample #4 = 5.4920 (-0.0030)
Avg % Abs = 5.4890 (-0.0053)
STD DEV = 0.0070 (0.0068)
REL STD DEV = 0.128 (127.629)

Channel 2 Summary:

Sample % Abs (% Abs Ref)
Sample #1 = 9.8990 (-0.0070)
Sample #2 = 9.8850 (0.0070)
Sample #3 = 9.8850 (0.0150)
Sample #4 = 9.8730 (0.0190)
Avg % Abs = 9.8810 (0.0137)
STD DEV = 0.0069 (0.0061)
REL STD DEV = 0.070 (44.708)

AUTO CAL DATA

Channel 1 Summary:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.111
Std Dev = 0.01 Rel Std Dev = 9.41
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.879
Std Dev = 0.02 Rel Std Dev = 2.40
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.967
Std Dev = 0.01 Rel Std Dev = 0.73
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.784
Std Dev = 0.01 Rel Std Dev = 0.37
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.489
Std Dev = 0.01 Rel Std Dev = 0.13
Zero Order Coef = -277.09
First Order Coef = 2479.68
Second Order Coef = 31.26
Standard Deviation = 28.458767

Channel 2 Summary:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.02 Rel Std Dev = 21.16
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.506
Std Dev = 0.00 Rel Std Dev = 0.19
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.559
Std Dev = 0.02 Rel Std Dev = 0.54
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.855
Std Dev = 0.01 Rel Std Dev = 0.18
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.881
Std Dev = 0.01 Rel Std Dev = 0.07
Zero Order Coef = -108.69
First Order Coef = 1307.52
Second Order Coef = 14.97
Standard Deviation = 24.632324

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.099	0.0006
0.200	0.201	-0.0007
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Channel 1 Data:
Sample #1 = 3052.00
Sample #2 = 3097.00
Sample #3 = 3121.00
Sample #4 = 2983.00
Average Result = 3067.0000
STD DEV = 73.7292
REL STD DEV = 2.404

Channel 2 Data:
Sample #1 = 3409.00
Sample #2 = 3423.00
Sample #3 = 3441.00
Sample #4 = 3387.00
Average Result = 3417.0000
STD DEV = 27.4955
REL STD DEV = 0.805

Try Gas H2O Adjust Results
Barometric Pressure = 1009
3 um H2O Adjust (mg/l*10,000) = 742
9 um H2O Adjust (mg/l*10,000) = 392
*** AUTO CAL PASS

CAL
ADJUSTMENT
SP

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	-0.0005
0.100	0.099	0.0009
0.200	0.201	-0.0006
0.300	0.300	0.0002

MX

JACKSONVILLE 30
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/19/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:38
Control Test	0.079	11:38
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

*Post
 Cal
 Adjust
 Stability
 Checks*

JACKSONVILLE 50
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/19/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:52
Control Test	0.079	11:53
Air Blank	0.000	11:53
Control Test	0.079	11:54
Air Blank	0.000	11:54
Control Test	0.079	11:55
Air Blank	0.000	11:56
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

JACKSONVILLE 50
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/19/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:46
Control Test	0.050	11:47
Air Blank	0.000	11:48
Control Test	0.048	11:48
Air Blank	0.000	11:49
Control Test	0.049	11:49
Air Blank	0.000	11:50
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	

SP

Operator's Signature

JACKSONVILLE 50
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001083
 08/19/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:58
Control Test	0.200	11:58
Air Blank	0.000	11:59
Control Test	0.199	12:00
Air Blank	0.000	12:00
Control Test	0.200	12:01
Air Blank	0.000	12:01
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SP

Operator's Signature

MX

2020.08.2
 BK¹
 15:16:41
 -04'00'