



INSTRUMENT PROCESSING SHEET

Agency FHP Tampa

S/N 80-001113

Florida Department of
Law Enforcement

Date In 09/21/2020 DI Completion Date 09/23/2020

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>TDG</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>TDG</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>169</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5092</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11/12/2021</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	201905A 05/14/2021	0.080	MP5093	201905B 05/14/2021	0.200	MP5094	201904D 04/30/2021	0.080 DGS	N/A	AG931603 11/12/2021	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>TDG</u> ☒ Lab Temp °C <u>21.63</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +- .2 Serial #: <u>MP5092</u> ☒ 34°C +- .2 Serial #: <u>MP5093</u> ☒ 34°C +- .2 Serial #: <u>MP5094</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5092	201905A 05/14/2021																																																
0.080	MP5093	201905B 05/14/2021																																																
0.200	MP5094	201904D 04/30/2021																																																
0.080 DGS	N/A	AG931603 11/12/2021																																																
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.09.24 08:01:46 -04'00'	Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1015</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use David Eliezer Reyes Rivera Digitally signed by David Eliezer Reyes Rivera Date: 2020.09.23 13:07:37 -04'00' <u>2020.09.24</u> <u>08:01:03</u> Tech Review / Date Admin Review / Date		Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP5092	0.080	MP5093	0.200	MP5094																																				
Simulator	Serial Number																																																	
0.000	SD1014																																																	
Interferent	SD1015																																																	
0.050	MP5092																																																	
0.080	MP5093																																																	
0.200	MP5094																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:45

Date of Inspection: 09/23/2020

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.082
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081
0.000	0.050	0.081	0.204	0.081

Standard Deviations	0.0000	0.0000	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0000 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D HAUGHEY

Signature and Printed Name

09/23/2020
Date

DERR
Digitally signed
by DERR
Date: 2020.09.23
13:07:01 -04'00'

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1113	FHP Tampa	09/22/2020	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2020 Software: 8100.27 Testg/210LTime <table><tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Air Blank	0.000	11:11	Control Test	0.050	11:11	Air Blank	0.000	11:12	Control Test	0.050	11:13	Air Blank	0.000	11:13	Control Test	0.050	11:14	Air Blank	0.000	11:14	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2020 Software: 8100.27 Testg/210LTime <table><tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr><tr><td>Control Test</td><td>0.082</td><td>11:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr><tr><td>Control Test</td><td>0.081</td><td>11:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr><tr><td>Control Test</td><td>0.081</td><td>11:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0813</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr></table> Operator's Signature	Air Blank	0.000	11:17	Control Test	0.082	11:18	Air Blank	0.000	11:19	Control Test	0.081	11:19	Air Blank	0.000	11:20	Control Test	0.081	11:21	Air Blank	0.000	11:21	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2020 Software: 8100.27 Testg/210LTime <table><tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr><tr><td>Control Test</td><td>0.203</td><td>11:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr><tr><td>Control Test</td><td>0.203</td><td>11:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr><tr><td>Control Test</td><td>0.202</td><td>11:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2027</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2849</td><td></td></tr></table> Operator's Signature	Air Blank	0.000	11:24	Control Test	0.203	11:24	Air Blank	0.000	11:25	Control Test	0.203	11:26	Air Blank	0.000	11:26	Control Test	0.202	11:27	Air Blank	0.000	11:28	Control Test Stats			Average	0.2027		Std Dev	0.0006		Rel Std Dev(%)	0.2849		DGS FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 09/22/2020 Software: 8100.27 Testg/210LTime <table><tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr><tr><td>Control Test</td><td>0.083</td><td>11:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr><tr><td>Control Test</td><td>0.083</td><td>11:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr><tr><td>Control Test</td><td>0.083</td><td>11:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0830</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> Operator's Signature	Air Blank	0.000	11:29	Control Test	0.083	11:29	Air Blank	0.000	11:30	Control Test	0.083	11:30	Air Blank	0.000	11:30	Control Test	0.083	11:31	Air Blank	0.000	11:31	Control Test Stats			Average	0.0830		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	11:11																																																																																																																																					
Control Test	0.050	11:11																																																																																																																																					
Air Blank	0.000	11:12																																																																																																																																					
Control Test	0.050	11:13																																																																																																																																					
Air Blank	0.000	11:13																																																																																																																																					
Control Test	0.050	11:14																																																																																																																																					
Air Blank	0.000	11:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0500																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:17																																																																																																																																					
Control Test	0.082	11:18																																																																																																																																					
Air Blank	0.000	11:19																																																																																																																																					
Control Test	0.081	11:19																																																																																																																																					
Air Blank	0.000	11:20																																																																																																																																					
Control Test	0.081	11:21																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7099																																																																																																																																						
Air Blank	0.000	11:24																																																																																																																																					
Control Test	0.203	11:24																																																																																																																																					
Air Blank	0.000	11:25																																																																																																																																					
Control Test	0.203	11:26																																																																																																																																					
Air Blank	0.000	11:26																																																																																																																																					
Control Test	0.202	11:27																																																																																																																																					
Air Blank	0.000	11:28																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2027																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2849																																																																																																																																						
Air Blank	0.000	11:29																																																																																																																																					
Control Test	0.083	11:29																																																																																																																																					
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.083	11:30																																																																																																																																					
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.083	11:31																																																																																																																																					
Air Blank	0.000	11:31																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0830																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

Comments:

2020.09.2
4 07:59:57
-04'00"

DERR

Digitally signed by DERR
Date: 2020.09.23 13:06:29
+08'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-0011113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-0011113</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/23/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>11:45</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/23/2020

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

DERR
Digitally signed
by DERR
Date: 2020.09.23
13:05:57 -0400


2020.09.23
4 07:59:29
-04'00'



INSTRUMENT PROCESSING SHEET

Agency Tampa FHP

S/N 80-001113

Florida Department of
Law Enforcement

Date In 6/16/2020

DI Completion Date 06/18/2020

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By TDG

☒ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☒ Case ☒ Handle
☒ Keyboard ☒ Dry Gas Shelf
☒ Feet ☒ Breath Tube
☒ Ports ☒ Screws Tight

Other Equipment/ Accessories:
☒ Power cord ☐ Printer Cable
☒ Static Bag ☐ 12V DC Cable

Notes: _____

Quality Checks Performed By TDG

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 170
☒ Flow Verification (L/s)
Flow Column # ATP104
32 mm 0.156 (.139 - .169)
36 mm 0.171 (.156 - .190)
53 mm 0.242 (.228 - .278)
103 mm 0.503 (.447 - .547)
☒ Barometric Pressure Check
Gauge ID # 68639
☒ Stability Checks

Flow Calibration Performed By _____

Flow Column # _____
☐ 5L/min - 17mm
☐ 15L/min - 53mm
☐ 30L/min - 103mm
☐ R-Value _____
☐ Post Calibration Verification (L/s)
Flow Column # _____
32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Final Release Date

FDLE
Alcohol
Testing
Program

Digitally signed
by FDLE Alcohol
Testing Program
Date: 2020.07.01
08:07:21 -04'00'

Simulator	Serial #	Lot #/Exp
0.050	MP5092	201905A 05/14/2021
0.080	MP5093	201905B 05/14/2021
0.200	MP5094	201904D 04/30/2021
0.080 DGS	N/A	AG003005 01/30/2022

Maintenance Performed By MH

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☒ Other DG Cal

Temperature Checks Performed By TDG

☒ Lab Temp °C 22.95
External Digital Therm. ID#: 300918
☒ 34°C +/-2 Serial #: MP5092
☒ 34°C +/-2 Serial #: MP5093
☒ 34°C +/-2 Serial #: MP5094

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By MH

Barometric Pressure ID# 28199
Gauge 1016 Instrument 1017
Mouth Alcohol Solution Lot # 2019-B
Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	SD1014
Interferent	SD1015
0.050	MP5092
0.080	MP5093
0.200	MP5094

Attachments

☒ Form 41 ☐ Post-Stability Checks
☒ Stability Checks ☐ Flow Calibration
☒ Calibration Certificate ☐ Form 40
☐ Calibration Adjustment ☐ Other _____

Notes/Suggested Service: Emailed 2020.06.26 08:53:14 -04'00'

Tech Review: Removed "Emailed"

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

2020.06.26 09:19:14 -04'00' SR 2020.07.01 08:04:36 -04'00' Brett Kirkland

Tech Review / Date _____ Admin Review / Date _____

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Time of Inspection: 11:30

Date of Inspection: 06/18/2020

Serial Number: 80-001113

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.081	0.202	0.077
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.081	0.201	0.078
0.000	0.051	0.080	0.202	0.077

Standard Deviations	0.0003	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

2020.06.
26
09:19:46
-04'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY
Signature and Printed Name

06/18/2020
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-00 1113	Lampa FHP	06/17/2020	MG

0.05g/210L 0.047 to 0.053 ☒	0.08g/210L 0.077 to 0.083 ☒	0.20g/210L 0.194 to 0.206 ☒	DGS 0.08g/210L 0.077 to 0.083 ☒																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 06/17/2020 Software: 8100.27 SN 80-001113	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 06/17/2020 Software: 8100.27 SN 80-001113	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 06/17/2020 Software: 8100.27 SN 80-001113	FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 06/17/2020 Software: 8100.27 SN 80-001113																																																																																																																																																
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:49	Control Test	0.050	09:49	Air Blank	0.000	09:50	Control Test	0.050	09:51	Air Blank	0.000	09:51	Control Test	0.049	09:52	Air Blank	0.000	09:53	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	09:54	Control Test	0.081	09:55	Air Blank	0.000	09:55	Control Test	0.080	09:56	Air Blank	0.000	09:56	Control Test	0.080	09:57	Air Blank	0.000	09:58	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td>Control Test</td><td>0.202</td><td>10:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2013</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:06	Control Test	0.202	10:07	Air Blank	0.000	10:07	Control Test	0.201	10:08	Air Blank	0.000	10:09	Control Test	0.201	10:09	Air Blank	0.000	10:10	Control Test Stats			Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	10:11	Control Test	0.078	10:11	Air Blank	0.000	10:12	Control Test	0.078	10:12	Air Blank	0.000	10:13	Control Test	0.078	10:13	Air Blank	0.000	10:13	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.050	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.050	09:51																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test	0.049	09:52																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test	0.081	09:55																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test	0.080	09:56																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test	0.080	09:57																																																																																																																																																	
Air Blank	0.000	09:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.202	10:07																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.201	10:08																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.201	10:09																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2013																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2868																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.078	10:11																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test	0.078	10:12																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test	0.078	10:13																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
065																																																																																																																																																			
Jacob Hunter Operator's Signature	Jacob Hunter Operator's Signature	Jacob Hunter Operator's Signature	Jacob Hunter Operator's Signature																																																																																																																																																

Jacob Hunter
Operator's Signature

Jacob Hunter
Operator's Signature

Jacob Hunter
Operator's Signature

Jacob Hunter
Operator's Signature

Operator's Signature

2020.06.2
6 09:20:41
-04'00'

BK
2020.07.
01
08:06:15
-04'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001113</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/18/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:30</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

06/18/2020

Date

MICHAEL D HAUGHEY,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1

2020.06.
26
09:21:18
-04'00'

2020.07.
01
08:06:38
-04'00'

BK



INSTRUMENT PROCESSING SHEET

Agency FHP TampaS/N 80-001113Florida Department of
Law EnforcementDate In 05/04/2020 DI Completion Date 05/08/2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>MH</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE MAY 28 2020 Alcohol Testing Program	Quality Checks Performed By <u>MH</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>170</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11/12/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG931603 11/12/2021	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>MH</u> ☒ Lab Temp °C <u>22.69</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/- .2 Serial #: <u>MP4863</u> ☒ 34°C +/- .2 Serial #: <u>MP4864</u> ☒ 34°C +/- .2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080	MP4864	201905B 05/14/2021																																																											
0.200	MP5097	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG931603 11/12/2021																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1018</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												
Notes/Suggested Service: <u>E-mailed</u> ☒ APPROVED 05/12/2020	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SR 5/26/20</u> Tech Review / Date <u>Brett Kirkland 5/28/2020</u> Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:36

Date of Inspection: 05/08/2020

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
Standard Deviations	0.0005	0.0004	0.0000	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

05/08/2020
Date

SP
13K
5/28/2020

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001113	FHP Tampa	05/08/2020	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model: 8000 05/08/2020 Software: 8100.27 SN 80-001113 Test 9/210L Time Air Blank 0.000 10:32 Control Test 0.049 10:32 Air Blank 0.000 10:33 Control Test 0.049 10:34 Air Blank 0.000 10:34 Control Test 0.049 10:35 Air Blank 0.000 10:35 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model: 8000 05/08/2020 Software: 8100.27 SN 80-001113 Test 9/210L Time Air Blank 0.000 10:39 Control Test 0.080 10:39 Air Blank 0.000 10:40 Control Test 0.080 10:41 Air Blank 0.000 10:41 Control Test 0.079 10:42 Air Blank 0.000 10:42 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247	FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model: 8000 05/08/2020 Software: 8100.27 SN 80-001113 Test 9/210L Time Air Blank 0.000 10:44 Control Test 0.202 10:45 Air Blank 0.000 10:45 Control Test 0.201 10:46 Air Blank 0.000 10:47 Control Test 0.200 10:47 Air Blank 0.000 10:48 Control Test Stats Average 0.2010 Std Dev 0.0010 Rel Std Dev(%) 0.4975	FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model: 8000 05/08/2020 Software: 8100.27 SN 80-001113 Test 9/210L Time Air Blank 0.000 10:50 Control Test 0.079 10:50 Air Blank 0.000 10:51 Control Test 0.079 10:51 Air Blank 0.000 10:52 Control Test 0.078 10:52 Air Blank 0.000 10:52 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339
Operator's Signature MX	Operator's Signature MX	Operator's Signature MX	Operator's Signature MX

SP
BKH
5/28/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001113, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001113</u>	UNCERTAINTY * $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/08/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>14:36</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/08/2020

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

SP 13K 5/20/2020

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Alain Hernandez on 04/07/2020

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001113

Bill To Address:
FHP Troop C

Ship to Address:
FDLE ATP

Attn: Michael Haughey

4700 Terminal Drive Suite 1

Fort Myers FL 33907

Reason for Return:

Pumps air through breath tube instead of simulator,
thus reads 0.000 on control tests.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Corporal Alain Hernandez

Phone #: 813.558.1800

Email: AlainHernandez-Ginarte@flhsmv.go

ATP Contact Name: Michael Haughey

ATP Email: michaelhaughey@fdle.state.fl.us

0.05 g/200

FL High-Speed
Intoxilyzer - Alcohol Analyzer
Model 8000
04/07/2020
Software: 8100.27

Test	Result	Time
Air Blank	0.000	09:28
Control Test	0.000	09:28
Air Blank	0.000	09:29
Control Test	0.000	09:30
Air Blank	0.000	09:31
Control Test	0.000	09:31
Air Blank	0.000	09:31
Control Test	0.000	09:31
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.00000	

MX
Operator's Signature

0.05 g/200

FL High-Speed
Intoxilyzer - Alcohol Analyzer
Model 8000
04/07/2020
Software: 8100.27

Test	Result	Time
Air Blank	0.000	09:32
Control Test	0.000	09:33
Air Blank	0.000	09:34
Control Test	0.000	09:34
Air Blank	0.000	09:35
Control Test	0.000	09:35
Air Blank	0.000	09:36
Control Test	0.000	09:36
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.0000	

MX
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 08:23

Date of Inspection: 04/07/2020

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BY-PASSED NOT. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

04/07/2020
Date



INSTRUMENT PROCESSING SHEET

Agency Tampa FHP

S/N 80-001113



Florida Department of
Law Enforcement

Date In 01/29/2020

DI Completion Date 2/4/20

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>RAW</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>02</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>171</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
Final Release Date FDLE FEB 06 2020 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1013</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>A5916501</u> <u>06-14-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>DR1279</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>SD1013</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>A5916501</u> <u>06-14-2021</u>	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>SP</u> ☒ Lab Temp °C <u>21.8</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C ±.2 Serial #: <u>MP5088</u> ☒ 34°C ±.2 Serial #: <u>MP5089</u> ☒ 34°C ±.2 Serial #: <u>MP5090</u>																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	<u>SD1012</u>	<u>201905A</u> <u>05-14-2021</u>																																																															
0.080	<u>DR1279</u>	<u>201905B</u> <u>05-14-2021</u>																																																															
0.200	<u>SD1013</u>	<u>201904D</u> <u>04-30-2021</u>																																																															
0.080 DGS	N/A	<u>A5916501</u> <u>06-14-2021</u>																																																															
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number	Lot Number	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	<u>MP5086</u>																																																																
Interferent	<u>MP5087</u>																																																																
0.050	<u>MP5088</u>																																																																
0.080	<u>MP5089</u>																																																																
0.200	<u>MP5090</u>																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>QPM 2/4/20</u> Tech Review / Date <u>Brett Kirklund 202/4/2020</u> Admin Review / Date																																																															



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001113, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001113</u>	UNCERTAINTY * $\pm$	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/04/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/04/2020

Date

Shayla Platt

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 January 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Adam YSK
2/4/2020

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:42

Date of Inspection: 02/04/2020

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.081	0.201	0.078
0.000	0.049	0.081	0.202	0.079
0.000	0.049	0.081	0.201	0.078
0.000	0.050	0.081	0.201	0.078
0.000	0.049	0.081	0.201	0.078
0.000	0.049	0.081	0.202	0.078
0.000	0.049	0.081	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.049	0.080	0.203	0.078
0.000	0.050	0.081	0.203	0.078

Standard Deviations	0.0004	0.0003	0.0007	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

JPM
13K
2/4/2020

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

02/04/2020
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001113
01/30/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.049	13:54
Air Blank	0.000	13:54
Control Test	0.049	13:55
Air Blank	0.000	13:55
Control Test	0.049	13:56
Air Blank	0.000	13:57
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SS

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001113
01/30/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.080	13:59
Air Blank	0.000	13:59
Control Test	0.080	14:00
Air Blank	0.000	14:00
Control Test	0.080	14:01
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

SS

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001113
01/30/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:05
Control Test	0.200	14:06
Air Blank	0.000	14:06
Control Test	0.200	14:07
Air Blank	0.000	14:07
Control Test	0.199	14:08
Air Blank	0.000	14:09
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SS

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001113
01/30/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:12
Control Test	0.079	14:12
Air Blank	0.000	14:13
Control Test	0.079	14:13
Air Blank	0.000	14:14
Control Test	0.078	14:14
Air Blank	0.000	14:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry

SS

Operator's Signature

SSM
13K
2/4/2020

Return Material Authorization

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Alain Ginarte on 8/3/19

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001113

Bill To Address:

FHP Tampa

Ship to Address:

FDLE-ATP Tallahassee

Reason for Return:

Ambient Fails and agency reports the dry gas regulator is leaking

(See attached)

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Alain Ginarte

Phone #: 386-747-6641

Email: Hernandez-Ginarte@flhsmv.gov

ATP Contact Name: Shayla Platt

ATP Email: shaylaplatt@fdle.state.fl.us

gsm
13/2
2/4/2020



INSTRUMENT PROCESSING SHEET

Agency FHP TampaS/N 80-001113Florida Department of
Law EnforcementDate In 7/26/2019

DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake Performed By <u>DP</u>		Quality Checks Performed By <u>SP</u>		Flow Calibration Performed By _____																																																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By _____ ☐ Lab Temp °C External Digital Therm. ID#: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____ ☐ 34°C +/- .2 Serial #: _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Final Release Date _____																																																																			
Calibration Adjustment Performed By _____		Department Inspection Performed By _____																																																																	
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th><th>Lot Number</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table> Attachments <table border="1"><tr><td>☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____</td></tr></table> ☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____				Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____
Simulator	Serial Number	Lot Number	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number	Lot Number	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000																																																																			
Interferent																																																																			
0.050																																																																			
0.080																																																																			
0.200																																																																			
☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																																																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____																																																																			

DP
B/L
2/4/2020

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:45

Date of Inspection: 08/01/2019

Serial Number: 80-001113
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

PDM
 TBK
 2/4/2000

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

08/01/2019
Date

INTOXILYZER 8000
Instrument Initialization
14:48 08/01/2019

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001113
08/01/2019
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:07
Control Test	INT*	15:08
Air Blank	AMB**	15:08
Air Blank	AMB**	15:09

*Interferent Detect
**Ambient Fail

[Signature]

Operator's Signature

[Signature]
13K
2/4/2020