



INSTRUMENT PROCESSING SHEET

Agency Deland PDS/N 80-001114Florida Department of
Law EnforcementDate In 11-06-2020 DI Completion Date 11-13-2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ 	Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>210</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG011102 04-20-2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>20.77</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5086</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP5088	201905A 05-14-2021																																																												
0.080	MP5089	201905B 05-14-2021																																																												
0.200	MP5090	201904D 04-30-2021																																																												
0.080 DGS	N/A	AG011102 04-20-2022																																																												
Final Release Date FDLE Alcohol Testing Program	Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.17 06:50:28 -05'00'																																																													
Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge <u>1020</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20420</td> <td>09-22-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06-05-2021</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A</td> <td>05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5091</td> <td>201905B</td> <td>05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D</td> <td>04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102</td> <td>04-20-2022</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20420	09-22-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	08819080A1	06-05-2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	05-14-2021	0.080	MP5091	201905B	05-14-2021	0.200	MP5090	201904D	04-30-2021	0.080 DGS	N/A	AG011102	04-20-2022	Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2020.11.16 <u>Michael D. Haughey</u> 12:11:09 -05'00' Tech Review / Date 2020.11.1 <u>CR</u> 7:06:49:38 Admin Review / Date		Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	MP5091	N/A	N/A																																																											
0.040	MP5082	20060	02-10-2022																																																											
0.100	MP5083	20420	09-22-2022																																																											
0.200	MP5084	20160	03-18-2022																																																											
0.300	MP5085	20030	01-21-2022																																																											
0.080 DGS	N/A	08819080A1	06-05-2021																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	MP5088	201905A	05-14-2021																																																											
0.080	MP5091	201905B	05-14-2021																																																											
0.200	MP5090	201904D	04-30-2021																																																											
0.080 DGS	N/A	AG011102	04-20-2022																																																											
Simulator	Serial Number																																																													
0.000	G11621																																																													
Interferent	MP5087																																																													
0.050	MP5088																																																													
0.080	MP5086																																																													
0.200	MP5090																																																													
Notes/Suggested Service: <u>Wet 0.080 stability check not nominal, checked temp and tightened seal and reran checked simulator with a different instrument. Value still not nominal, performed opt bench cal adjust. Post wet 0.080 still not nominal, switched simulator from MP5089 to MP5091 and 0.080 post stability was nominal.</u>																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND P.D.

Time of Inspection: 10:48

Date of Inspection: 11/13/2020

Serial Number: 80-001114

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.048	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.048	0.080	0.200	0.079
0.000	0.048	0.079	0.200	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.078

Standard Deviations	0.0005	0.0005	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name


11/13/2020
Date


 2020.1
 1.17
 06:49:1
 3
 -05'00'

Stability Checks

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:41
Control Test	0.049	14:42
Air Blank	0.000	14:43
Control Test	0.048	14:43
Air Blank	0.000	14:44
Control Test	0.048	14:44
Air Blank	0.000	14:45
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	




Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:47
Control Test	0.061	14:48
Air Blank	0.000	14:49
Control Test	0.062	14:49
Air Blank	0.000	14:50
Control Test	0.062	14:50
Air Blank	0.000	14:51
Control Test Stats		
Average	0.0617	
Std Dev	0.0006	
Rel Std Dev(%)	0.9362	

wet 1



Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:56
Control Test	0.063	14:57
Air Blank	0.000	14:57
Control Test	0.062	14:58
Air Blank	0.000	14:58
Control Test	0.062	14:59
Air Blank	0.000	15:00
Control Test Stats		
Average	0.0623	
Std Dev	0.0006	
Rel Std Dev(%)	0.9262	

wet 2



Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:01
Control Test	0.200	15:02
Air Blank	0.000	15:02
Control Test	0.200	15:03
Air Blank	0.000	15:04
Control Test	0.200	15:04
Air Blank	0.000	15:05
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:06
Control Test	0.079	15:06
Air Blank	0.000	15:07
Control Test	0.079	15:07
Air Blank	0.000	15:07
Control Test	0.079	15:08
Air Blank	0.000	15:08
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry



Operator's Signature



Operator's Signature

2020.11
17
06:48:5
2-05'00'

MX



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001114, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001114	UNCERTAINTY* $\pm$	
Owning Agency:	DELAND P.D.	0.050 g/ 210 L	0.004
Calibration Date:	11/13/2020	0.080 g/ 210 L	0.005
Calibration Time:	10:48	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/13/2020

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.11.
17
06:48:30
-05'00'

Optical Bench Cal Adj.

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SM 80-001114
11/06/2020 15:38:26

Auto Calibration
Max Power Res Value = 21
Auto Range Res Value = 5

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11576, Sum Io = 14284

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1960 (-0.0090)
Sample #2 = 0.1740 (0.0410)
Sample #3 = 0.1690 (0.0890)
Sample #4 = 0.1600 (0.1070)
Avg % Abs = 0.1677 (0.0790)
STD DEV = 0.0071 (0.0341)
REL STD DEV = 4.231 (43.187)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1740 (-0.0040)
Sample #2 = 0.1630 (0.0300)
Sample #3 = 0.1630 (0.0400)
Sample #4 = 0.1730 (0.0440)
Avg % Abs = 0.1663 (0.0380)
STD DEV = 0.0058 (0.0072)
REL STD DEV = 3.471 (18.577)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11558, Sum Io = 14278

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.9060 (-0.0130)
Sample #2 = 0.8850 (0.0060)
Sample #3 = 0.8950 (0.0300)
Sample #4 = 0.8950 (0.0390)
Avg % Abs = 0.8917 (0.0250)
STD DEV = 0.0058 (0.0171)
REL STD DEV = 0.647 (68.235)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6180 (0.0060)
Sample #2 = 1.6050 (0.0190)
Sample #3 = 1.6130 (0.0320)
Sample #4 = 1.5980 (0.0300)
Avg % Abs = 1.6053 (0.0270)
STD DEV = 0.0075 (0.0070)
REL STD DEV = 0.468 (25.926)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11550, Sum Io = 14274

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9800 (-0.0150)
Sample #2 = 1.9670 (-0.0050)
Sample #3 = 1.9810 (-0.0070)
Sample #4 = 1.9850 (-0.0030)
Avg % Abs = 1.9777 (-0.0050)
STD DEV = 0.0095 (0.0020)
REL STD DEV = 0.478 (40.000)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6850 (-0.0050)
Sample #2 = 3.6770 (0.0030)
Sample #3 = 3.6530 (0.0270)
Sample #4 = 3.6760 (0.0160)
Avg % Abs = 3.6687 (0.0153)
STD DEV = 0.0136 (0.0120)
REL STD DEV = 0.370 (78.351)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11551, Sum Io = 14272

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7690 (-0.0120)
Sample #2 = 3.7590 (0.0150)
Sample #3 = 3.7600 (0.0170)
Sample #4 = 3.7500 (0.0150)
Avg % Abs = 3.7563 (0.0157)
STD DEV = 0.0055 (0.0012)
REL STD DEV = 0.147 (7.370)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.0280 (-0.0080)
Sample #2 = 7.0340 (0.0070)
Sample #3 = 7.0000 (0.0310)
Sample #4 = 7.0340 (0.0180)
Avg % Abs = 7.0127 (0.0187)
STD DEV = 0.0186 (0.0120)
REL STD DEV = 0.265 (64.360)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 11549, Sum Io = 14270

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.4400 (-0.0140)
Sample #2 = 5.4150 (-0.0040)
Sample #3 = 5.4120 (0.0050)
Sample #4 = 5.4230 (-0.0120)
Avg % Abs = 5.4167 (-0.0037)
STD DEV = 0.0057 (0.0085)
REL STD DEV = 0.105 (231.952)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0970 (-0.0130)
Sample #2 = 10.0880 (0.0070)
Sample #3 = 10.0910 (0.0140)
Sample #4 = 10.0890 (-0.0050)
Avg % Abs = 10.0893 (0.0053)
STD DEV = 0.0015 (0.0096)
REL STD DEV = 0.015 (180.169)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.168
Std Dev = 0.01 Rel Std Dev = 4.23
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.892
Std Dev = 0.01 Rel Std Dev = 0.65
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.978
Std Dev = 0.01 Rel Std Dev = 0.48
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.756
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.417
Std Dev = 0.01 Rel Std Dev = 0.10
Zero Order Coef = -399.78
First Order Coef = 2529.75
Second Order Coef = 32.93
Standard Deviation = 32.495480

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.166
Std Dev = 0.01 Rel Std Dev = 3.47
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.605
Std Dev = 0.01 Rel Std Dev = 0.47
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.669
Std Dev = 0.01 Rel Std Dev = 0.37
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.013
Std Dev = 0.02 Rel Std Dev = 0.26
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.089
Std Dev = 0.00 Rel Std Dev = 0.02
Zero Order Coef = -207.00
First Order Coef = 1294.43
Second Order Coef = 13.95
Standard Deviation = 24.239378

MX

2020.11.

17

06:48:07

-05'00"

Optical Bench Cal Adj.

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0005
0.040	0.040	0.0005
0.100	0.099	0.0006
0.200	0.201	-0.0009
0.300	0.300	0.0003

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	-0.0000
0.100	0.099	0.0007
0.200	0.201	-0.0007
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 2792.00
 Sample #2 = 2824.00
 Sample #3 = 2782.00
 Sample #4 = 2829.00
 Average Result = 2811.6667
 STD DEV = 25.8134
 REL STD DEV = 0.918

***** CHANNEL 2

Sample #1 = 3186.00
 Sample #2 = 3166.00
 Sample #3 = 3194.00
 Sample #4 = 3171.00
 Average Result = 3177.0000
 STD DEV = 14.9332
 REL STD DEV = 0.470

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1017
 3 um H2O Adjust (mg/l*10,000) = 998
 9 um H2O Adjust (mg/l*10,000) = 632

*** AUTO CAL PASS

MX

2020.1
 1.17
 06:47:4
 3
 -05'00'

Post stability checks

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:24
Control Test	0.049	16:24
Air Blank	0.000	16:25
Control Test	0.049	16:26
Air Blank	0.000	16:26
Control Test	0.048	16:27
Air Blank	0.000	16:27
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:29
Control Test	0.062	16:29
Air Blank	0.000	16:30
Control Test	0.061	16:30
Air Blank	0.000	16:31
Control Test	0.061	16:32
Air Blank	0.000	16:32
Control Test Stats		
Average	0.0613	
Std Dev	0.0006	
Rel Std Dev(%)	0.9413	

wet 1

MP5089

Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:33
Control Test	0.199	16:34
Air Blank	0.000	16:34
Control Test	0.198	16:35
Air Blank	0.000	16:36
Control Test	0.199	16:36
Air Blank	0.000	16:37
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Control Test	0.079	16:39
Air Blank	0.000	16:40
Control Test	0.079	16:40
Air Blank	0.000	16:41
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Dry

Operator's Signature

DELAND P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001114
11/09/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:54
Control Test	0.078	15:54
Air Blank	0.000	15:55
Control Test	0.079	15:56
Air Blank	0.000	15:56
Control Test	0.079	15:57
Air Blank	0.000	15:57
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet 2

MP5091

Operator's Signature

2020.11.
17
06:47:23
-05'00'