



INSTRUMENT PROCESSING SHEET

Agency Pasco CO SOS/N 80-001117

Florida Department of Law Enforcement

Date In 08-12-2020DI Completion Date 8/19/20 Ship P/U H/D CMI EE

Intake Performed By <u>KAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>212</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.183</u> (.139 - .169) 36 mm <u>0.199</u> (.156 - .190) 53 mm <u>0.273</u> (.228 - .278) 103 mm <u>0.546</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By <u>SP</u> Flow Column # <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>214</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547)
---	--	--

Final Release Date Digitally signed by FDLE Alcohol Testing Program Date: 2020.08.20 08:10:42 -04'00'	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1021</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>DR1275</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>SD1019</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1021	201905A 05-14-2021	0.080	DR1275	201905B 05-14-2021	0.200	SD1019	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____
Simulator	Serial #	Lot #/Exp															
0.050	SD1021	201905A 05-14-2021															
0.080	DR1275	201905B 05-14-2021															
0.200	SD1019	201904D 04-30-2021															
0.080 DGS	N/A	AG931603 11-12-2021															

Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1009</u> ID # <u>26932</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>200600</td> <td>2-10-22</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>4-6-22</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>3-8-22</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>1-21-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03519050A4</td> <td>4-5-21</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	200600	2-10-22	0.100	MP5083	20190	4-6-22	0.200	MP5084	20160	3-8-22	0.300	MP5085	20030	1-21-22	0.080 DGS	N/A	03519050A4	4-5-21	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1010</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP50869 SP</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP50869 SP	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																						
0.000	MP5091	N/A	N/A																																						
0.040	MP5082	200600	2-10-22																																						
0.100	MP5083	20190	4-6-22																																						
0.200	MP5084	20160	3-8-22																																						
0.300	MP5085	20030	1-21-22																																						
0.080 DGS	N/A	03519050A4	4-5-21																																						
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP50869 SP																																								
0.200	MP5090																																								

<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A</td> <td>5-14-21</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B</td> <td>5-14-21</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D</td> <td>4-30-21</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603</td> <td>11-12-21</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	5-14-21	0.080	MP5089	201905B	5-14-21	0.200	MP5090	201904D	4-30-21	0.080 DGS	N/A	AG931603	11-12-21	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
Simulator	Serial Number	Lot Number	Expiration																			
0.050	MP5088	201905A	5-14-21																			
0.080	MP5089	201905B	5-14-21																			
0.200	MP5090	201904D	4-30-21																			
0.080 DGS	N/A	AG931603	11-12-21																			

Notes/Suggested Service: _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
David Eliezer Reyes Rivera Digitally signed by David Eliezer Reyes Rivera Date: 2020.08.19 16:15:16 -0400	Brett Kirkland 2020.08.20 08:07:52 -04'00' Tech Review / Date Admin Review / Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2729 Fort Knox Blvd.
 Bldg. 2, Suite 1300
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001117, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001117</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/19/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>15:16</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/19/2020 Date
Shayla Platt
 SHAYLA D PLATT,
 Department Inspector

FDLE/ATP Form 69 April 2020
 Issuing Authority: Alcohol Testing Program
 Service • Integrity • Respect • Quality
 Page 1 of 1

DERR

2020.08.20
 08:08:14
 -04:00
 BK

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 15:16

Date of Inspection: 08/19/2020

Serial Number: 80-001117
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.048	0.079	0.201	0.078
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.079

Standard Deviations	0.0004	0.0003	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

08/19/2020
Date

DERR
Digitally signed by DERR
Date: 2020.08.19 16:13:18 -0400

BK 2020.08.20 08:08:41 -0400

stability checks

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/18/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:26
Control Test	0.046	12:27
Air Blank	0.000	12:27
Control Test	0.047	12:28
Air Blank	0.000	12:28
Control Test	0.047	12:29
Air Blank	0.000	12:30
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/18/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:31
Control Test	0.077	12:32
Air Blank	0.000	12:33
Control Test	0.078	12:33
Air Blank	0.000	12:34
Control Test	0.077	12:35
Air Blank	0.000	12:35
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	


PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/18/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:40
Control Test	0.200	12:41
Air Blank	0.000	12:41
Control Test	0.200	12:42
Air Blank	0.000	12:43
Control Test	0.200	12:43
Air Blank	0.000	12:44
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


wet



Operator's Signature



Operator's Signature



Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/18/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:46
Control Test	0.080	12:47
Air Blank	0.000	12:47
Control Test	0.078	12:47
Air Blank	0.000	12:48
Control Test	0.079	12:48
Air Blank	0.000	12:49
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Dry



Operator's Signature

BK 2020.08.20
08:08:59
-0400'

DERR Digitally signed
by DERR
Date: 2020.08.19
16:12:24 -0400'

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

PRSCO COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
08/19/2020
SN 86-001117
08:57:45

Auto Calibration
Max Power Res Value = 107
Auto Range Res Value = 80

CAL ADJUSTMENT
#80-00117 SP

DERR
2020.08.20
08:09:17
BK
0400

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.040 0.0002
0.100 0.099 0.0007
0.200 0.201 -0.0008
0.300 0.300 0.0003

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0004
0.200 0.200 -0.0005
0.300 0.300 0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l ****
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 2985.00
Sample #2 = 3004.00
Sample #3 = 2992.00
Sample #4 = 3002.00
Average Result = 2999.3333
STD DEV = 6.4291
REL STD DEV = 0.214

**** CHANNEL 2
Sample #1 = 3361.00
Sample #2 = 3384.00
Sample #3 = 3389.00
Sample #4 = 3407.00
Average Result = 3393.3333
STD DEV = 12.0968
REL STD DEV = 0.356

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1009
3 um H2O Adjust (mg/l*10,000) = 810
9 um H2O Adjust (mg/l*10,000) = 416
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.040 0.0002
0.100 0.099 0.0007
0.200 0.201 -0.0008
0.300 0.300 0.0003

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0004
0.200 0.200 -0.0005
0.300 0.300 0.0002

**** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.113
Std Dev = 0.01 Rel Std Dev = 8.01
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.840
Std Dev = 0.01 Rel Std Dev = 0.79
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.908
Std Dev = 0.02 Rel Std Dev = 1.09
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.647
Std Dev = 0.02 Rel Std Dev = 0.45
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.257
Std Dev = 0.01 Rel Std Dev = 0.20
Zero Order Coef = -270.70
First Order Coef = 2539.31
Second Order Coef = 43.15
Standard Deviation = 28.224184

**** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.124
Std Dev = 0.01 Rel Std Dev = 5.65
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.549
Std Dev = 0.00 Rel Std Dev = 0.17
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.604
Std Dev = 0.01 Rel Std Dev = 0.35
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.869
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.877
Std Dev = 0.02 Rel Std Dev = 0.19
Zero Order Coef = -152.79
First Order Coef = 1300.12
Second Order Coef = 16.28
Standard Deviation = 15.559872

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9310 (-0.0030)
Sample #2 = 6.8640 (0.0540)
Sample #3 = 6.8720 (0.0670)
Sample #4 = 6.8700 (0.0690)
Avg % Abs = 6.8687 (0.0633)
STD DEV = 0.0042 (0.0081)
REL STD DEV = 0.061 (12.860)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3000 (-0.0100)
Sample #2 = 5.2580 (0.0370)
Sample #3 = 5.2460 (0.0370)
Sample #4 = 5.2670 (0.0470)
Avg % Abs = 5.2570 (0.0403)
STD DEV = 0.0105 (0.0058)
REL STD DEV = 0.200 (14.314)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.9540 (-0.0030)
Sample #2 = 9.8980 (0.0880)
Sample #3 = 9.8610 (0.1030)
Sample #4 = 9.8720 (0.0920)
Avg % Abs = 9.8770 (0.0943)
STD DEV = 0.0190 (0.0078)
REL STD DEV = 0.192 (8.234)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9290 (-0.0170)
Sample #2 = 1.9260 (0.0060)
Sample #3 = 1.9120 (0.0020)
Sample #4 = 1.8850 (0.0350)
Avg % Abs = 1.9077 (0.0143)
STD DEV = 0.0208 (0.0180)
REL STD DEV = 1.092 (125.646)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6140 (-0.0020)
Sample #2 = 3.6160 (0.0150)
Sample #3 = 3.5940 (0.0390)
Sample #4 = 3.5990 (0.0340)
Avg % Abs = 3.6037 (0.0293)
STD DEV = 0.0127 (0.0127)
REL STD DEV = 0.351 (43.167)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9290 (-0.0170)
Sample #2 = 1.9260 (0.0060)
Sample #3 = 1.9120 (0.0020)
Sample #4 = 1.8850 (0.0350)
Avg % Abs = 1.9077 (0.0143)
STD DEV = 0.0208 (0.0180)
REL STD DEV = 1.092 (125.646)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6140 (-0.0020)
Sample #2 = 3.6160 (0.0150)
Sample #3 = 3.5940 (0.0390)
Sample #4 = 3.5990 (0.0340)
Avg % Abs = 3.6037 (0.0293)
STD DEV = 0.0127 (0.0127)
REL STD DEV = 0.351 (43.167)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9290 (-0.0170)
Sample #2 = 1.9260 (0.0060)
Sample #3 = 1.9120 (0.0020)
Sample #4 = 1.8850 (0.0350)
Avg % Abs = 1.9077 (0.0143)
STD DEV = 0.0208 (0.0180)
REL STD DEV = 1.092 (125.646)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6140 (-0.0020)
Sample #2 = 3.6160 (0.0150)
Sample #3 = 3.5940 (0.0390)
Sample #4 = 3.5990 (0.0340)
Avg % Abs = 3.6037 (0.0293)
STD DEV = 0.0127 (0.0127)
REL STD DEV = 0.351 (43.167)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0080)
Sample #2 = 3.6310 (0.0300)
Sample #3 = 3.6450 (0.0380)
Sample #4 = 3.6540 (0.0270)
Avg % Abs = 3.6467 (0.0317)
STD DEV = 0.0166 (0.0057)
REL STD DEV = 0.454 (17.957)

Post Cal Adjustment Stability Checks # 80-001117

PRSCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/19/2020
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:19
Control Test	0.079	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Control Test	0.079	11:22
Air Blank	0.000	
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DAS

SP

Operator's Signature

DERR
Digitally signed
by DERR
Date: 2020.08.20
16:07:08 -0400

BK
082936
0400

PRSCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/19/2020
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:41
Control Test	0.200	11:41
Air Blank	0.000	11:42
Control Test	0.200	11:43
Air Blank	0.000	11:43
Control Test	0.199	11:44
Air Blank	0.000	11:45
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

SP

Operator's Signature

PRSCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/19/2020
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:36
Control Test	0.079	11:37
Air Blank	0.000	11:38
Control Test	0.079	11:39
Air Blank	0.000	11:39
Control Test	0.079	11:40
Air Blank	0.000	
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PRSCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/19/2020
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	11:31
Control Test	0.048	11:32
Air Blank	0.000	11:32
Control Test	0.048	11:33
Air Blank	0.000	11:33
Control Test	0.048	11:34
Air Blank	0.000	11:35
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

FLOW CAL ADJUSTMENT

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001117
08/19/2020
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 7.070
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.871
3: Rate (Liters/min) = 30
SQRT(Diff) = 21.023
Dependent Data Scale Factor = 10000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 692
Rounded Intercept = -693297
Correlation = 0.99900

SR

DERR
Digitally signed
by DERR
Date: 2020.08.19
16:00:58 -04'00'

BK
2020.08.20
08:09:58
-04'00'