FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency South Daytona PD S/N 80-001246

Date In 11-13-2020 DI Completion Date 11-16-2020 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By IS

- ☑ Annual
- ☐ Registration
- ☑ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☑ Static Bag ☐ 12V DC Cable

Notes: _____

Quality Checks Performed By IS

- ☑ Breath Tube Screen
- ☑ Replace External O-Rings
- ☑ Instrument Set Up Verified
- ☑ R-Value 197
- ☑ Flow Verification (L/s)

Flow Column # ATP-105

32 mm	0.148	(.139 - .169)
36 mm	0.164	(.156 - .190)
53 mm	0.242	(.228 - .278)
103 mm	0.503	(.447 - .547)

- ☑ Barometric Pressure Check

Gauge ID # 30793

- ☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A
		05-14-2021
0.080	MP5086	201905B
		05-14-2021
0.200	MP5090	201904D
		04-30-2021
0.080 DGS	N/A	AG011102
		04-20-2022

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By IS

- ☑ Lab Temp °C 20.82

External Digital Therm. ID#: 300503

- ☑ 34°C +-.2 Serial #: MP5088
- ☑ 34°C +-.2 Serial #: MP5086
- ☑ 34°C +-.2 Serial #: MP5090

Final Release Date

FDLE Alcohol Testing Program

Digitally signed by FDLE Alcohol Testing Program
Date: 2020.11.23 12:58:26 -05'00'

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By IS

Barometric Pressure ID# 30793

Gauge 1021 Instrument 1021

Mouth Alcohol Solution Lot # 2019-B

Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	G11621
Interferent	MP5087
0.050	MP5088
0.080	MP5086
0.200	MP5090

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☑ Other Form 51

Notes/Suggested Service: _____

- ☑ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☑ **Return to/Place into Evidentiary Use**
- ☐ **Remove Out of Evidentiary Use**
- ☑ **Conduct an Agency Inspection Before Evidentiary Use**

Michael D Haughey 2020.11.19 14:31:21 -05'00' — Tech Review / Date

2020.11.23 12:57:40 -05'00' — Admin Review / Date

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001246 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001246
Owning Agency:	SOUTH DAYTONA P.D.
Calibration Date:	11/16/2020
Calibration Time:	14:11

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

11/16/2020
Date

Israel Soto
ISRAEL SOTO,
Department Inspector

2020.11.23 12:57:21 -05'00'

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SOUTH DAYTONA P.D. Serial Number: 80-001246
Time of Inspection: 14:11 Date of Inspection: 11/16/2020 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.050	0.081	0.204	0.080
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.081	0.204	0.079
0.000	0.050	0.082	0.204	0.080
0.000	0.050	0.081	0.204	0.079
0.000	0.050	0.081	0.204	0.079
0.000	0.050	0.082	0.204	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.050	0.081	0.204	0.079
0.000	0.050	0.081	0.204	0.079

Standard Deviations	0.0000	0.0004	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

2020.11.23 12:56:57 -05'00'

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

11/16/2020
Date

FDLE/ATP Form 41 –Revised August 2005

Stability Checks

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:08
Control Test	0.049	12:08
Air Blank	0.000	12:09
Control Test	0.049	12:09
Air Blank	0.000	12:10
Control Test	0.049	12:11
Air Blank	0.000	12:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:12
Control Test	0.081	12:13
Air Blank	0.000	12:14
Control Test	0.081	12:14
Air Blank	0.000	12:15
Control Test	0.081	12:16
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.205	12:19
Air Blank	0.000	12:19
Control Test	0.203	12:20
Air Blank	0.000	12:21
Control Test	0.203	12:21
Air Blank	0.000	12:22
Control Test Stats		
Average	0.2037	
Std Dev	0.0012	
Rel Std Dev(%)	0.5670	

Operator's Signature

wet

SOUTH DAYTONA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001246
11/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:23
Control Test	0.083	12:23
Air Blank	0.000	12:24
Control Test	0.080	12:24
Air Blank	0.000	12:25
Control Test	0.081	12:25
Air Blank	0.000	12:25
Control Test Stats		
Average	0.0813	
Std Dev	0.0015	
Rel Std Dev(%)	1.8781	

Dry

Operator's Signature

MK

2020.11.23 12:56:34 -05'00'

Return Material Authorization

Ship to: ☑ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Anthony Labombard on 9/2/2020

Items Returned: Instrument ☑ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001246

Bill To Address:	Ship to Address:
South Daytona PD	FDLE-Tallahassee

Reason for Return:

Flow Sensor Replacement R- value @ 98

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to $_____.

☑ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anthony Labombard

Phone #: (386) 322-6180 ext. 33521 Email: alabombard@southdaytona.org

ATP Contact Name: _____ ATP Email: _____

2020.11.23 12:56:12 -05'00'

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency South Daytona Police Department S/N 80-001246

Date In 04/09/2020 DI Completion Date _____ ☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Intake Performed By RAW

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case ☑ Handle
- ☑ Keyboard ☑ Dry Gas Shelf
- ☑ Feet ☑ Breath Tube
- ☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

Quality Checks Performed By SP

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 98
- ☐ Flow Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 30793

- ☐ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050		
0.080		
0.200		
0.080 DGS	N/A	

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By SP

- ☒ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☒ Breath Tube Replacement SP
- ☐ Other _____

Temperature Checks Performed By _____

- ☐ Lab Temp °C _____

External Digital Therm. ID#: _____

- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____
- ☐ 34°C +-.2 Serial #: _____

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By _____

Barometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot # _____

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- ☐ Form 41
- ☐ Stability Checks
- ☐ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: Sending to Repair for flow sensor replacement.

SP

- ☐ Instrument Complies with Chapter 11D-8, FAC
- ☒ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☐ Return to/Place into Evidentiary Use
- ☒ Remain Out of Evidentiary Use
- ☐ Conduct an Agency Inspection Before Evidentiary Use

MK

Tech Review / Date _____ Admin Review / Date _____

2020.11.23
12:55:46 -05'00'

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SOUTH DAYTONA P.D.
Time of Inspection: 10:53 Date of Inspection: 09/01/2020 Serial Number: 80-001246
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

MK

N/A Compliance not determined

2020.11.23 12:55:26 -05'00'

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto ISRAEL SOTO
Signature and Printed Name

09/01/2020
Date

FDLE/ATP Form 40 – March 2004