



INSTRUMENT PROCESSING SHEET

Agency Baker ~~CO~~ Sherriffs County SO SP S/N 80-001287

Florida Department of Law Enforcement Date In 7/28/2020 DI Completion Date 8-19-20  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>KAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>1.95</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.144</u> (.139 -.169) 36 mm <u>0.160</u> (.156 -.190) 53 mm <u>0.226</u> (.228 -.278) 103 mm <u>0.480</u> (.447 -.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By <u>SP</u> Flow Column # <u>ATP102</u> <input type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>190</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.144</u> (.139 -.169) 36 mm <u>.164</u> (.156 -.190) 53 mm <u>.238</u> (.228 -.278) 103 mm <u>.507</u> (.447 -.547)																																	
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.08.20 08:17:11 -04'00'		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____		<b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.6</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>																																	
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1009</u> ID # <u>26932</u>		<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1010</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5091</td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td>MP5082</td><td>20060</td><td>2-10-22</td></tr> <tr><td>0.100</td><td>MP5083</td><td>20190</td><td>4-6-22</td></tr> <tr><td>0.200</td><td>MP5084</td><td>20160</td><td>3-8-22</td></tr> <tr><td>0.300</td><td>MP5085</td><td>20030</td><td>1-21-22</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>03519050A4</td><td>4-5-21</td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	2-10-22	0.100	MP5083	20190	4-6-22	0.200	MP5084	20160	3-8-22	0.300	MP5085	20030	1-21-22	0.080 DGS	N/A	03519050A4	4-5-21				
Simulator	Serial Number	Lot Number	Expiration																																		
0.000	MP5091	N/A	N/A																																		
0.040	MP5082	20060	2-10-22																																		
0.100	MP5083	20190	4-6-22																																		
0.200	MP5084	20160	3-8-22																																		
0.300	MP5085	20030	1-21-22																																		
0.080 DGS	N/A	03519050A4	4-5-21																																		
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td>MP5088</td><td>201905A</td><td>5-14-21</td></tr> <tr><td>0.080</td><td>MP5089</td><td>201905B</td><td>5-14-21</td></tr> <tr><td>0.200</td><td>MP5090</td><td>201904D</td><td>4-30-21</td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td>AG931603</td><td>11-12-21</td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	5-14-21	0.080	MP5089	201905B	5-14-21	0.200	MP5090	201904D	4-30-21	0.080 DGS	N/A	AG931603	11-12-21	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP5086</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5089</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																		
0.050	MP5088	201905A	5-14-21																																		
0.080	MP5089	201905B	5-14-21																																		
0.200	MP5090	201904D	4-30-21																																		
0.080 DGS	N/A	AG931603	11-12-21																																		
Simulator	Serial Number																																				
0.000	MP5086																																				
Interferent	MP5087																																				
0.050	MP5088																																				
0.080	MP5089																																				
0.200	MP5090																																				
Notes/Suggested Service: _____		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																	
		Digitally signed by David Eliezer Reyes Rivera David Eliezer Reyes Rivera Reyes Rivera Date: 2020.08.19 16:36:53 -04'00'		2020.08.20 08:14:10 -04'00'																																	
		Tech Review / Date		Admin Review / Date																																	



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001287, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001287</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/19/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>15:02</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/19/2020

Date

Shayla Platt  
SHAYLAD PLATT,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Digitally signed  
by DERR  
Date: 2020.08.19  
16:36:03 -04'00'

**DERR**

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 15:02

Date of Inspection: 08/19/2020

Serial Number: 80-001287  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.050	0.081	0.202	0.078
0.000	0.049	0.081	0.202	0.078
Standard Deviations	0.0005	0.0004	0.0000	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

08/19/2020  
Date

Digitally signed  
by DERR  
Date: 2020.08.19  
16:35:22 -0400

# Stability Checks

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
07/31/2020  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.048	13:54
Air Blank	0.000	13:54
Control Test	0.046	13:55
Air Blank	0.000	13:56
Control Test	0.046	13:56
Air Blank	0.000	13:57
Control Test Stats		
Average	0.0467	
Std Dev	0.0012	
Rel Std Dev(%)	2.4744	

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
07/31/2020  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	13:59
Control Test	0.077	13:59
Air Blank	0.000	14:00
Control Test	0.076	14:01
Air Blank	0.000	14:01
Control Test	0.077	14:02
Air Blank	0.000	14:03
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

wet

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
07/31/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:04
Control Test	0.200	14:05
Air Blank	0.000	14:06
Control Test	0.198	14:06
Air Blank	0.000	14:07
Control Test	0.197	14:08
Air Blank	0.000	14:08
Control Test Stats		
Average	0.1983	
Std Dev	0.0015	
Rel Std Dev(%)	0.7702	

  
-----  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
07/31/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:09
Control Test	0.077	14:10
Air Blank	0.000	14:10
Control Test	0.078	14:11
Air Blank	0.000	14:11
Control Test	0.078	14:11
Air Blank	0.000	14:12
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

DRY

  
-----  
Operator's Signature

BK  
2020.08.20  
08:15:08  
-04'00'  
Digitally signed  
by DERR  
Date: 2020.08.19  
16:34:36 -04'00'

DERR

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	
0.000	0.000	-0.0002
0.040	0.040	0.0001
0.100	0.100	0.0004
0.200	0.201	-0.0005
0.300	0.300	0.0002

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<<< CHANNEL 1 >>>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.064  
 Std Dev = 0.00 Rel Std Dev = 3.13  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.799  
 Std Dev = 0.01 Rel Std Dev = 0.88  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.881  
 Std Dev = 0.00 Rel Std Dev = 0.21  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.645  
 Std Dev = 0.01 Rel Std Dev = 0.40  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 5.306  
 Std Dev = 0.00 Rel Std Dev = 0.01  
 Zero Order Coef = -150.93  
 First Order Coef = 2535.39  
 Second Order Coef = 34.55  
 Standard Deviation = 18.789291

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 6.9630 (-0.0080)  
 Sample #2 = 6.8780 (0.0600)  
 Sample #3 = 6.8410 (0.1010)  
 Sample #4 = 6.6450 (0.1130)  
 Avg % Abs = 6.8547 (0.0980)  
 Std Dev = 0.0203 (0.0167)  
 REL STD DEV = 0.296 (17.044)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.5510 (-0.0050)  
 Sample #2 = 1.4990 (0.0240)  
 Sample #3 = 1.5010 (0.0280)  
 Sample #4 = 1.4940 (0.0290)  
 Avg % Abs = 1.4980 (0.0270)  
 Std Dev = 0.0036 (0.0026)  
 REL STD DEV = 0.241 (9.799)

<<<<< CHANNEL 1 >>>>>  
 Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum To = 12869, Sum To = 12490

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.0750 (-0.0290)  
 Sample #2 = 0.0660 (-0.0040)  
 Sample #3 = 0.0620 (0.0000)  
 Sample #4 = 0.0640 (0.0020)  
 Avg % Abs = 0.0640 (-0.0007)  
 Std Dev = 0.0020 (0.0031)  
 REL STD DEV = 3.125 (458.258)

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	
0.000	0.000	-0.0004
0.040	0.039	0.0006
0.100	0.100	0.0000
0.200	0.200	-0.0004
0.300	0.300	0.0002

<<<<< CHANNEL 2 >>>>>  
 Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.097  
 Std Dev = 0.01 Rel Std Dev = 11.27  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.498  
 Std Dev = 0.00 Rel Std Dev = 0.24  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.589  
 Std Dev = 0.01 Rel Std Dev = 0.31  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.855  
 Std Dev = 0.02 Rel Std Dev = 0.30  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 9.882  
 Std Dev = 0.02 Rel Std Dev = 0.19  
 Zero Order Coef = -106.03  
 First Order Coef = 1298.85  
 Second Order Coef = 15.86  
 Standard Deviation = 20.666615

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 10.0220 (-0.0050)  
 Sample #2 = 9.9030 (0.1330)  
 Sample #3 = 9.8700 (0.1470)  
 Sample #4 = 9.8720 (0.1420)  
 Avg % Abs = 9.8817 (0.1407)  
 Std Dev = 0.0185 (0.0071)  
 REL STD DEV = 0.187 (5.044)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.6280 (-0.0080)  
 Sample #2 = 3.6020 (0.0400)  
 Sample #3 = 3.5850 (0.0490)  
 Sample #4 = 3.5810 (0.0550)  
 Avg % Abs = 3.5893 (0.0480)  
 Std Dev = 0.0112 (0.0075)  
 REL STD DEV = 0.311 (15.729)

<<<<< CHANNEL 2 >>>>>  
 Sol Value = 0.040 g/210L \*\*\*  
 Fit value = 0.1905 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum To = 12870, Sum To = 12491

<<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.8020 (-0.0080)  
 Sample #2 = 0.8060 (-0.0010)  
 Sample #3 = 0.8000 (0.0080)  
 Sample #4 = 0.7920 (0.0160)  
 Avg % Abs = 0.7993 (0.0077)  
 Std Dev = 0.0070 (0.0085)  
 REL STD DEV = 0.879 (110.933)

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3610 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
 Sample #1 = 3288.00  
 Sample #2 = 3314.00  
 Sample #3 = 3317.00  
 Sample #4 = 3329.00  
 Average Result = 3320.0000  
 STD DEV = 7.9373  
 REL STD DEV = 0.239  
 \*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
 Sample #1 = 3357.00  
 Sample #2 = 3379.00  
 Sample #3 = 3391.00  
 Sample #4 = 3350.00  
 Average Result = 3373.3333  
 STD DEV = 21.0792  
 REL STD DEV = 0.625

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

Digitally signed  
 by DERR  
 Date: 2020.08.19  
 16:34:01 -04'00'

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1009  
 3 um H2O Adjust (mg/l\*10,000) = 489  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

CAL ADJUSTMENT  
 # 80-001287 SP

DERR

BK  
 2020.08.20  
 08:15:29  
 -04'00'

Post Cal Adjust Stability Checks # 80-001287

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
08/19/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.049	11:23
Air Blank	0.000	11:24
Control Test	0.049	11:24
Air Blank	0.000	11:25
Control Test	0.049	11:25
Air Blank	0.000	11:26
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
08/19/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:31
Control Test	0.080	11:32
Air Blank	0.000	11:32
Control Test	0.081	11:33
Air Blank	0.000	11:33
Control Test	0.080	11:34
Air Blank	0.000	11:34
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SP  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
08/19/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:36
Control Test	0.202	11:37
Air Blank	0.000	11:37
Control Test	0.203	11:38
Air Blank	0.000	11:39
Control Test	0.202	11:39
Air Blank	0.000	11:40
Control Test Stats		
Average	0.2023	
Std Dev	0.0006	
Rel Std Dev(%)	0.2853	

SP  
Operator's Signature

BAKER COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001287  
08/19/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:16
Control Test	0.079	11:16
Air Blank	0.000	11:17
Control Test	0.079	11:17
Air Blank	0.000	11:17
Control Test	0.079	11:18
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP  
Operator's Signature

FLOW CAL ADJUSTMENT

BAKER COUNTY SO  
Intoxilyzer - Alconoi Analyzer  
Model 8000 SN 80-001287  
08/19/2020  
Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
1: Rate (Liters/min) = 5  
   SQRT(Diff) ) = 8.426  
2: Rate (Liters/min) = 15  
   SQRT(Diff) ) = 11.531  
3: Rate (Liters/min) = 30  
   SQRT(Diff) ) = 20.688  
Dependent Data Scale Factor = 100000 L/min  
Independent Data Scale Factor = 256  
Rounded Slope = 761  
Rounded Intercept = -972292  
Correlation = 0.98681

BK 2020.08.20  
08:16:15 -0400

SP

Digitally signed  
by DERR  
Date: 2020.08.19  
16:27:53 -0400

DERR

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO  
Time of Inspection: 12:44

Date of Inspection: 07/31/2020

Serial Number: 80-001287  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT

*SA*  
*N/A compliance not determined*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto* ISRAEL SOTO  
Signature and Printed Name

07/31/2020  
Date