




INSTRUMENT PROCESSING SHEET

Agency Marianna PDS/N 80-0001305

Florida Department of Law Enforcement

Date In 02/20/2020DI Completion Date 2/26/20 Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
Final Release Date <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FDLE</div> <div style="text-align: center; font-weight: bold; font-size: 1.2em;">FEB 26 2020</div> <div style="text-align: center; font-weight: bold;">Alcohol Testing Program</div>	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>SD1011</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 05-14-2021	0.080	DR1279	201905B 05-14-2021	0.200	SD1011	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.2</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>																									
Simulator	Serial #	Lot #/Exp																																								
0.050	SD1012	201905A 05-14-2021																																								
0.080	DR1279	201905B 05-14-2021																																								
0.200	SD1011	201904D 04-30-2021																																								
0.080 DGS	N/A	AG931603 11-12-2021																																								
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1012</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>19080</td> <td>3-4-21</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>19160</td> <td>7-9-21</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>19040</td> <td>1-29-21</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>19010</td> <td>1-3-21</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>6-5-21</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	19080	3-4-21	0.100	MP5083	19160	7-9-21	0.200	MP5084	19040	1-29-21	0.300	MP5085	19010	1-3-21	0.080 DGS	N/A	08819080A1	6-5-21	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
Simulator	Serial Number	Lot Number	Expiration																																							
0.000	MP5091	N/A	N/A																																							
0.040	MP5082	19080	3-4-21																																							
0.100	MP5083	19160	7-9-21																																							
0.200	MP5084	19040	1-29-21																																							
0.300	MP5085	19010	1-3-21																																							
0.080 DGS	N/A	08819080A1	6-5-21																																							
Simulator	Serial Number																																									
0.000	MP5086																																									
Interferent	MP5087																																									
0.050	MP5088																																									
0.080	MP5089																																									
0.200	MP5090																																									
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="text-align: center;"> <u>SPM 2/26/20</u> Tech Review / Date </div> <div style="text-align: center;"> <u>Brith Kinkland 2/26/2020</u> Admin Review / Date </div> </div>																																									



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001305, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001305</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MARIANNA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/26/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:32</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/26/2020 Date
Shayla Platt
SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 January 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

RAM BK 2/26/2020

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARIANNA PD

Time of Inspection: 11:32

Date of Inspection: 02/26/2020

Serial Number: 80-001305

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.201	0.082
0.000	0.049	0.078	0.201	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081

Standard Deviations	0.0000	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

JBM
YBK
2/26/2020

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

02/26/2020
Date

stability checks

MARIANNA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001305
 02/21/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.047	09:15
Air Blank	0.000	09:16
Control Test	0.048	09:16
Air Blank	0.000	09:17
Control Test	0.048	09:18
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	



 Operator's Signature

MARIANNA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001305
 02/21/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:19
Control Test	0.077	09:20
Air Blank	0.000	09:21
Control Test	0.078	09:21
Air Blank	0.000	09:22
Control Test	0.079	09:23
Air Blank	0.000	09:23
Control Test Stats		
Average	0.0780	
Std Dev	0.0010	
Rel Std Dev(%)	1.2821	


wet



 Operator's Signature

MARIANNA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001305
 02/21/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:24
Control Test	0.198	09:25
Air Blank	0.000	09:26
Control Test	0.200	09:26
Air Blank	0.000	09:27
Control Test	0.201	09:28
Air Blank	0.000	09:28
Control Test Stats		
Average	0.1997	
Std Dev	0.0015	
Rel Std Dev(%)	0.7650	



 Operator's Signature

MARIANNA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001305
 02/21/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.081	09:30
Air Blank	0.000	09:31
Control Test	0.081	09:31
Air Blank	0.000	09:32
Control Test	0.081	09:32
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry



 Operator's Signature

G8800
 BK
 2/26/2020

MARIANNA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 02/26/2020

Auto Calibration
 Max Power Res Value = 31
 Auto Range Res Value = 19

SN 60-001305
 07:33:03

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.089
 Std Dev = 0.01 Rel Std Dev = 11.74
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.823
 Std Dev = 0.00 Rel Std Dev = 0.51
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.900
 Std Dev = 0.02 Rel Std Dev = 1.02
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.653
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.314
 Std Dev = 0.01 Rel Std Dev = 0.14
 Zero Order Coef = -218.46
 First Order Coef = 2548.46
 Second Order Coef = 33.84
 Standard Deviation = 13.590017

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.130
 Std Dev = 0.01 Rel Std Dev = 5.13
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.555
 Std Dev = 0.01 Rel Std Dev = 0.47
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.640
 Std Dev = 0.02 Rel Std Dev = 0.56
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.020
 Std Dev = 0.00 Rel Std Dev = 0.06
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.116
 Std Dev = 0.01 Rel Std Dev = 0.06
 Zero Order Coef = -147.55
 First Order Coef = 1290.86
 Second Order Coef = 13.26
 Standard Deviation = 31.667791

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12764, Sum Io = 13610

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12768, Sum Io = 13610

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8680 (0.0030)
 Sample #2 = 1.8780 (0.0010)
 Sample #3 = 1.9140 (0.0040)
 Sample #4 = 1.9080 (0.0030)
 Avg % Abs = 1.9000 (0.0027)
 STD DEV = 0.0193 (0.0015)
 REL STD DEV = 1.015 (57.282)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12768, Sum Io = 13610

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6150 (-0.0150)
 Sample #2 = 3.6170 (-0.0170)
 Sample #3 = 3.6500 (-0.0310)
 Sample #4 = 3.6540 (-0.0320)
 Avg % Abs = 3.6403 (-0.0267)
 STD DEV = 0.0203 (0.0084)
 REL STD DEV = 0.558 (31.449)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12766, Sum Io = 13613

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8370 (0.0000)
 Sample #2 = 0.8240 (0.0060)
 Sample #3 = 0.8260 (0.0240)
 Sample #4 = 0.8180 (0.0120)
 Avg % Abs = 0.8227 (0.0140)
 STD DEV = 0.0042 (0.0092)
 REL STD DEV = 0.506 (65.465)

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8370 (0.0000)
 Sample #2 = 0.8240 (0.0060)
 Sample #3 = 0.8260 (0.0240)
 Sample #4 = 0.8180 (0.0120)
 Avg % Abs = 0.8227 (0.0140)
 STD DEV = 0.0042 (0.0092)
 REL STD DEV = 0.506 (65.465)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

<<<<< CHANNEL 1 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

<<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8370 (0.0000)
 Sample #2 = 0.8240 (0.0060)
 Sample #3 = 0.8260 (0.0240)
 Sample #4 = 0.8180 (0.0120)
 Avg % Abs = 0.8227 (0.0140)
 STD DEV = 0.0042 (0.0092)
 REL STD DEV = 0.506 (65.465)

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12771, Sum Io = 13615

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1013
 3 um H2O Adjust (mg/l*10,000) = 784
 9 um H2O Adjust (mg/l*10,000) = 478

 REL STD DEV = 0.165

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1013
 3 um H2O Adjust (mg/l*10,000) = 784
 9 um H2O Adjust (mg/l*10,000) = 478
 ***** AUTO CAL PASS

CAL ADJUSTMENT
 #80-001305

Qasm
 TBK 2/26/2020

Post Cal Adjust Stability Checks

MARIANNA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001305
02/26/2020
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:46
Control Test	0.049	09:47
Air Blank	0.000	09:47
Control Test	0.049	09:48
Air Blank	0.000	09:48
Control Test	0.049	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

MARIANNA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001305
02/26/2020
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:24
Control Test	0.079	09:25
Air Blank	0.000	09:25
Control Test	0.079	09:26
Air Blank	0.000	09:27
Control Test	0.079	09:27
Air Blank	0.000	09:28
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

MARIANNA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001305
02/26/2020
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:18
Control Test	0.201	09:19
Air Blank	0.000	09:19
Control Test	0.201	09:20
Air Blank	0.000	09:21
Control Test	0.200	09:21
Air Blank	0.000	09:22
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

SP

Operator's Signature

MARIANNA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001305
02/26/2020
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	09:50
Control Test	0.081	09:50
Air Blank	0.000	09:51
Control Test	0.080	09:51
Air Blank	0.000	09:52
Control Test	0.080	09:52
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

SP

Operator's Signature

OPDM BK 2/26/2020