



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PDS/N 80-001306Florida Department of
Law EnforcementDate In 10-26-2020 DI Completion Date 10-28-2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>234</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.20</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5089</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP5088	201905A 05-14-2021																																																												
0.080	MP5089	201905B 05-14-2021																																																												
0.200	MP5090	201904D 04-30-2021																																																												
0.080 DGS	N/A	AG931603 11-12-2021																																																												
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.10.29 13:47:37 -04'00'																																																														
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1015</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	MP5086																																																													
Interferent	MP5087																																																													
0.050	MP5088																																																													
0.080	MP5089																																																													
0.200	MP5090																																																													
Notes/Suggested Service: _____ Performed extra 0.080 wet stability check to check seal on 0.080 simulator. _____ _____ _____																																																														
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2020.10.29 Michael D. Haughney 09:10:44 -04'00' Tech Review / Date 2020.10.29 13:47:11 Admin Review / Date																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 11:13

Date of Inspection: 10/28/2020

Serial Number: 80-001306
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.048	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080

Standard Deviations	0.0003	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

10/28/2020
Date

MX
SR
2020.10.
29
13:46:56
-04'00'

stability checks

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.049	08:54
Air Blank	0.000	08:55
Control Test	0.049	08:55
Air Blank	0.000	08:56
Control Test	0.048	08:56
Air Blank	0.000	08:57
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1663	

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test	0.078	09:00
Air Blank	0.000	09:00
Control Test	0.078	09:01
Air Blank	0.000	09:02
Control Test Stats		
Average	0.0787	
Std Dev	0.0012	
Rel Std Dev(%)	1.4678	

wet 2
0.08

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.202	09:03
Air Blank	0.000	09:04
Control Test	0.201	09:04
Air Blank	0.000	09:05
Control Test	0.201	09:06
Air Blank	0.000	09:06
Control Test Stats		
Average	0.2013	
Std Dev	0.0006	
Rel Std Dev(%)	0.2868	

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.080	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:09
Control Test	0.081	09:10
Air Blank	0.000	09:10
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.079	08:19
Air Blank	0.000	08:20
Control Test	0.079	08:21
Air Blank	0.000	08:21
Control Test	0.079	08:22
Air Blank	0.000	08:22
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet 1
0.08

Operator's Signature

MX

2020.10.
29
13:46:35
-04'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001306, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001306</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GULF BREEZE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/28/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Date

10/28/2020

Israel Soto
ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.1
0.29
13:46:22
04:00'



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PDS/N 80-001306Florida Department of
Law EnforcementDate In 1/16/2020DI Completion Date 1/22/20☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DP</u> ☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____		Quality Checks Performed By <u>SP</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>231</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.167</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.484</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE JAN 23 2020 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u> <u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u> <u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>6-14-21</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>	0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>4-30-21</u>	0.080 DGS	N/A	<u>AG916501</u> <u>6-14-21</u>	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	<u>SD1012</u>	<u>201905A</u> <u>5-14-21</u>																																									
0.080	<u>DR1279</u>	<u>201905B</u> <u>5-14-21</u>																																									
0.200	<u>MP5090</u>	<u>201904D</u> <u>4-30-21</u>																																									
0.080 DGS	N/A	<u>AG916501</u> <u>6-14-21</u>																																									
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1027</u> ID # <u>26932</u>		Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2018-B</u> Acetone Stock Solution Lot # <u>2019-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>08819080A1</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	<u>MP5091</u>	N/A	N/A																																								
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																								
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																								
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																								
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																								
0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>																																								
Simulator	Serial Number																																										
0.000	<u>MP5086</u>																																										
Interferent	<u>MP5087</u>																																										
0.050	<u>MP5088</u>																																										
0.080	<u>MP5089</u>																																										
0.200	<u>MP5090</u>																																										
☒ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____																																									
Notes/Suggested Service: _____ _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>GPm 1/22/20</u> <u>Brett Kirkland 1/22/2020</u> Tech Review / Date Admin Review / Date																																									



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001306, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001306</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GULF BREEZE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/22/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>10:44</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2020

Date

Shayla Platt

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Room
13K
1/22/2020

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Software: 8100.27

^ab0W

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

QPM
BK
1/22/2020

Completed this inspection in accordance with
Shayla Platt
Signature and Title

Signature and Printed Name

FDLE/ATP Form 41 – Revised August 2005

Stability Checks # 80-001306

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/16/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	13:42
Control Test	0.048	13:42
Air Blank	0.000	13:43
Control Test	0.048	13:44
Air Blank	0.000	13:44
Control Test	0.049	13:45
Air Blank	0.000	13:45
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP

Operator's Signature

Adam

13K

1/22/2020

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/16/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	13:47
Control Test	0.077	13:47
Air Blank	0.000	13:48
Control Test	0.078	13:48
Air Blank	0.000	13:49
Control Test	0.078	13:50
Air Blank	0.000	13:50
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/16/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	14:00
Control Test	0.199	14:01
Air Blank	0.000	14:01
Control Test	0.199	14:02
Air Blank	0.000	14:03
Control Test	0.199	14:03
Air Blank	0.000	14:04
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/16/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:52
Control Test	0.080	13:52
Air Blank	0.000	13:52
Control Test	0.080	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DS

SP

Operator's Signature

Post Cal-Adjust

Checks # 80-001306
Stability

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/17/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.050	08:50
Air Blank	0.000	08:51
Control Test	0.049	08:51
Air Blank	0.000	08:52
Control Test	0.049	08:52
Air Blank	0.000	08:53
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SP

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/17/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.080	08:46
Air Blank	0.000	08:46
Control Test	0.079	08:47
Air Blank	0.000	08:47
Control Test	0.080	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/17/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.199	08:39
Air Blank	0.000	08:39
Control Test	0.199	08:40
Air Blank	0.000	08:41
Control Test	0.200	08:41
Air Blank	0.000	08:42
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

SP

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
01/17/2020
Software: 8100.27
SN 80-001306

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.080	08:57
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:58
Control Test	0.080	08:58
Air Blank	0.000	08:59
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

NGS

SP

Operator's Signature

DDMM
SK
1/22/2020

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD

Time of Inspection: 13:40

Date of Inspection: 01/16/2020

Serial Number: 80-001306

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

JDM
BK
1/22/2020

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

Signature and Printed Name

SHAYLA D PLATT

01/16/2020
Date