



INSTRUMENT PROCESSING SHEET

Agency Madison County Sheriff's Office S/N 80-001307

Florida Department of Law Enforcement

Date In 05/28/2020 DI Completion Date 06/10/2020 Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>160</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.136</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.222</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By <u>SP</u> Flow Column # <u>ATP103</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>157</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>1.148</u> (.139 - .169) 36 mm <u>1.179</u> (.156 - .190) 53 mm <u>1.250</u> (.228 - .278) 103 mm <u>1.503</u> (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.06.11 15:56:57 -04'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>SP8962</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>SP8962</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>																																									
0.080	<u>SP8962</u>	<u>201905B</u> <u>05-14-2021</u>																																									
0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>																																									
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																									
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1012/1014</u> ID # <u>30793</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>03519080A4</u></td> <td><u>4-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	<u>MP5091</u>	N/A	N/A																																								
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																								
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																								
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																								
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																								
0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>																																								
Simulator	Serial Number																																										
0.000	<u>MP5086</u>																																										
Interferent	<u>MP5087</u>																																										
0.050	<u>MP5088</u>																																										
0.080	<u>MP5089</u>																																										
0.200	<u>MP5090</u>																																										
Notes/Suggested Service: <u>Cal adjust solutions used for second cal adjust.</u> <u>0.040 Lot# 20060 2-10-22</u> <u>0.100 Lot# 20190 4-6-22</u> <u>0.200 Lot# 20160 3-18-22</u> <u>0.300 Lot# 20030 1-21-22</u>		Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <u>X2</u> <input type="checkbox"/> Other _____																																									
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<table border="0"> <tr> <td><u>Michael D. Haghighi</u> 2020.06.10 14:56:16 -04'00'</td> <td><u>Brett Kishland</u> 2020.06.11 15:52:55 -04'00'</td> </tr> <tr> <td>Tech Review / Date</td> <td>Admin Review / Date</td> </tr> </table>		<u>Michael D. Haghighi</u> 2020.06.10 14:56:16 -04'00'	<u>Brett Kishland</u> 2020.06.11 15:52:55 -04'00'	Tech Review / Date	Admin Review / Date																																				
<u>Michael D. Haghighi</u> 2020.06.10 14:56:16 -04'00'	<u>Brett Kishland</u> 2020.06.11 15:52:55 -04'00'																																										
Tech Review / Date	Admin Review / Date																																										



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001307, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001307</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/06/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:39</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/06/2020 Date
Shayla Platt
SHAYLAD PLATT,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MH
BK 2020.06.11
15:53:27
-0400

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 13:39

Date of Inspection: 06/06/2020

Serial Number: 80-001307
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079

Standard Deviations	0.0000	0.0000	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

MA

BK 2020.06.11
15:53:52
-0400'

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/06/2020
Date

Stability Checks

ADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/27/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:21
Control Test	0.047	14:21
Air Blank	0.000	14:22
Control Test	0.046	14:22
Air Blank	0.000	14:23
Control Test	0.047	14:24
Air Blank	0.000	14:24
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	



 Operator's Signature

ADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/27/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:26
Control Test	0.077	14:26
Air Blank	0.000	14:27
Control Test	0.077	14:28
Air Blank	0.000	14:28
Control Test	0.077	14:29
Air Blank	0.000	14:29
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



 Operator's Signature

ADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/27/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:31
Control Test	0.197	14:32
Air Blank	0.000	14:33
Control Test	0.198	14:33
Air Blank	0.000	14:34
Control Test	0.198	14:34
Air Blank	0.000	14:35
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	



 Operator's Signature

ADISON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001307
 05/27/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.080	14:38
Air Blank	0.000	14:38
Control Test	0.079	14:38
Air Blank	0.000	14:39
Control Test	0.079	14:39
Air Blank	0.000	14:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Dry



 Operator's Signature

MAH
 BK 2020.06.1
 11:54:17
 -04'00'

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/05/2020 17:26:40

Auto Calibration
Max Power Res Value = 40
Auto Range Res Value = 32

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12805, Sum Io = 12689

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1180 (-0.0120)
Sample #2 = 0.1090 (0.0000)
Sample #3 = 0.0850 (0.0390)
Sample #4 = 0.0860 (0.0540)
Avg % Abs = 0.0933 (0.0310)
STD DEV = 0.0136 (0.0279)
REL STD DEV = 14.547 (89.918)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0230)
Sample #2 = 0.1030 (0.0070)
Sample #3 = 0.0950 (0.0230)
Sample #4 = 0.0880 (0.0350)
Avg % Abs = 0.0953 (0.0217)
STD DEV = 0.0075 (0.0140)
REL STD DEV = 7.873 (64.835)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12793, Sum Io = 12682

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0620 (-0.0030)
Sample #2 = 0.0350 (0.0090)
Sample #3 = 0.0510 (0.0370)
Sample #4 = 0.0580 (0.0420)
Avg % Abs = 0.0480 (0.0293)
STD DEV = 0.0118 (0.0178)
REL STD DEV = 1.390 (60.633)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5430 (0.0020)
Sample #2 = 1.5080 (0.0230)
Sample #3 = 1.5260 (0.0190)
Sample #4 = 1.5360 (0.0330)
Avg % Abs = 1.5233 (0.0250)
STD DEV = 0.0142 (0.0072)
REL STD DEV = 0.931 (28.844)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12781, Sum Io = 12675

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3600 (-0.0400)
Sample #2 = 5.3490 (0.0040)
Sample #3 = 5.3510 (-0.0080)
Sample #4 = 5.3680 (0.0270)
Avg % Abs = 5.3560 (0.0077)
STD DEV = 0.0104 (0.0178)
REL STD DEV = 0.195 (231.988)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.9170 (-0.0240)
Sample #2 = 9.8980 (0.0130)
Sample #3 = 9.9080 (-0.0020)
Sample #4 = 9.8970 (0.0270)
Avg % Abs = 9.9010 (0.0127)
STD DEV = 0.0061 (0.0145)
REL STD DEV = 0.061 (114.496)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12774, Sum Io = 12669

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7100 (-0.0230)
Sample #2 = 3.6900 (-0.0110)
Sample #3 = 3.6680 (0.0190)
Sample #4 = 3.7010 (0.0170)
Avg % Abs = 3.6863 (0.0083)
STD DEV = 0.0168 (0.0168)
REL STD DEV = 0.456 (201.276)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8740 (-0.0150)
Sample #2 = 6.8610 (-0.0100)
Sample #3 = 6.8470 (0.0040)
Sample #4 = 6.8490 (0.0150)
Avg % Abs = 6.8523 (0.0030)
STD DEV = 0.0076 (0.0125)
REL STD DEV = 0.111 (417.666)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12771, Sum Io = 12671

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9160 (-0.0170)
Sample #2 = 1.9170 (0.0190)
Sample #3 = 1.9070 (0.0110)
Sample #4 = 1.9080 (0.0180)
Avg % Abs = 1.9107 (0.0160)
STD DEV = 0.0055 (0.0044)
REL STD DEV = 0.288 (27.243)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5520 (-0.0070)
Sample #2 = 3.5370 (0.0380)
Sample #3 = 3.5570 (0.0180)
Sample #4 = 3.5540 (0.0280)
Avg % Abs = 3.5493 (0.0280)
STD DEV = 0.0108 (0.0100)
REL STD DEV = 0.304 (35.714)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.093
Std Dev = 0.01 Rel Std Dev = 14.55
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.048
Std Dev = 0.01 Rel Std Dev = 1.39
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 5.356
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.686
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 1.911
Std Dev = 0.01 Rel Std Dev = 0.29
Zero Order Coef = -1745.29
First Order Coef = 9316.82
Second Order Coef = -1535.01
Standard Deviation = 2770.889160

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.01 Rel Std Dev = 7.87
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.523
Std Dev = 0.01 Rel Std Dev = 0.93
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 9.901
Std Dev = 0.01 Rel Std Dev = 0.06
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.852
Std Dev = 0.01 Rel Std Dev = 0.11
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 3.549
Std Dev = 0.01 Rel Std Dev = 0.30
Zero Order Coef = -1420.42
First Order Coef = 4940.06
Second Order Coef = -442.21
Standard Deviation = 2737.310059

MA

BK 2020.06.11
15:54:41 -04'00"

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.019	0.0187
0.040	0.106	-0.0661
0.100	0.087	0.0135
0.200	0.247	-0.0465
0.300	0.219	0.0805

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.020	0.0200
0.040	0.107	-0.0667
0.100	0.087	0.0130
0.200	0.245	-0.0450
0.300	0.221	0.0786

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 7788.00
Sample #2 = 7638.00
Sample #3 = 7756.00
Sample #4 = 7802.00
Average Result = 7732.0000
STD DEV = 84.5931
REL STD DEV = 1.094

***** CHANNEL 2
Sample #1 = 8487.00
Sample #2 = 8390.00
Sample #3 = 8430.00
Sample #4 = 8509.00
Average Result = 8443.0000
STD DEV = 60.5558
REL STD DEV = 0.717

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1012
3 um H2O Adjust (mg/l*10,000) = -3922
9 um H2O Adjust (mg/l*10,000) = -4633
**** AUTO CAL PASS

Cal Adjustment #1
Sims not prepared properly
will repeat cal adjustsp

Post Cal Adjust Stability Checks

05

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:12
Control Test	0.128	19:13
Air Blank	0.000	19:13
Control Test	0.128	19:14
Air Blank	0.000	19:14
Control Test	0.128	19:15
Air Blank	0.000	19:15
Control Test Stats		
Average	0.1280	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

08

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:24
Control Test	0.191	18:25
Air Blank	0.000	18:25
Control Test	0.190	18:26
Air Blank	0.000	18:26
Control Test	0.191	18:27
Air Blank	0.000	18:28
Control Test Stats		
Average	0.1907	
Std Dev	0.0006	
Rel Std Dev(%)	0.3028	

SP

Operator's Signature

200

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:18
Control Test	0.245	18:19
Air Blank	0.000	18:19
Control Test	0.246	18:20
Air Blank	0.000	18:21
Control Test	0.245	18:21
Air Blank	0.000	18:22
Control Test Stats		
Average	0.2453	
Std Dev	0.0006	
Rel Std Dev(%)	0.2353	

SP

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:37
Control Test	0.078	18:38
Air Blank	0.000	18:38
Control Test	0.079	18:38
Air Blank	0.000	18:39
Control Test	0.078	18:39
Air Blank	0.000	18:40
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

SP

Operator's Signature

MA

BK

2020.06.11
15:55:05
-0400'

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001307
06/06/2020 09:08:39

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 33

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12901, Sum Io = 12732

**** CHANNEL 1 ****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1160 (0.0000)
Sample #2 = 0.1250 (-0.0010)
Sample #3 = 0.0990 (0.0090)
Sample #4 = 0.1050 (0.0190)
Avg % Abs = 0.1097 (0.0090)
STD DEU = 0.0136 (0.0100)
REL STD DEU = 12.414 (111.111)

**** CHANNEL 2 ****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1100 (-0.0220)
Sample #2 = 0.0920 (-0.0120)
Sample #3 = 0.1260 (-0.0300)
Sample #4 = 0.1010 (-0.0180)
Avg % Abs = 0.1063 (-0.0200)
STD DEU = 0.0176 (0.0092)
REL STD DEU = 16.567 (45.826)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12896, Sum Io = 12732

**** CHANNEL 1 ****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8540 (-0.0020)
Sample #2 = 0.8690 (-0.0010)
Sample #3 = 0.8340 (-0.0030)
Sample #4 = 0.8360 (0.0240)
Avg % Abs = 0.8463 (0.0067)
STD DEU = 0.0197 (0.0150)
REL STD DEU = 2.322 (225.666)

**** CHANNEL 2 ****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5150 (0.0020)
Sample #2 = 1.5200 (-0.0100)
Sample #3 = 1.4750 (0.0010)
Sample #4 = 1.5190 (-0.0040)
Avg % Abs = 1.5047 (-0.0043)
STD DEU = 0.0257 (0.0055)
REL STD DEU = 1.708 (127.098)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12894, Sum Io = 12731

**** CHANNEL 1 ****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9530 (-0.0130)
Sample #2 = 1.9220 (-0.0020)
Sample #3 = 1.9060 (0.0170)
Sample #4 = 1.9280 (0.0000)
Avg % Abs = 1.9187 (0.0050)
STD DEU = 0.0114 (0.0104)
REL STD DEU = 0.593 (208.806)

**** CHANNEL 2 ****
Sample % Abs (% Abs Ref)
Sample #1 = 3.5980 (-0.0170)
Sample #2 = 3.5860 (-0.0160)
Sample #3 = 3.5530 (-0.0030)
Sample #4 = 3.5660 (-0.0060)
Avg % Abs = 3.5683 (-0.0083)
STD DEU = 0.0166 (0.0068)
REL STD DEU = 0.466 (81.682)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12893, Sum Io = 12732

**** CHANNEL 1 ****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6820 (-0.0060)
Sample #2 = 3.6570 (-0.0070)
Sample #3 = 3.6580 (0.0260)
Sample #4 = 3.6490 (0.0060)
Avg % Abs = 3.6547 (0.0083)
STD DEU = 0.0049 (0.0166)
REL STD DEU = 0.135 (199.479)

**** CHANNEL 2 ****
Sample % Abs (% Abs Ref)
Sample #1 = 6.8340 (-0.0010)
Sample #2 = 6.8030 (0.0080)
Sample #3 = 6.8080 (0.0220)
Sample #4 = 6.8170 (0.0080)
Avg % Abs = 6.8093 (0.0127)
STD DEU = 0.0071 (0.0081)
REL STD DEU = 0.104 (63.812)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12892, Sum Io = 12730

**** CHANNEL 1 ****
Sample % Abs (% Abs Ref)
Sample #1 = 5.3050 (-0.0150)
Sample #2 = 5.3280 (0.0000)
Sample #3 = 5.2870 (0.0120)
Sample #4 = 5.2900 (0.0350)
Avg % Abs = 5.3017 (0.0157)
STD DEU = 0.0229 (0.0178)
REL STD DEU = 0.431 (113.526)

**** CHANNEL 2 ****
Sample % Abs (% Abs Ref)
Sample #1 = 9.7930 (-0.0110)
Sample #2 = 9.8170 (-0.0060)
Sample #3 = 9.7930 (-0.0040)
Sample #4 = 9.7860 (0.0170)
Avg % Abs = 9.7987 (0.0023)
STD DEU = 0.0163 (0.0127)
REL STD DEU = 0.166 (546.043)

**** AUTO CAL DATA ****
**** CHANNEL 1 ****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.110
Std Dev = 0.01 Rel Std Dev = 12.41
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.846
Std Dev = 0.02 Rel Std Dev = 2.32
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.919
Std Dev = 0.01 Rel Std Dev = 0.59
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.655
Std Dev = 0.00 Rel Std Dev = 0.13
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.302
Std Dev = 0.02 Rel Std Dev = 0.43
Zero Order Coef = -277.45
First Order Coef = 2549.04
Second Order Coef = 37.18
Standard Deviation = 8.497903

**** CHANNEL 2 ****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.106
Std Dev = 0.02 Rel Std Dev = 16.57
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.505
Std Dev = 0.03 Rel Std Dev = 1.71
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.568
Std Dev = 0.02 Rel Std Dev = 0.47
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.809
Std Dev = 0.01 Rel Std Dev = 0.10
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.799
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -121.29
First Order Coef = 1306.75
Second Order Coef = 16.59
Standard Deviation = 19.262873

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.040	0.0005
0.100	0.100	0.0002
0.200	0.200	-0.0005
0.300	0.300	0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1

Sample #1 = 3043.00
Sample #2 = 3031.00
Sample #3 = 3109.00
Sample #4 = 2984.00
Average Result = 3041.3333
STD DEU = 63.1374
REL STD DEU = 2.076

**** CHANNEL 2
Sample #1 = 3357.00
Sample #2 = 3359.00
Sample #3 = 3381.00
Sample #4 = 3360.00

Average Result = 3366.6667
STD DEU = 12.4231
REL STD DEU = 0.369

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 768
9 um H2O Adjust (mg/l*10,000) = 443
**** AUTO CAL PASS

Cal Adjust #2
SP

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	-0.0000
0.100	0.100	0.0002
0.200	0.200	-0.0002
0.300	0.300	0.0001

MA

BK 2020.06.1
1 15:55:26
-0400'

Post Cal Adjust Stability Checks

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer SN 80-001307
Model 8000
06/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:17
Control Test	0.048	10:18
Air Blank	0.000	10:18
Control Test	0.048	10:19
Air Blank	0.000	10:20
Control Test	0.048	10:20
Air Blank	0.000	10:21
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer SN 80-001307
Model 8000
06/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.078	10:30
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:32
Control Test	0.078	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer SN 80-001307
Model 8000
06/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.199	10:46
Air Blank	0.000	10:47
Control Test	0.200	10:47
Air Blank	0.000	10:48
Control Test	0.198	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

SP

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer SN 80-001307
Model 8000
06/16/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:27
Control Test	0.080	10:27
Air Blank	0.000	10:27
Control Test	0.080	10:28
Air Blank	0.000	10:28
Control Test	0.079	10:29
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

MSB

MSB
BK
2020.06.11
155551-0400

Flow Cal Adjust SP

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer SN 80-001307
Model 8000
06/06/2020
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SORT(0:iff) = 3.160
2: Rate (Liters/min) = 15
SORT(0:iff) = 9.695
3: Rate (Liters/min) = 30
SORT(0:iff) = 19.973
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 580
Rounded Intercept = 42180
Correlation = 0.99992

MA

BK 2020.06.1
11:55:13
-0400