FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Orange County Sheriff's Office S/N 80-001420

Date In 09/11/2020 DI Completion Date 09-15-2020 ☑Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By RAW

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☑ Static Bag ☐ 12V DC Cable

Notes:

Final Release Date

FDLE Alcohol Testing Program

Digitally signed by FDLE Alcohol Testing Program
Date: 2020.09.17 10:37:12 -04'00'

Quality Checks Performed By IS

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 210
☑ Flow Verification (L/s)
Flow Column # ATP-105
32 mm 0.160 (.139 - .169)
36 mm 0.171 (.156 - .190)
53 mm 0.246 (.228 - .278)
103 mm 0.535 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28427
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A
		05-14-2021
0.080	MP5089	201905B
		05-14-2021
0.200	MP5090	201904D
		04-30-2021
0.080 DGS	N/A	AG931603
		11-12-2021

Flow Calibration Performed By

Flow Column #
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value
☐ Post Calibration Verification (L/s)
Flow Column #
32 mm (.139 - .169)
36 mm (.156 - .190)
53 mm (.228 - .278)
103 mm (.447 - .547)

Maintenance Performed By

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other

Temperature Checks Performed By IS

☑ Lab Temp °C 21.81
External Digital Therm. ID#: 300503
☑ 34°C +-.2 Serial #: MP5088
☑ 34°C +-.2 Serial #: MP5089
☑ 34°C +-.2 Serial #: MP5090

Calibration Adjustment Performed By

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By IS

Barometric Pressure ID# 30793
Gauge 1013 Instrument 1012
Mouth Alcohol Solution Lot # 2019-B
Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	MP5086
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other

Notes/Suggested Service:
Tech Review: Added check to Ship box AA 9-16-2020

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

Michael D Haughey 2020.09.16 12:47:01 -04'00'
Tech Review / Date

2020.09.17 10:36:40 -04'00'
Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 16:21 Date of Inspection: 09/15/2020
Serial Number: 80-001420
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.082
0.000	0.050	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081

Standard Deviations	0.0003	0.0000	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

2020.09.17 10:36:11 -04'00'

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO
Signature and Printed Name

09/15/2020
Date

FDLE/ATP Form 41 –Revised August 2005

Stability Checks

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
09/15/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:36
Control Test	0.049	14:37
Air Blank	0.000	14:37
Control Test	0.049	14:38
Air Blank	0.000	14:38
Control Test	0.048	14:39
Air Blank	0.000	14:39

Control Test Stats

Average	0.0487
Std Dev	0.0006
Rel Std Dev(%)	1.1863

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
09/15/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:41
Control Test	0.079	14:42
Air Blank	0.000	14:42
Control Test	0.079	14:43
Air Blank	0.000	14:43
Control Test	0.079	14:44
Air Blank	0.000	14:45

Control Test Stats

Average	0.0790
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
09/15/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:46
Control Test	0.200	14:46
Air Blank	0.000	14:47
Control Test	0.200	14:48
Air Blank	0.000	14:48
Control Test	0.201	14:49
Air Blank	0.000	14:49

Control Test Stats

Average	0.2003
Std Dev	0.0006
Rel Std Dev(%)	0.2882

Operator's Signature

wet

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
09/15/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:50
Control Test	0.081	14:51
Air Blank	0.000	14:51
Control Test	0.082	14:51
Air Blank	0.000	14:52
Control Test	0.082	14:52
Air Blank	0.000	14:53

Control Test Stats

Average	0.0817
Std Dev	0.0006
Rel Std Dev(%)	0.7070

Dry

Operator's Signature

2020.09.17 10:35:12 -04'00'

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001420 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001420
Owning Agency:	ORANGE COUNTY S.O.
Calibration Date:	09/15/2020
Calibration Time:	16:21

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

09/15/2020
Date

Israel Soto
ISRAEL SOTO,
Department Inspector

2020.09.17 10:35:39 -04'00'

FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Orange County Sheriff's Office S/N 80-001420

Date In 06/25/2020 DI Completion Date 6/28/20 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By RAW

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☑ Case
- ☑ Keyboard
- ☑ Feet
- ☑ Ports
- ☑ Handle
- ☑ Dry Gas Shelf
- ☑ Breath Tube
- ☑ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord
- ☑ Static Bag
- ☐ Printer Cable
- ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE Alcohol Testing Program

Digitally signed by FDLE Alcohol Testing Program
Date: 2020.07.06 08:37:26 -04'00'

Quality Checks Performed By DH

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 209
- ☒ Flow Verification (L/s)

Flow Column # ATP-103

32 mm	0.152	(.139 - .169)
36 mm	0.167	(.156 - .190)
53 mm	0.238	(.228 - .278)
103 mm	0.511	(.447 - .547)

- ☒ Barometric Pressure Check

Gauge ID # 28427

- ☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A 05-14-2021
0.080	MP5089	201905B 05-14-2021
0.200	MP5090	201904D 04-30-2021
0.080 DGS	N/A	AG931603 11-12-2021

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By SP

- ☒ Lab Temp °C 22.4

External Digital Therm. ID#: 300505

- ☒ 34°C +-.2 Serial #: MP5088
- ☒ 34°C +-.2 Serial #: MP5089
- ☒ 34°C +-.2 Serial #: MP5090

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 26932

Gauge 1014 Instrument 1013

Mouth Alcohol Solution Lot # 2019-B

Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	MP5086
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Attachments

- ☒ Form 41
- ☒ Stability Checks
- ☒ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ☐ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

Michael D Haughey 2020.07.06 08:21:47 -04'00' — Tech Review / Date

Brett Kirkland 2020.07.06 08:35:10 -04'00' — Admin Review / Date

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001420 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001420
Owning Agency:	ORANGE COUNTY S.O.
Calibration Date:	06/28/2020
Calibration Time:	08:24

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

06/28/2020
Date

Shayla Platt
SHAYLA D PLATT,
Department Inspector

BK 6 2020.07.0 08:35:36 -04'00'

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 08:24

Date of Inspection: 06/28/2020

Serial Number: 80-001420
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.080	0.201	0.082
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.082
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.201	0.081
Standard Deviations	0.0000	0.0004	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

MH

BK 2020.07.06 08:36:06 -04'00'

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT
Signature and Printed Name

06/28/2020
Date

FDLE/ATP Form 41 –Revised August 2005

Stability Checks

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
06/25/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:31
Control Test	0.048	13:31
Air Blank	0.000	13:32
Control Test	0.048	13:33
Air Blank	0.000	13:33
Control Test	0.048	13:34
Air Blank	0.000	13:34

Control Test Stats

Average	0.0480
Std Dev	0.0000
Rel Std Dev(%)	0.0000

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
06/25/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:36
Control Test	0.078	13:36
Air Blank	0.000	13:37
Control Test	0.079	13:38
Air Blank	0.000	13:38
Control Test	0.078	13:39
Air Blank	0.000	13:39

Control Test Stats

Average	0.0783
Std Dev	0.0006
Rel Std Dev(%)	0.7370

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
06/25/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:40
Control Test	0.201	13:41
Air Blank	0.000	13:42
Control Test	0.200	13:42
Air Blank	0.000	13:43
Control Test	0.200	13:44
Air Blank	0.000	13:44

Control Test Stats

Average	0.2003
Std Dev	0.0006
Rel Std Dev(%)	0.2882

Operator's Signature

wet

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
06/25/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:45
Control Test	0.079	13:46
Air Blank	0.000	13:46
Control Test	0.080	13:47
Air Blank	0.000	13:47
Control Test	0.080	13:47
Air Blank	0.000	13:48

Control Test Stats

Average	0.0797
Std Dev	0.0006
Rel Std Dev(%)	0.7247

MH
BK 2020.07.06 08:36:26 -04'00'

Dry

Operator's Signature