



INSTRUMENT PROCESSING SHEET

Agency Bicayne National ParkS/N 80-001442Florida Department of
Law EnforcementDate In 08/03/2020 DI Completion Date 08/03/2020 Ship P/U H/D CMI EE

Intake Performed By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>242</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.132</u> (.139 - .169) 36 mm <u>0.152</u> (.156 - .190) 53 mm <u>0.214</u> (.228 - .278) 103 mm <u>0.441</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 1/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 1/30/2022	Flow Calibration Performed By <u>DERR</u> Flow Column # <u>ATP101</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>243</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201905A 05/14/2021																																																											
0.080	SD3968	201905B 05/14/2021																																																											
0.200	SD3969	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG003005 1/30/2022																																																											
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.08.06 14:48:43 -04'00'	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.60C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>SD3969</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1013</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3968																																																												
0.200	SD3969																																																												
Notes/Suggested Service: <u>Flow calibration to bring values closer to nominal.</u> _____ _____ _____ _____	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<table style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; text-align: center;"> 2020.08.03 <i>Michael D. Hagley</i> 15:22:10 -04'00' </td> <td style="width:50%; text-align: center;"> 2020.08.06 14:42:38 <i>Brett Kirkland</i> -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review / Date</td> <td style="text-align: center;">Admin Review / Date</td> </tr> </table>	2020.08.03 <i>Michael D. Hagley</i> 15:22:10 -04'00'	2020.08.06 14:42:38 <i>Brett Kirkland</i> -04'00'	Tech Review / Date	Admin Review / Date																																																							
2020.08.03 <i>Michael D. Hagley</i> 15:22:10 -04'00'	2020.08.06 14:42:38 <i>Brett Kirkland</i> -04'00'																																																												
Tech Review / Date	Admin Review / Date																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK
Time of Inspection: 14:47

Date of Inspection: 08/03/2020

Serial Number: 80-001442
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077
0.000	0.050	0.080	0.201	0.077

Standard Deviations	0.0005	0.0003	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

MX
BK 2020.08.06
14:44:12
-04'00"

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera _____ DAVID E REYES-RIVERA
Signature and Printed Name

08/03/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001442	Biscayne National Park	08/03/2020	<i>PR</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>BISCAYNE NAT PARK Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001442 08/03/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:54 Control Test 0.050 10:55 Air Blank 0.000 10:55 Control Test 0.049 10:56 Air Blank 0.000 10:57 Control Test 0.050 10:57 Air Blank 0.000 10:58 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625</p> <p><i>Paul Deep Kap Lauer</i> Operator's Signature</p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>BISCAYNE NAT PARK Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001442 08/03/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:00 Control Test 0.080 11:00 Air Blank 0.000 11:01 Control Test 0.080 11:02 Air Blank 0.000 11:02 Control Test 0.080 11:03 Air Blank 0.000 11:03 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>Paul Deep Kap Lauer</i> Operator's Signature</p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>BISCAYNE NAT PARK Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001442 08/03/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:07 Control Test 0.203 11:06 Air Blank 0.000 11:08 Control Test 0.202 11:08 Air Blank 0.000 11:09 Control Test 0.202 11:10 Air Blank 0.000 11:11 Control Test Stats Average 0.2023 Std Dev 0.0006 Rel Std Dev(%) 0.2853</p> <p><i>Paul Deep Kap Lauer</i> Operator's Signature</p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>BISCAYNE NAT PARK Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001442 08/03/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:14 Control Test 0.078 11:14 Air Blank 0.000 11:15 Control Test 0.078 11:15 Air Blank 0.000 11:16 Control Test 0.078 11:16 Air Blank 0.000 11:17 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>Paul Deep Kap Lauer</i> Operator's Signature</p>

BISCAYNE NAT PARK
Intoxilyzer - Alcotest Analyzer
Model 8000 SN 80-001442
08/03/2020
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SORT(Diff) = 6.707
2: Rate (Liters/min) = 15
SORT(Diff) = 11.000
3: Rate (Liters/min) = 30
SORT(Diff) = 19.492
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 753
Rounded Intercept = -725056
Correlation = 0.99738

Flow Calibration
SN: 80-001442
Agency: Biscayne National Park
Date: 8/3/2020
By: DERR <i>[Signature]</i>

MX

BK 2020.08.0
6 14:46:09
-04'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001442, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001442</u>	UNCERTAINTY* ±
Owning Agency:	<u>BISCAYNE NAT PARK</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/03/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>14:47</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/03/2020 Date
 **DAVID E REYES-RIVERA,**
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MX
BK 2020.08.0
6 14:46:58
-04'00'