



INSTRUMENT PROCESSING SHEET

Agency Lee CSO

S/N 80-001722

Florida Department of
Law Enforcement

Date In 07/08/2020 DI Completion Date 07/14/2020

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>TDG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>TDG</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>163</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.07.17 11:42:42 -04'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>300504</u> External Digital Therm. ID#: <u>21.74</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5093</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201905A 05/14/2021																																																											
0.080	SD3968	201905B 05/14/2021																																																											
0.200	SD3969	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG003005 01/30/2022																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div> 2020.07.16 12:16:12 -04'00' </div> <div> 2020.07.17 11:40:19 -04'00' </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 5px;"> <div>Tech Review / Date</div> <div>Admin Review / Date</div> </div>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP5092	0.080	MP5093	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP5092																																																												
0.080	MP5093																																																												
0.200	MP5097																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 12:56

Date of Inspection: 07/14/2020

Serial Number: 80-001722
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.080	0.197	0.079
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.080	0.198	0.078
0.000	0.048	0.080	0.197	0.079
0.000	0.048	0.080	0.198	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0004	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:


The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

MICHAEL D HAUGHEY

07/14/2020
Date

 2020.07.1
6 12:15:21
-04'00'

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001722	Lee CSO	07/08/2020	MG

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001722 07/08/2020 Software: 8100.27	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001722 07/08/2020 Software: 8100.27	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001722 07/08/2020 Software: 8100.27	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001722 07/08/2020 Software: 8100.27																																																																																																																																																
065			065																																																																																																																																																
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr><tr><td>Control Test</td><td>0.049</td><td>12:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></table> <p><i>Jack Hunter</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:15	Control Test	0.048	12:15	Air Blank	0.000	12:16	Control Test	0.048	12:17	Air Blank	0.000	12:17	Control Test	0.049	12:18	Air Blank	0.000	12:18	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></table> <p><i>Jack Hunter</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:19	Control Test	0.079	12:20	Air Blank	0.000	12:21	Control Test	0.079	12:21	Air Blank	0.000	12:22	Control Test	0.078	12:23	Air Blank	0.000	12:23	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.200</td><td>12:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr></table> <p><i>Jack Hunter</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.200	12:26	Air Blank	0.000	12:26	Control Test	0.198	12:27	Air Blank	0.000	12:27	Control Test	0.198	12:28	Air Blank	0.000	12:29	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p><i>Jack Hunter</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:30	Control Test	0.080	12:30	Air Blank	0.000	12:31	Control Test	0.080	12:31	Air Blank	0.000	12:32	Control Test	0.080	12:32	Air Blank	0.000	12:33	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.048	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.048	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.049	12:18																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:19																																																																																																																																																	
Control Test	0.079	12:20																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.079	12:21																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.078	12:23																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.200	12:26																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.198	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.198	12:28																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5812																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test	0.080	12:30																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test	0.080	12:31																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test	0.080	12:32																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001722, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number: 80-001722
Owning Agency: LEE COUNTY SO
Calibration Date: 07/14/2020
Calibration Time: 12:56

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$). The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/14/2020

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2020.07.16
12:14:15
-04'00'

BK 2020.07.17
11:41:31
-04'00'