



INSTRUMENT PROCESSING SHEET

Agency NAS Key WestS/N 80-002170

Florida Department of Law Enforcement

Date In 06/09/2020 DI Completion Date 06/12/2020 Ship P/U H/D CMI EE

Intake Performed By <u>TDG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Needs battery change</u>	Quality Checks Performed By <u>TDG</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	201905A 05/14/2021	0.080	MP5093	201905B 05/14/2021	0.200	MP5094	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP5092	201905A 05/14/2021																																																												
0.080	MP5093	201905B 05/14/2021																																																												
0.200	MP5094	201904D 04/30/2021																																																												
0.080 DGS	N/A	AG003005 01/30/2022																																																												
Final Release Date <div style="text-align: center; padding: 10px;"> FDLE JUN 29 2020 Alcohol Testing Program </div>	Maintenance Performed By <u>MH</u> <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.38</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5097</u>																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1017</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	SD1014																																																													
Interferent	SD1015																																																													
0.050	MP4863																																																													
0.080	MP4864																																																													
0.200	MP5097																																																													
Notes/Suggested Service: <u>Emailed</u> <div style="margin-top: 10px;"> <input checked="" type="checkbox"/> APPROVED <u>6/15/2020</u> </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="margin-top: 10px;"> <u>SP 6/28/20</u> <u>Scott Kurland</u> <u>6/29/2020</u> Tech Review / Date Admin Review / Date </div>																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST
Time of Inspection: 12:45

Date of Inspection: 06/12/2020

Serial Number: 80-002170
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0003	0.0000	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY
Signature and Printed Name

06/12/2020
Date

SP
BK
6/29/2020

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-00 2170	NAS Key West	6/10/2020	MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																				
<p>0.047 to 0.053</p> <p>NAS KEY TEST Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-002170 06/10/2020 Software: 8100.27</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </table>	Air Blank	0.000	12:08	Control Test	0.050	12:09	Air Blank	0.000	12:09	Control Test	0.049	12:10	Air Blank	0.000	12:10	Control Test	0.049	12:11	Air Blank	0.000	12:11	Control Test Status			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>0.077 to 0.083</p> <p>NAS KEY TEST Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-002170 06/10/2020 Software: 8100.27</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	12:14	Control Test	0.080	12:14	Air Blank	0.000	12:15	Control Test	0.080	12:15	Air Blank	0.000	12:16	Control Test	0.080	12:17	Air Blank	0.000	12:17	Control Test Status			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>0.194 to 0.206</p> <p>NAS KEY TEST Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-002170 06/10/2020 Software: 8100.27</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2692</td><td></td></tr> </table>	Air Blank	0.000	12:20	Control Test	0.199	12:20	Air Blank	0.000	12:21	Control Test	0.200	12:22	Air Blank	0.000	12:22	Control Test	0.200	12:23	Air Blank	0.000	12:23	Control Test Status			Average	0.1997		Std Dev	0.0005		Rel Std Dev(%)	0.2692		<p>DGS</p> <p>NAS KEY TEST Intoxilizer - Alcohol Analyzer Model: 8000 SN: 80-002170 06/10/2020 Software: 8100.27</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr> <tr><td>Control Test Status</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7539</td><td></td></tr> </table>	Air Blank	0.000	12:27	Control Test	0.079	12:28	Air Blank	0.000	12:28	Control Test	0.078	12:28	Air Blank	0.000	12:29	Control Test	0.079	12:29	Air Blank	0.000	12:30	Control Test Status			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7539	
Air Blank	0.000	12:08																																																																																																																																					
Control Test	0.050	12:09																																																																																																																																					
Air Blank	0.000	12:09																																																																																																																																					
Control Test	0.049	12:10																																																																																																																																					
Air Blank	0.000	12:10																																																																																																																																					
Control Test	0.049	12:11																																																																																																																																					
Air Blank	0.000	12:11																																																																																																																																					
Control Test Status																																																																																																																																							
Average	0.0493																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1703																																																																																																																																						
Air Blank	0.000	12:14																																																																																																																																					
Control Test	0.080	12:14																																																																																																																																					
Air Blank	0.000	12:15																																																																																																																																					
Control Test	0.080	12:15																																																																																																																																					
Air Blank	0.000	12:16																																																																																																																																					
Control Test	0.080	12:17																																																																																																																																					
Air Blank	0.000	12:17																																																																																																																																					
Control Test Status																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	12:20																																																																																																																																					
Control Test	0.199	12:20																																																																																																																																					
Air Blank	0.000	12:21																																																																																																																																					
Control Test	0.200	12:22																																																																																																																																					
Air Blank	0.000	12:22																																																																																																																																					
Control Test	0.200	12:23																																																																																																																																					
Air Blank	0.000	12:23																																																																																																																																					
Control Test Status																																																																																																																																							
Average	0.1997																																																																																																																																						
Std Dev	0.0005																																																																																																																																						
Rel Std Dev(%)	0.2692																																																																																																																																						
Air Blank	0.000	12:27																																																																																																																																					
Control Test	0.079	12:28																																																																																																																																					
Air Blank	0.000	12:28																																																																																																																																					
Control Test	0.078	12:28																																																																																																																																					
Air Blank	0.000	12:29																																																																																																																																					
Control Test	0.079	12:29																																																																																																																																					
Air Blank	0.000	12:30																																																																																																																																					
Control Test Status																																																																																																																																							
Average	0.0787																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7539																																																																																																																																						
<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>																																																																																																																																				


SR
BK
6/29/2020

Post-Battery Charge

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-002170	NAS Key West	06/12/2020	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>NAS KEY WEST Intoxilyzer - Alconoc Analyzer Model 8000 SN 80-002170 06/12/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>9/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	9/210L	Time	Air Blank	0.000	09:06	Control Test	0.049	09:07	Air Blank	0.000	09:08	Control Test	0.049	09:08	Air Blank	0.000	09:09	Control Test	0.049	09:10	Air Blank	0.000	09:10	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>NAS KEY WEST Intoxilyzer - Alconoc Analyzer Model 8000 SN 80-002170 06/12/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>9/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	9/210L	Time	Air Blank	0.000	09:12	Control Test	0.080	09:12	Air Blank	0.000	09:13	Control Test	0.080	09:13	Air Blank	0.000	09:14	Control Test	0.080	09:15	Air Blank	0.000	09:15	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>NAS KEY WEST Intoxilyzer - Alconoc Analyzer Model 8000 SN 80-002170 06/12/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>9/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>09:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table>	Test	9/210L	Time	Air Blank	0.000	09:18	Control Test	0.199	09:18	Air Blank	0.000	09:19	Control Test	0.200	09:19	Air Blank	0.000	09:20	Control Test	0.199	09:21	Air Blank	0.000	09:21	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		<p>NAS KEY WEST Intoxilyzer - Alconoc Analyzer Model 8000 SN 80-002170 06/12/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>9/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table>	Test	9/210L	Time	Air Blank	0.000	09:23	Control Test	0.078	09:23	Air Blank	0.000	09:23	Control Test	0.079	09:24	Air Blank	0.000	09:24	Control Test	0.078	09:24	Air Blank	0.000	09:24	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.049	09:07																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test	0.049	09:08																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.049	09:10																																																																																																																																																	
Air Blank	0.000	09:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	09:12																																																																																																																																																	
Control Test	0.080	09:12																																																																																																																																																	
Air Blank	0.000	09:13																																																																																																																																																	
Control Test	0.080	09:13																																																																																																																																																	
Air Blank	0.000	09:14																																																																																																																																																	
Control Test	0.080	09:15																																																																																																																																																	
Air Blank	0.000	09:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	09:18																																																																																																																																																	
Control Test	0.199	09:18																																																																																																																																																	
Air Blank	0.000	09:19																																																																																																																																																	
Control Test	0.200	09:19																																																																																																																																																	
Air Blank	0.000	09:20																																																																																																																																																	
Control Test	0.199	09:21																																																																																																																																																	
Air Blank	0.000	09:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2896																																																																																																																																																		
Test	9/210L	Time																																																																																																																																																	
Air Blank	0.000	09:23																																																																																																																																																	
Control Test	0.078	09:23																																																																																																																																																	
Air Blank	0.000	09:23																																																																																																																																																	
Control Test	0.079	09:24																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test	0.078	09:24																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Operator's Signature: <i>MX</i>	Operator's Signature: <i>MX</i>	Operator's Signature: <i>MX</i>	Operator's Signature: <i>MX</i>																																																																																																																																																

SP
SK
6/29/2020



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002170, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002170</u>	UNCERTAINTY* ±
Owning Agency:	<u>NAS KEY WEST</u>	0.050 g/ 210 L
Calibration Date:	<u>06/12/2020</u>	0.080 g/ 210 L
Calibration Time:	<u>12:45</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/12/2020

Date

MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

13X
6/29/2020

Florida Department of Law Enforcement Alcohol Testing Program

BK
6/29/2021

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST

Serial Number: 80-002170

Time of Inspection: 10:56

Date of Inspection: 06/10/2020

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

AI BYPASSED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/10/2020

Date

SP

Forgot to print before DI upload MHT