



# INSTRUMENT PROCESSING SHEET

Agency Pasco County Sheriff's OfficeS/N 80-003347Florida Department of  
Law EnforcementDate In 02/14/2020DI Completion Date 2/26/20 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Left keyboard peg loose.</u>	<b>Quality Checks</b> Performed By <u>JS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>154</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.307</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1012</u></td> <td><u>201905A</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201905B</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG916501</u> <u>06-14-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1012</u>	<u>201905A</u> <u>05-17-2021</u>	0.080	<u>DR1279</u>	<u>201905B</u> <u>05-17-2021</u>	0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	<u>N/A</u>	<u>AG916501</u> <u>06-14-2021</u>	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	<u>SD1012</u>	<u>201905A</u> <u>05-17-2021</u>																																																												
0.080	<u>DR1279</u>	<u>201905B</u> <u>05-17-2021</u>																																																												
0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>																																																												
0.080 DGS	<u>N/A</u>	<u>AG916501</u> <u>06-14-2021</u>																																																												
<b>Final Release Date</b> <p style="text-align: center;"><b>FDLE</b></p> <p style="text-align: center;">FEB 27 2020</p> <p style="text-align: center;">Alcohol Testing Program</p>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																													
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1012</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>08819080A1</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	<u>N/A</u>	<u>08819080A1</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1012</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>																																																											
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																																											
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																																											
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																																											
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																																											
0.080 DGS	<u>N/A</u>	<u>08819080A1</u>	<u>6-5-21</u>																																																											
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																																											
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																																											
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																																											
0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>																																																											
Simulator	Serial Number																																																													
0.000	<u>MP5086</u>																																																													
Interferent	<u>MP5087</u>																																																													
0.050	<u>MP5088</u>																																																													
0.080	<u>MP5089</u>																																																													
0.200	<u>MP5090</u>																																																													
Notes/Suggested Service: _____ _____ _____ _____ _____		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<u>SPM 2/26/20 Brett Kirkland 2/27/2020</u> Tech Review / Date Admin Review / Date																																																												



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003347, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003347</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/26/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Shayla Platt*

02/26/2020

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*Shayla Platt*  
*15h/2/27/2020*

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 13:13

Date of Inspection: 02/26/2020

Serial Number: 80-003347  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.010 / 0.000	0.050	0.080	0.200	0.081
0.000 / 0.000	0.050	0.080	0.200	0.081
0.000 / 0.000	0.050	0.080	0.201	0.080
0.000 / 0.000	0.050	0.080	0.201	0.080
0.000 / 0.000	0.049	0.079	0.201	0.081
0.000 / 0.000	0.049	0.080	0.200	0.080
0.000 / 0.000	0.050	0.079	0.201	0.080
0.000 / 0.000	0.050	0.079	0.200	0.080
0.000 / 0.000	0.050	0.080	0.200	0.080
0.000 / 0.000	0.050	0.079	0.201	0.080

Standard Deviations	0.0004	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

**Remarks:**

A F / M A: DID NOT PROVIDE SUFFICIENT SAMPLE. 00: Control Outside Tolerance. *Too much Mouth Alcohol SP*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

02/26/2020  
Date

*SPM  
BIC  
2/27/2020*

# Stability Checks

PASCO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003347  
 02/14/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.046	13:30
Air Blank	0.000	13:30
Control Test	0.047	13:31
Air Blank	0.000	13:32
Control Test	0.047	13:32
Air Blank	0.000	13:33
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	



Operator's Signature

PASCO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003347  
 02/14/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:35
Control Test	0.077	13:35
Air Blank	0.000	13:36
Control Test	0.077	13:36
Air Blank	0.000	13:37
Control Test	0.077	13:38
Air Blank	0.000	13:38
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet



Operator's Signature

PASCO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003347  
 02/14/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.197	13:40
Air Blank	0.000	13:41
Control Test	0.197	13:41
Air Blank	0.000	13:42
Control Test	0.198	13:42
Air Blank	0.000	13:43
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	



Operator's Signature

PASCO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003347  
 02/14/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:44
Control Test	0.080	13:44
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Control Test	0.081	13:46
Air Blank	0.000	13:46
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry



Operator's Signature

BKM  
 2/27/2020

PASCOCO COUNTY SO  
 Intoxilyzer - Alcolco Analyzer  
 Model 8010  
 02/26/2020  
 SN 80-003347  
 08:33:05  
 Auto Calibration  
 Max Power Res Value = 83  
 Auto Range Res Value = 55

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.5200 (-0.0110)  
 Sample #2 = 1.5080 (-0.0190)  
 Sample #3 = 1.5000 (-0.0190)  
 Sample #4 = 1.4980 (-0.0150)  
 Avg % Abs = 1.5020 (-0.0177)  
 STD DEV = 0.0053 (0.0023)  
 REL STD DEV = 0.352 (13.072)

Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12723, Sum Io = 13162  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.7740 (-0.0140)  
 Sample #2 = 1.7370 (0.0120)  
 Sample #3 = 1.7250 (0.0230)  
 Sample #4 = 1.7470 (0.0050)  
 Avg % Abs = 1.7377 (0.0133)  
 STD DEV = 0.0090 (0.0091)  
 REL STD DEV = 0.519 (68.053)

Sol Value = 0.300 g/210L \*\*\*  
 Fit value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12715, Sum Io = 13157  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 4.9190 (-0.0040)  
 Sample #2 = 4.9700 (-0.0240)  
 Sample #3 = 4.9740 (-0.0280)  
 Sample #4 = 4.9130 (0.0150)  
 Avg % Abs = 4.9523 (-0.0123)  
 STD DEV = 0.0341 (0.0238)  
 REL STD DEV = 0.689 (192.614)

<<<<< CHANNEL 2 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.5470 (-0.0130)  
 Sample #2 = 3.5390 (0.0010)  
 Sample #3 = 3.5230 (0.0170)  
 Sample #4 = 3.5530 (-0.0080)  
 Avg % Abs = 3.5383 (0.0033)  
 STD DEV = 0.0150 (0.0127)  
 REL STD DEV = 0.424 (379.868)

Sol Value = 0.200 g/210L \*\*\*  
 Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12720, Sum Io = 13158  
 <<<<< CHANNEL 1 >>>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.4370 (-0.0130)  
 Sample #2 = 3.3880 (0.0100)  
 Sample #3 = 3.4190 (0.0020)  
 Sample #4 = 3.3910 (0.0280)  
 Avg % Abs = 3.3993 (0.0133)  
 STD DEV = 0.0171 (0.0133)  
 REL STD DEV = 0.503 (99.875)

Solution Stats Quadratic Fit Chan 1  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 -0.000 0.001  
 0.040 0.041 -0.0007  
 0.100 0.099 0.0012  
 0.200 0.201 -0.0008  
 0.300 0.300 0.0003

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0004  
 0.040 0.040 0.0000  
 0.100 0.099 0.0010  
 0.200 0.201 -0.0011  
 0.300 0.300 0.0004

Sol Value = 0.160 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3167.00  
 Sample #2 = 3265.00  
 Sample #3 = 3221.00  
 Sample #4 = 3224.00  
 Average Result = 3235.6667  
 STD DEV = 24.5832  
 REL STD DEV = 0.760

<<<<< CHANNEL 2 >>>>>  
 Sample #1 = 3454.00  
 Sample #2 = 3423.00  
 Sample #3 = 3460.00  
 Sample #4 = 3451.00  
 Average Result = 3444.6667  
 STD DEV = 19.2959  
 REL STD DEV = 0.560  
 <<<<< CHANNEL 1 >>>>>  
 Sample #1 = 3454.00  
 Sample #2 = 3423.00  
 Sample #3 = 3460.00  
 Sample #4 = 3451.00  
 Average Result = 3444.6667  
 STD DEV = 19.2959  
 REL STD DEV = 0.560

Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1013  
 3 um H2O Adjust (mg/l\*10,000) = 573  
 9 um H2O Adjust (mg/l\*10,000) = 365  
 \*\*\*\*\*  
 \*\*\*\*\* AUTO CAL PASS

CAL ADJUSTMENT  
 #80-003347 SP

Ryan BK 3/2/27/2020

# Post Cal - Adjust Stability Checks

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003347  
02/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.049	10:20
Air Blank	0.000	10:21
Control Test	0.049	10:21
Air Blank	0.000	10:22
Control Test	0.049	10:23
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003347  
02/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16
Control Test	0.080	10:17
Air Blank	0.000	10:17
Control Test	0.079	10:18
Air Blank	0.000	10:18
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003347  
02/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.199	10:10
Air Blank	0.000	10:10
Control Test	0.199	10:11
Air Blank	0.000	10:12
Control Test	0.199	10:12
Air Blank	0.000	10:13
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003347  
02/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.080	10:25
Air Blank	0.000	10:26
Control Test	0.080	10:26
Air Blank	0.000	10:27
Control Test	0.080	10:27
Air Blank	0.000	10:28
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP

Operator's Signature

QAM  
2/27/2020