



INSTRUMENT PROCESSING SHEET

Agency FHP Gainesville


S/N 80-003407

Florida Department of
Law Enforcement

Date In 12/21/2020 DI Completion Date 12/29/2020

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Keyboard peg broken, keyboard does not close, feet loose.</u>	Quality Checks Performed By <u>RAW</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>220</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.156</u> (.139 - .169) 36 mm <u>.175</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																												
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.08 09:13:38 -05'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-21</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B 10-05-22</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-22</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-21	0.080	MP5086	202010B 10-05-22	0.200	MP5090	202010D 10-06-22	0.080 DGS	N/A	AG011102 04-20-22	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>BK</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.62</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5086</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP5088	201905A 05-14-21																																																												
0.080	MP5086	202010B 10-05-22																																																												
0.200	MP5090	202010D 10-06-22																																																												
0.080 DGS	N/A	AG011102 04-20-22																																																												
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>BK</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1026</u> Instrument <u>1023</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>G11621</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5086</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </tbody> </table>		Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	G11621																																																													
Interferent	MP5087																																																													
0.050	MP5088																																																													
0.080	MP5086																																																													
0.200	MP5090																																																													
Notes/Suggested Service: <u>Replaced external screw on keyboard and tightened screw on feet. BK</u>	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
	Israel Soto <small>Digitally signed by Israel Soto Date: 2021.01.06 10:53:03 -05'00'</small> Tech Review / Date _____ Admin Review / Date _____																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003407, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003407</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/210 L 0.004
Calibration Date:	<u>12/29/2020</u>	0.080 g/210 L 0.005
Calibration Time:	<u>14:36</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

12/29/2020

Date

Brett Kirkland

BRETT H KIRKLAND,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1

25
2021.01.0
8:09:12:06
-05'00'

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:36

Date of Inspection: 12/29/2020

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.081	0.200	0.078
0.000	0.050	0.082	0.200	0.078
0.000	0.050	0.082	0.200	0.078
0.000	0.050	0.082	0.201	0.078
0.000	0.050	0.082	0.201	0.077
0.000	0.050	0.082	0.200	0.077
0.000	0.050	0.082	0.200	0.077
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.077
0.000	0.050	0.082	0.200	0.078
Standard Deviations	0.0003	0.0004	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

AS

2021.01.
08
09:11:39
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett Kirkland

BRETT H KIRKLAND

Signature and Printed Name

12/29/2020
Date

80-003407
Stabilities
12/29/20

RAW

25

2021.0
1.08
09:08:33
-05'00"

FL Highway Patrol
Intoxilyzer
Model 8000
12/29/2020
Software: 8100.27

Intoxilyzer
Model 8000
12/29/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.078	10:59
Air Blank	0.000	11:00
Control Test	0.079	11:00
Air Blank	0.000	11:01
Control Test	0.078	11:01
Air Blank	0.000	11:01
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

0.08g/210L

PGS

RAW

Operator's Signature

FL Highway Patrol
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:19
Control Test	0.201	11:19
Air Blank	0.000	11:20
Control Test	0.200	11:21
Air Blank	0.000	11:21
Control Test	0.200	11:22
Air Blank	0.000	11:22
Control Test Stats		
Average	0.2003	
Std Dev	0.0016	
Rel Std Dev(%)	0.2882	

0.20g/210L

RAW

Operator's Signature

FL Highway Patrol
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Control Test	0.081	11:13
Air Blank	0.000	11:14
Control Test	0.081	11:15
Air Blank	0.000	11:15
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08g/210L

wet

RAW

Operator's Signature

FL Highway Patrol
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:06
Control Test	0.051	11:07
Air Blank	0.000	11:07
Control Test	0.050	11:08
Air Blank	0.000	11:09
Control Test	0.051	11:09
Air Blank	0.000	11:10
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

0.05g/210L
wet

RAW

Operator's Signature