



INSTRUMENT PROCESSING SHEET

Agency Patrick AFBS/N 80-005146Florida Department of
Law EnforcementDate In 12/29/2020DI Completion Date 12/29/2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>RAW</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>RAW</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>132</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.156</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.527</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																																
Final Release Date FDLE Alcohol Testing Program	Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.08 09:25:47 -05'00'																																																																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 5/14/21</td> </tr> <tr> <td>0.080</td> <td>MP5088 MP5086</td> <td>202010B 10/5/22</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10/6/22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-22</td> </tr> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 5/14/21	0.080	MP5088 MP5086	202010B 10/5/22	0.200	MP5090	202010D 10/6/22	0.080 DGS	N/A	AG011102 04-20-22																																																																	
Simulator	Serial #	Lot #/Exp																																																																																
0.050	MP5088	201905A 5/14/21																																																																																
0.080	MP5088 MP5086	202010B 10/5/22																																																																																
0.200	MP5090	202010D 10/6/22																																																																																
0.080 DGS	N/A	AG011102 04-20-22																																																																																
Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																																																		
Temperature Checks Performed By <u>BK</u> ☒ Lab Temp °C <u>20.62</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5086</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																																																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="4">Calibration Adjustment</th> <th colspan="4">Department Inspection</th> </tr> <tr> <td colspan="4"> Performed By _____ Barometric Pressure Gauge _____ ID # _____ </td> <td colspan="4"> Performed By <u>BK</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> </td> </tr> <tr> <td colspan="4"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </table> </td> <td colspan="4"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr><td>0.000</td><td>G11621</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5086</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </table> </td> </tr> <tr> <td colspan="4"> ☐ Post Calibration Adjustment Stability Checks </td> <td colspan="4"> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____ </td> </tr> <tr> <td colspan="4" style="vertical-align: top;"> Notes/Suggested Service: <u>Tech Review corrected 0.080</u> <u>Simulator Serial Number to MP5086</u> <u>and Form 40 correction added</u> <u>"not a compliance check" RAW</u> <u>1/6/2021</u> </td> <td colspan="4" style="vertical-align: top;"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Israel Soto</u> 01-04-2021 Tech Review / Date Admin Review / Date 2021.01.08 09:24:23 -05'00' </td> </tr> </table>			Calibration Adjustment				Department Inspection				Performed By _____ Barometric Pressure Gauge _____ ID # _____				Performed By <u>BK</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u>				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr><td>0.000</td><td>G11621</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5086</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </table>				Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090	☐ Post Calibration Adjustment Stability Checks				Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____				Notes/Suggested Service: <u>Tech Review corrected 0.080</u> <u>Simulator Serial Number to MP5086</u> <u>and Form 40 correction added</u> <u>"not a compliance check" RAW</u> <u>1/6/2021</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Israel Soto</u> 01-04-2021 Tech Review / Date Admin Review / Date 2021.01.08 09:24:23 -05'00'			
Calibration Adjustment				Department Inspection																																																																														
Performed By _____ Barometric Pressure Gauge _____ ID # _____				Performed By <u>BK</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u>																																																																														
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr><td>0.000</td><td>G11621</td></tr> <tr><td>Interferent</td><td>MP5087</td></tr> <tr><td>0.050</td><td>MP5088</td></tr> <tr><td>0.080</td><td>MP5086</td></tr> <tr><td>0.200</td><td>MP5090</td></tr> </table>				Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	MP5090																																			
Simulator	Serial Number	Lot Number	Expiration																																																																															
0.000		N/A	N/A																																																																															
0.040																																																																																		
0.100																																																																																		
0.200																																																																																		
0.300																																																																																		
0.080 DGS	N/A																																																																																	
Simulator	Serial Number																																																																																	
0.000	G11621																																																																																	
Interferent	MP5087																																																																																	
0.050	MP5088																																																																																	
0.080	MP5086																																																																																	
0.200	MP5090																																																																																	
☐ Post Calibration Adjustment Stability Checks				Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																																														
Notes/Suggested Service: <u>Tech Review corrected 0.080</u> <u>Simulator Serial Number to MP5086</u> <u>and Form 40 correction added</u> <u>"not a compliance check" RAW</u> <u>1/6/2021</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Israel Soto</u> 01-04-2021 Tech Review / Date Admin Review / Date 2021.01.08 09:24:23 -05'00'																																																																														



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005146, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005146</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PATRICK AFB</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/29/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

12/29/2020

Date

Brett H Kirkland
BRETT H KIRKLAND,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1

2021.01.
08
09:22:59
-05'00'

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB

Time of Inspection: 13:57

Date of Inspection: 12/29/2020

Serial Number: 80-005146

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.080	0.200	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080

Standard Deviations	0.0005	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

AS

2021.01.
08
09:22:30
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brett H Kirkland

BRETT H KIRKLAND

Signature and Printed Name

12/29/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB

Time of Inspection: 11:47

Date of Inspection: 12/29/2020

Serial Number: 80-005146

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot #:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

*not a compliance check
RAW-116/21*

✓S

2021.01.
08
09:21:50
-05'00'

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Richard A Williams

RICHARD A WILLIAMS

Signature and Printed Name

12/29/2020

Date

Stabilized
80-005146
12/29/20

PAW

PATK
Intoxilyzer
Model 8000
12/29/2020
Software: 8100.27
SN 80-005146

Test	g/210L	Time
Air Blank	0.000	11:48
Control Test	0.049	11:49
Air Blank	0.000	11:49
Control Test	0.049	11:50
Air Blank	0.000	11:50
Control Test	0.049	11:51
Air Blank	0.000	11:52
Control Test	0.049	11:52
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05g/210L
wet

PAW

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27 -
SN 80-005146

Test	g/210L	Time
Air Blank	0.000	11:54
Control Test	0.079	11:55
Air Blank	0.000	11:56
Control Test	0.080	11:56
Air Blank	0.000	11:57
Control Test	0.080	11:57
Air Blank	0.000	11:58
Control Test	0.000	11:58
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

0.08g/210L
wet

PAW

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27
SN 80-005146

Test	g/210L	Time
Air Blank	0.000	12:00
Control Test	0.201	12:01
Air Blank	0.000	12:01
Control Test	0.202	12:02
Air Blank	0.000	12:02
Control Test	0.200	12:03
Air Blank	0.000	12:04
Control Test	0.000	12:04
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

0.20g/210L
wet

PAW

Operator's Signature

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000
12/29/2020
Software: 8100.27
SN 80-005146

Test	g/210L	Time
Air Blank	0.000	12:07
Control Test	0.080	12:08
Air Blank	0.000	12:08
Control Test	0.081	12:09
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:10
Control Test	0.000	12:10
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

0.08g/210L
wet

PAW

Operator's Signature

25
2021.01.08
09:21:26
-05'00'