



# INSTRUMENT PROCESSING SHEET

Agency Hernando CountyS/N 80-005249

Florida Department of Law Enforcement

Date In 3/2/2020DI Completion Date 4/1/2020 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>232</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
<b>Final Release Date</b> _____ _____	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>201A05A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1021</u>	<u>201A05A</u> <u>05-14-2021</u>	0.080	<u>DR1275</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD1021</u>	<u>201A05A</u> <u>05-14-2021</u>																																																											
0.080	<u>DR1275</u>	<u>201905B</u> <u>05-14-2021</u>																																																											
0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>																																																											
0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>																																																											
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1018</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>08819080A1</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	<u>N/A</u>	<u>08819080A1</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1011</u> Instrument <u>1007-SP-1011</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>																																																										
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																																										
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																																										
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																																										
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																																										
0.080 DGS	<u>N/A</u>	<u>08819080A1</u>	<u>6-5-21</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																																										
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																																										
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																																										
0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>MP5086</u>																																																												
Interferent	<u>MP5087</u>																																																												
0.050	<u>MP5088</u>																																																												
0.080	<u>MP5089</u>																																																												
0.200	<u>MP5090</u>																																																												
Notes/Suggested Service: <u>Performed cal adjustment due to barometric pressure difference greater than 1%. SP</u> _____ _____ _____	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <u>x2</u> <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____																																																												



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005249, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005249</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HERNANDO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/01/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:47</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Shayla Platt*  
 \_\_\_\_\_  
 SHAYLA D PLATT,  
 Department Inspector

04/01/2020

Date

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 13:47

Date of Inspection: 04/01/2020

Serial Number: 80-005249  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.078	0.198	0.079
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.079	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079

Standard Deviations	0.0005	0.0005	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

04/01/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 11:03

Date of Inspection: 03/29/2020

Serial Number: 80-005249  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: 5

**Remarks:**

A F / M A: Sequence AbortedSequence Aborted.

- Barometric pressure difference will require calibration adjustment. SP

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

03/29/2020  
Date

# Stability Checks

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/13/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:40
Control Test	0.047	13:40
Air Blank	0.000	13:41
Control Test	0.047	13:42
Air Blank	0.000	13:42
Control Test	0.047	13:43
Air Blank	0.000	13:43
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/13/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:46
Control Test	0.077	13:46
Air Blank	0.000	13:47
Control Test	0.077	13:48
Air Blank	0.000	13:48
Control Test	0.077	13:49
Air Blank	0.000	13:49
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/13/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.200	13:51
Air Blank	0.000	13:52
Control Test	0.200	13:53
Air Blank	0.000	13:53
Control Test	0.200	13:54
Air Blank	0.000	13:54
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



-----  
 Operator's Signature



-----  
 Operator's Signature



-----  
 Operator's Signature

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/13/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:56
Control Test	0.080	13:56
Air Blank	0.000	13:57
Control Test	0.080	13:57
Air Blank	0.000	13:58
Control Test	0.080	13:58
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry



-----  
 Operator's Signature

# Stability Checks

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/29/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:58
Control Test	0.047	06:59
Air Blank	0.000	07:00
Control Test	0.048	07:00
Air Blank	0.000	07:01
Control Test	0.048	07:01
Air Blank	0.000	07:02
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

MP5088



Operator's Signature

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/29/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:04
Control Test	0.078	07:04
Air Blank	0.000	07:05
Control Test	0.078	07:06
Air Blank	0.000	07:06
Control Test	0.078	07:07
Air Blank	0.000	07:07
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MP5089

wet



Operator's Signature

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/29/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:17
Control Test	0.080	07:17
Air Blank	0.000	07:18
Control Test	0.081	07:18
Air Blank	0.000	07:19
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Dry



Operator's Signature

HERNANDO COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-005249  
 03/29/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:10
Control Test	0.198	07:11
Air Blank	0.000	07:11
Control Test	0.198	07:12
Air Blank	0.000	07:13
Control Test	0.199	07:13
Air Blank	0.000	07:14
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

MP5090



Operator's Signature

*Repeated stabilities  
 after suspecting  
 issue w/ previous  
 simulators.*



HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
03/29/2020  
SN 80-005249  
13:01:55  
Auto Calibration  
Max Power Res Value = 59  
Auto Range Res Value = 38

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5560 (-0.0200)  
Sample #2 = 1.5240 (0.0190)  
Sample #3 = 1.5360 (0.0070)  
Sample #4 = 1.5650 (0.0070)  
Avg % Abs = 1.5417 (0.0110)  
STD DEV = 0.0211 (0.0069)  
REL STD DEV = 1.357 (62.984)

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12459, Sum Io = 13043  
<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1600 (0.0040)  
Sample #2 = 0.0340 (0.0220)  
Sample #3 = 0.0700 (0.0230)  
Sample #4 = 0.0480 (0.0560)  
Avg % Abs = 0.0507 (0.0337)  
STD DEV = 0.0181 (0.0193)  
REL STD DEV = 35.818 (57.468)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<< CHANNEL 1 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.051  
Std Dev = 0.02 Rel Std Dev = 35.82  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.771  
Std Dev = 0.03 Rel Std Dev = 3.85  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.834  
Std Dev = 0.02 Rel Std Dev = 0.98  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.593  
Std Dev = 0.01 Rel Std Dev = 0.15  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.202  
Std Dev = 0.03 Rel Std Dev = 0.53  
Zero Order Coef = -101.94  
First Order Coef = 2550.78  
Second Order Coef = 40.44  
Standard Deviation = 44.649723

<<<< CHANNEL 2 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.152  
Std Dev = 0.01 Rel Std Dev = 8.24  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.542  
Std Dev = 0.02 Rel Std Dev = 1.37  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.563  
Std Dev = 0.01 Rel Std Dev = 0.21  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.845  
Std Dev = 0.02 Rel Std Dev = 0.27  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.831  
Std Dev = 0.02 Rel Std Dev = 0.25  
Zero Order Coef = -180.63  
First Order Coef = 1322.40  
Second Order Coef = 14.98  
Standard Deviation = 35.070839

Solution Stats Quadratic Fit Chan 1  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.001 -0.0006  
0.040 0.040 0.0003  
0.100 0.099 0.0011  
0.200 0.201 -0.0013  
0.300 0.300 0.0005

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.000 -0.0004  
0.040 0.040 0.0002  
0.100 0.099 0.0008  
0.200 0.201 -0.0010  
0.300 0.300 0.0004

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5410 (0.0040)  
Sample #2 = 3.5600 (-0.0180)  
Sample #3 = 3.5580 (0.0040)  
Sample #4 = 3.5720 (-0.0040)  
Avg % Abs = 3.5633 (-0.0060)  
STD DEV = 0.0076 (0.0111)  
REL STD DEV = 0.212 (185.592)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12448, Sum Io = 13036  
<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8070 (-0.0180)  
Sample #2 = 0.7500 (0.0290)  
Sample #3 = 0.7580 (0.0290)  
Sample #4 = 0.8050 (0.0190)  
Avg % Abs = 0.7710 (0.0257)  
STD DEV = 0.0297 (0.0058)  
REL STD DEV = 3.854 (22.494)

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 9.8400 (-0.0090)  
Sample #2 = 9.8000 (0.0200)  
Sample #3 = 9.8400 (0.0060)  
Sample #4 = 9.8490 (0.0120)  
Avg % Abs = 9.8307 (0.0127)  
STD DEV = 0.0244 (0.0070)  
REL STD DEV = 0.248 (55.451)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12431, Sum Io = 13025  
<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5590 (-0.0120)  
Sample #2 = 3.5870 (0.0100)  
Sample #3 = 3.5960 (0.0170)  
Sample #4 = 3.5970 (0.0310)  
Avg % Abs = 3.5933 (0.0193)  
STD DEV = 0.0055 (0.0107)  
REL STD DEV = 0.153 (55.307)

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3385.00  
Sample #2 = 3404.00  
Sample #3 = 3447.00  
Sample #4 = 3366.00  
Average Result = 3405.6667  
STD DEV = 40.5257  
REL STD DEV = 1.190  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3325.00  
Sample #2 = 3340.00  
Sample #3 = 3334.00  
Sample #4 = 3336.00  
Average Result = 3336.6667  
STD DEV = 3.0551  
REL STD DEV = 0.092  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1017  
3 um H2O Adjust (mg/l\*10,000) = 404  
9 um H2O Adjust (mg/l\*10,000) = 473  
\*\*\*\* AUTO CAL PASS

CAL ADJUSTMENT SR  
#80-005249

# Post Cal Adjust Stability Checks # 80-005249

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
03/29/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:33
Control Test	0.048	14:34
Air Blank	0.000	14:34
Control Test	0.048	14:35
Air Blank	0.000	14:35
Control Test	0.048	14:36
Air Blank	0.000	14:37
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
03/29/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:39
Control Test	0.078	14:39
Air Blank	0.000	14:40
Control Test	0.078	14:41
Air Blank	0.000	14:41
Control Test	0.079	14:42
Air Blank	0.000	14:42
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

SP

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
03/29/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:44
Control Test	0.197	14:44
Air Blank	0.000	14:45
Control Test	0.198	14:46
Air Blank	0.000	14:46
Control Test	0.198	14:47
Air Blank	0.000	14:47
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

SP

Operator's Signature

HERNANDO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005249  
03/29/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:30
Control Test	0.080	14:30
Air Blank	0.000	14:31
Control Test	0.080	14:31
Air Blank	0.000	14:31
Control Test	0.080	14:32
Air Blank	0.000	14:32
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGB

SP

Operator's Signature



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 10:16

Date of Inspection: 04/01/2020

Serial Number: 80-005249  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT

COMPLIANCE NOT DETERMINED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/01/2020  
Date