




INSTRUMENT PROCESSING SHEET

Agency Fernadina Beach Police Department S/N 80-005538

Florida Department of Law Enforcement

Date In 06/05/2020 DI Completion Date 6/10/2020 Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>177</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.06.12 15:09:12 -04'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>07-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>07-30-2021</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.3</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>																																																											
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>																																																											
0.200	<u>MP5090</u>	<u>201904D</u> <u>07-30-2021</u>																																																											
0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		<u>N/A</u>	<u>N/A</u>	0.040				0.100				0.200				0.300				0.080 DGS	<u>N/A</u>			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	<u>N/A</u>			Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		<u>N/A</u>	<u>N/A</u>																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number																																																												
0.000	<u>MP5086</u>																																																												
Interferent	<u>MP5087</u>																																																												
0.050	<u>MP5088</u>																																																												
0.080	<u>MP5089</u>																																																												
0.200	<u>MP5090</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"> <tr> <td style="width:50%;"><u>Michael D. Hargis</u> 2020.06.11 14:46:52 -04'00'</td> <td style="width:50%;"><u>Brett Kirkland</u> 2020.06.12 15:04:16 -04'00'</td> </tr> </table>		<u>Michael D. Hargis</u> 2020.06.11 14:46:52 -04'00'	<u>Brett Kirkland</u> 2020.06.12 15:04:16 -04'00'																																																									
<u>Michael D. Hargis</u> 2020.06.11 14:46:52 -04'00'	<u>Brett Kirkland</u> 2020.06.12 15:04:16 -04'00'																																																												
	Tech Review / Date	Admin Review / Date																																																											



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005538, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005538</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FERNANDINA BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/10/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:47</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3). The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

MF

Shayla Platt

06/10/2020

Date

SHAYLA D PLATT,

Department Inspector

BK 2020.06.1
2 15:04:44
-0400

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FERNANDINA BEACH PD
Time of Inspection: 11:47

Date of Inspection: 06/10/2020

Serial Number: 80-005538
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.052	0.081	0.202	0.079
0.000	0.053	0.082	0.202	0.078
0.000	0.054	0.082	0.203	0.078
0.000	0.052	0.083	0.203	0.079
0.000	0.052	0.082	0.201	0.078
0.000	0.052	0.084	0.202	0.078
0.000	0.052	0.084	0.202	0.077
0.000	0.052	0.084	0.201	0.077
0.000	0.053	0.084	0.201	0.077
0.000	0.051	0.083	0.202	0.077

Standard Deviations	0.0008	0.0011	0.0007	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

MH

BK 2020.06.12
15:00:07
-0107

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

06/10/2020
Date

Stability Checks

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 06/08/2020
 Software: 8100.27

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 06/08/2020
 Software: 8100.27

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.051	09:51
Air Blank	0.000	09:52
Control Test	0.051	09:53
Air Blank	0.000	09:53
Control Test	0.051	09:54
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.082	10:00
Air Blank	0.000	10:00
Control Test	0.081	10:01
Air Blank	0.000	10:01
Control Test	0.082	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.204	10:05
Air Blank	0.000	10:06
Control Test	0.203	10:06
Air Blank	0.000	10:07
Control Test	0.202	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.2030	
Std Dev	0.0010	
Rel Std Dev(%)	0.4926	


Wet



Operator's Signature



Operator's Signature



Operator's Signature

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 06/08/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.081	10:11
Air Blank	0.000	10:11
Control Test	0.081	10:12
Air Blank	0.000	10:12
Control Test	0.081	10:12
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MY

BK 2020.06.12 15:05:29 -0400

Dry



Operator's Signature

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: James Smith on 2/25/2020

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005538

Bill To Address:
Fernandino Beach PD

Ship to Address:
FDLE-ATP Tallahassee

Reason for Return:
Instrument will not pump air to analyze simulator.

Please choose one of the following options:

1. I _____, authorize all repairs.

2. I _____, authorize repairs up to \$ _____.

3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: James Smith
Phone #: (904) 310-3211 Email: jsmith@fbfl.org
ATP Contact Name: Shayla Platt ATP Email: shaylplatt@fdle.state.fl.us

MH
BK 2020.06.12
15:05:52
-04'00



INSTRUMENT PROCESSING SHEET

Agency Fernadina Beach Police Department S/N 80-005538

Florida Department of Law Enforcement Date In 02/12/2020 DI Completion Date _____ Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ Final Release Date _____ _____	Quality Checks Performed By <u>DL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>174</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>06-17-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG916501</u> <u>06-17-2021</u>	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By _____ <input type="checkbox"/> Lab Temp °C _____ External Digital Therm. ID#: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____ <input type="checkbox"/> 34°C +/- .2 Serial #: _____																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>																																																											
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>																																																											
0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>																																																											
0.080 DGS	N/A	<u>AG916501</u> <u>06-17-2021</u>																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000																																																													
Interferent																																																													
0.050																																																													
0.080																																																													
0.200																																																													

MK
BK 2020.06.1
2 15:06:15
-04'00'

Stability Checks

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 02/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:08
Control Test	0.000	15:09
Air Blank	0.000	15:09
Control Test	0.000	15:10
Air Blank	0.000	15:11
Control Test	0.000	15:11
Air Blank	0.000	15:12
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05

SS

Operator's Signature

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 02/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:14
Control Test	0.000	15:15
Air Blank	0.000	15:15
Control Test	0.000	15:16
Air Blank	0.000	15:16
Control Test	0.000	15:17
Air Blank	0.000	15:17
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

0.08

SS

Operator's Signature

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 02/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:18
Control Test	0.000	15:19
Air Blank	0.000	15:19
Control Test	0.000	15:20
Air Blank	0.000	15:20
Control Test	0.000	15:21
Air Blank	0.000	15:21
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.200

SS

Operator's Signature

Air is not
 pumping through
 simulators

FERNANDINA BEACH PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-005538
 02/12/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:22
Control Test	0.080	15:22
Air Blank	0.000	15:23
Control Test	0.080	15:23
Air Blank	0.000	15:24
Control Test	0.080	15:24
Air Blank	0.000	15:25
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DRY

SS

Operator's Signature

MS

BK 2020.06.12
 15:06:40
 -0400