



INSTRUMENT PROCESSING SHEET

Agency Miami Dade PD

S/N 80-005550

Florida Department of Law Enforcement

Date In 04/15/2020 DI Completion Date 04/21/2020

Ship P/U H/D CMI EE

Intake, Quality Checks, Flow Calibration, Maintenance sections with various checkboxes and data entry fields.

Final Release Date section: FDLE APR 30 2020 Alcohol Testing Program

Table with 3 columns: Simulator, Serial #, Lot #/Exp. Rows include 0.050, 0.080, 0.200, and 0.080 DGS.

Temperature Checks section: Lab Temp 22.17, External Digital Therm. ID# 300504, 34°C +/-2 Serial #s.

Calibration Adjustment section: Barometric Pressure Gauge table, Post Calibration Adjustment Stability Checks table.

Department Inspection section: Barometric Pressure ID# 28199, Gauge 1014, Instrument 1013, Mouth Alcohol Solution Lot # 2019-B, Acetone Stock Solution Lot # 2019-A, Simulator table.

Attachments section: Checkboxes for Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, Other.

Notes/Suggested Service: E-mailed, APPROVED 04/21/2020

Instrument Compliance section: Instrument Complies with Chapter 11D-8, FAC, Return to/Place into Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use. Tech Review / Date and Admin Review / Date fields.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 11:15

Date of Inspection: 04/21/2020

Serial Number: 80-005550
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021 | 0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021 | 0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.203 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.205 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.051 | 0.083 | 0.205 | 0.081 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0004 | 0.0005 | 0.0000 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael D Haughey MICHAEL D HAUGHEY
Signature and Printed Name

04/21/2020
Date

*SP
BK
4/27/2020*

| | | | | |
|---------------------|----------------------|---------------|-------------|---------------------|
| TYPE OF TEST | SERIAL NUMBER | AGENCY | DATE | PERFORMED BY |
| Stabilities | 80-00 5550 | Miami Dade PD | 04/21/2020 | MA |

| | | | | | | |
|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|
| 0.05g/210L | 0.08g/210L | 0.20g/210L | 0.077 to 0.083 | 0.077 to 0.083 | 0.08g/210L | 0.077 to 0.083 |
| <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> | <input checked="" type="checkbox"/> |

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005550
04/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:01 |
| Control Test | 0.052 | 09:02 |
| Air Blank | 0.000 | 09:03 |
| Control Test | 0.051 | 09:03 |
| Air Blank | 0.000 | 09:04 |
| Control Test | 0.050 | 09:05 |
| Air Blank | 0.000 | 09:05 |
| Control Test Stats | | |
| Average | 0.0510 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.9608 | |

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005550
04/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:06 |
| Control Test | 0.081 | 09:07 |
| Air Blank | 0.000 | 09:08 |
| Control Test | 0.081 | 09:08 |
| Air Blank | 0.000 | 09:09 |
| Control Test | 0.082 | 09:10 |
| Air Blank | 0.000 | 09:10 |
| Control Test Stats | | |
| Average | 0.0813 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7099 | |

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005550
04/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:11 |
| Control Test | 0.203 | 09:12 |
| Air Blank | 0.000 | 09:12 |
| Control Test | 0.203 | 09:13 |
| Air Blank | 0.000 | 09:14 |
| Control Test | 0.202 | 09:14 |
| Air Blank | 0.000 | 09:15 |
| Control Test Stats | | |
| Average | 0.2027 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2849 | |

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005550
04/21/2020
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 09:16 |
| Control Test | 0.082 | 09:17 |
| Air Blank | 0.000 | 09:17 |
| Control Test | 0.082 | 09:18 |
| Air Blank | 0.000 | 09:18 |
| Control Test | 0.082 | 09:18 |
| Air Blank | 0.000 | 09:19 |
| Control Test Stats | | |
| Average | 0.0820 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

MIAMI DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005550
04/21/2020
Software: 8100.27

DGS

Operator's Signature

Operator's Signature

Operator's Signature


Operator's Signature

Operator's Signature

SP

BK

4/28/2020



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005550, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

| | | | |
|-------------------|----------------------|--------------------------------|-------|
| Serial Number: | <u>80-005550</u> | UNCERTAINTY* ± | |
| Owning Agency: | <u>MIAMI DADE PD</u> | 0.050 g/ 210 L | 0.004 |
| Calibration Date: | <u>04/21/2020</u> | 0.080 g/ 210 L | 0.005 |
| Calibration Time: | <u>11:15</u> | 0.200 g/ 210 L | 0.007 |
| | | 0.080 g/ 210 L Dry Gas Control | 0.005 |

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.
Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


MICHAEL D HAUGHEY,
Department Inspector

04/21/2020 Date

FDLE/ATP Form 69 January 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP
BRK
4/28/2020

Return Material Authorization

Ship to: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Reginald Myrtel on 4/14/2019

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005550

Bill To Address:
Miami Dade Police
1567 NW 79 AVE
Miami FL 33126

Ship to Address:
FDLE FAROC
ATP
4700 Terminal Dr Site 1
Fort Myers FL 33907

Reason for Return:
EEPROM issue. Ser # reported by instrument is
80-000000. Also passwords all reset to default.

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$ _____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sergeant Reginald Myrtel
Phone #: (305) 785-3706 Email: rjmyrtel@~~mpd~~mdpd.com

ATP Contact Name: Michael Haughey ATP Email: michael.haughey@fdle.state.fl.us

*SP ASK
4/27/2019*