

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD
Time of Inspection: 13:31

Date of Inspection: 06/09/2020

Serial Number: 80-005655
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.079	0.197	0.078

Standard Deviations	0.0006	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
BK
6/29/2020

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

06/09/2020
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005655	Fort Myers PD	06/09/2020	MG

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>
--	--	--	--

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2020 Software: 8100.27			FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2020 Software: 8100.27			FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2020 Software: 8100.27			FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 06/09/2020 Software: 8100.27		
Air Blank	0.000	10:01	Air Blank	0.000	10:08	Air Blank	0.000	10:13	Air Blank	0.000	10:18
Control Test	0.048	10:02	Control Test	0.078	10:09	Control Test	0.196	10:14	Control Test	0.079	10:18
Air Blank	0.000	10:02	Air Blank	0.000	10:09	Air Blank	0.000	10:14	Air Blank	0.000	10:19
Control Test	0.048	10:03	Control Test	0.078	10:10	Control Test	0.196	10:15	Control Test	0.079	10:19
Air Blank	0.000	10:03	Air Blank	0.000	10:11	Air Blank	0.000	10:16	Air Blank	0.000	10:20
Control Test	0.048	10:04	Control Test	0.078	10:11	Control Test	0.196	10:16	Control Test	0.078	10:20
Air Blank	0.000	10:05	Air Blank	0.000	10:12	Air Blank	0.000	10:17	Air Blank	0.000	10:21
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0480		Average	0.0780		Average	0.1960		Average	0.0787	
Std Dev	0.0000		Std Dev	0.0000		Std Dev	0.0000		Std Dev	0.0006	
Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.7339	



Jack Hunter
Operator's Signature

Jack Hunter
Operator's Signature

Jack Hunter
Operator's Signature

Jack Hunter
Operator's Signature

SR BK 6/29/2020



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005655, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005655</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FORT MYERS PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/09/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:31</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
 This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

SP 13K 6/29/2020

06/09/2020 Date

Michael D. Haughey
MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 April 2020
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality