



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CO SO

9/2/2020 SP

S/N 80-006385

Florida Department of Law Enforcement

Date In 08-14-2020

DI Completion Date 8/18/2020

Ship

P/U

H/D

CMI

EE

Intake Performed By <u>KAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>1.95</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.09.11 12:07:52 -04'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	MP5088	201905A 05-14-2021																																									
0.080	MP5089	201905B 05-14-2021																																									
0.200	MP5090	201904D 04-30-2021																																									
0.080 DGS	N/A	AG931603 11-12-2021																																									
		Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.45/22.9</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5090</u>																																									
Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1015</u> ID # <u>26932</u>		Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421/26932/28421/8P</u> Gauge <u>1012/1015</u> Instrument <u>1010/1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>2-10-22</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>4-6-22</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>3-18-22</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>1-21-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03519080A4</td> <td>4-5-21</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	2-10-22	0.100	MP5083	20190	4-6-22	0.200	MP5084	20160	3-18-22	0.300	MP5085	20030	1-21-22	0.080 DGS	N/A	03519080A4	4-5-21	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	MP5091	N/A	N/A																																								
0.040	MP5082	20060	2-10-22																																								
0.100	MP5083	20190	4-6-22																																								
0.200	MP5084	20160	3-18-22																																								
0.300	MP5085	20030	1-21-22																																								
0.080 DGS	N/A	03519080A4	4-5-21																																								
Simulator	Serial Number																																										
0.000	MP5086																																										
Interferent	MP5087																																										
0.050	MP5088																																										
0.080	MP5089																																										
0.200	MP5090																																										
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A</td> <td>5-14-21</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B</td> <td>5-14-21</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D</td> <td>4-30-20</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102</td> <td>4-20-22</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	5-14-21	0.080	MP5089	201905B	5-14-21	0.200	MP5090	201904D	4-30-20	0.080 DGS	N/A	AG011102	4-20-22	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	MP5088	201905A	5-14-21																																								
0.080	MP5089	201905B	5-14-21																																								
0.200	MP5090	201904D	4-30-20																																								
0.080 DGS	N/A	AG011102	4-20-22																																								
Notes/Suggested Service: <u>Tech Review: Instrument Requires cal adjustment due to .08 ARS & DGS difference greater than .003.</u> <u>SP 9/2/2020</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
2020.09.10 <u>Michael D. Hargray</u> 08:53:53 -04'00'		2020.09.11 <u>Brett Kirkland</u> 12:04:37 -04'00'																																									
Tech Review / Date		Admin Review / Date																																									



Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006385, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006385</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/02/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:38</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

09/02/2020

Date

SHAYLAD PLATT,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MX

BK

2020.09.11
12:05:05
-04'00

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 11:38

Date of Inspection: 09/02/2020

Serial Number: 80-006385
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.080	0.201	0.079
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.079
Standard Deviations	0.0005	0.0003	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

2020.09.
11
12:05:31
-04'00'

MX

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

09/02/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 15:16

Date of Inspection: 08/18/2020

Serial Number: 80-006385
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.047	0.077	0.196	0.080
0.000	0.047	0.076	0.196	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.047	0.076	0.196	0.080
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081
0.000	0.048	0.077	0.196	0.081
0.000	0.047	0.077	0.196	0.081

Standard Deviations	0.0003	0.0004	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Instrument requires cal adjustment. Will re-inspect sp

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

08/18/2020
Date

MX

BK

2020.09.1
11:20:58
-04'00'

Stability Checks

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006385
 08/17/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.047	11:29
Air Blank	0.000	11:29
Control Test	0.047	11:30
Air Blank	0.000	11:30
Control Test	0.048	11:31
Air Blank	0.000	11:32
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006385
 08/17/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:33
Control Test	0.077	11:34
Air Blank	0.000	11:34
Control Test	0.077	11:35
Air Blank	0.000	11:35
Control Test	0.077	11:36
Air Blank	0.000	11:37
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006385
 08/17/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:43
Control Test	0.196	11:44
Air Blank	0.000	11:44
Control Test	0.196	11:45
Air Blank	0.000	11:45
Control Test	0.196	11:46
Air Blank	0.000	11:47
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature



Operator's Signature



Operator's Signature

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006385
 08/17/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:49
Control Test	0.081	11:50
Air Blank	0.000	11:50
Control Test	0.080	11:50
Air Blank	0.000	11:51
Control Test	0.081	11:51
Air Blank	0.000	11:52
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Dry



Operator's Signature

MX

BK 2020.09.11 12:06:18 -04'00'

SR

BK

MX

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 0.000
 0.040 0.040 0.005
 0.100 0.101 0.0007
 0.200 0.200 0.0004
 0.300 0.300 0.0001

Sol Val = 0.000 mg/l or 0.000 g/210L
 % Abs = 0.165
 Std Dev = 0.02 Rel Std Dev = 30.65
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.762
 Std Dev = 0.03 Rel Std Dev = 4.20
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.820
 Std Dev = 0.02 Rel Std Dev = 1.01
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.474
 Std Dev = 0.03 Rel Std Dev = 0.76
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.091
 Std Dev = 0.02 Rel Std Dev = 0.39
 Zero Order Coef = -167.25
 First Order Coef = 2663.03
 Second Order Coef = 34.81
 Standard Deviation = 22.704950

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.9590 (-0.01303)
 Sample #2 = 6.9140 (0.0420)
 Sample #3 = 6.9600 (0.0180)
 Sample #4 = 6.9390 (0.0320)
 Avg % Abs = 6.9377 (0.0307)
 STD DEV = 0.0230 (0.0121)
 REL STD DEV = 0.332 (39.311)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5930 (-0.0050)
 Sample #2 = 1.5810 (0.0140)
 Sample #3 = 1.5600 (0.0070)
 Sample #4 = 1.5760 (-0.0020)
 Avg % Abs = 1.5723 (0.0063)
 STD DEV = 0.0110 (0.0080)
 REL STD DEV = 0.698 (126.644)

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 SN 80-006385
 Model 8000
 09/02/2020
 Auto Calibration
 Max Power Res Value = 93
 Auto Range Res Value = 71

***** CHANNEL 1 *****
 Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12707, 9um Io = 13976
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.7950 (-0.0080)
 Sample #2 = 1.8150 (0.0030)
 Sample #3 = 1.8400 (0.0020)
 Sample #4 = 1.8040 (0.0110)
 Avg % Abs = 1.8197 (0.0053)
 STD DEV = 0.0184 (0.0049)
 REL STD DEV = 1.014 (92.492)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.300 g/210L ***
 Fit Value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12694, 9um Io = 13969
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.1000 (-0.0040)
 Sample #2 = 5.1030 (0.0010)
 Sample #3 = 5.1010 (0.0030)
 Sample #4 = 5.0680 (0.0360)
 Avg % Abs = 5.0907 (0.0133)
 STD DEV = 0.0197 (0.0197)
 REL STD DEV = 0.386 (147.415)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.9930 (-0.0070)
 Sample #2 = 9.9890 (0.0110)
 Sample #3 = 10.0140 (0.0060)
 Sample #4 = 9.9690 (0.0280)
 Avg % Abs = 9.9907 (0.0157)
 STD DEV = 0.0225 (0.0108)
 REL STD DEV = 0.226 (68.845)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6620 (0.0030)
 Sample #2 = 3.6540 (0.0180)
 Sample #3 = 3.6600 (0.0220)
 Sample #4 = 3.6470 (0.0230)
 Avg % Abs = 3.6537 (0.0210)
 STD DEV = 0.0065 (0.0026)
 REL STD DEV = 0.178 (12.599)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1770 (-0.0130)
 Sample #2 = 0.1540 (-0.0070)
 Sample #3 = 0.1260 (0.0090)
 Sample #4 = 0.1890 (-0.0210)
 Avg % Abs = 0.1563 (-0.0063)
 STD DEV = 0.0316 (0.0150)
 REL STD DEV = 20.191 (237.017)

***** CHANNEL 1 *****
 Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3254.00
 Sample #2 = 3341.00
 Sample #3 = 3323.00
 Sample #4 = 3308.00
 Average Result = 3324.0000
 STD DEV = 16.5227
 REL STD DEV = 0.497

 ***** CHANNEL 2 *****
 Sample #1 = 3273.00
 Sample #2 = 3286.00
 Sample #3 = 3300.00
 Sample #4 = 3309.00
 Average Result = 3298.3333
 STD DEV = 11.5902
 REL STD DEV = 0.351

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.156
 Std Dev = 0.03 Rel Std Dev = 20.19
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.572
 Std Dev = 0.01 Rel Std Dev = 0.70
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.654
 Std Dev = 0.01 Rel Std Dev = 0.18
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.938
 Std Dev = 0.02 Rel Std Dev = 0.33
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.991
 Std Dev = 0.02 Rel Std Dev = 0.23
 Zero Order Coef = -190.41
 First Order Coef = 1298.85
 Second Order Coef = 14.95
 Standard Deviation = 14.133142

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5060 (-0.0020)
 Sample #2 = 3.4550 (0.0260)
 Sample #3 = 3.5040 (0.0100)
 Sample #4 = 3.4620 (0.0270)
 Avg % Abs = 3.4737 (0.0210)
 STD DEV = 0.0265 (0.0095)
 REL STD DEV = 0.763 (45.426)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12697, 9um Io = 13972
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5060 (-0.0020)
 Sample #2 = 3.4550 (0.0260)
 Sample #3 = 3.5040 (0.0100)
 Sample #4 = 3.4620 (0.0270)
 Avg % Abs = 3.4737 (0.0210)
 STD DEV = 0.0265 (0.0095)
 REL STD DEV = 0.763 (45.426)

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12704, 9um Io = 13975
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7600 (0.0000)
 Sample #2 = 0.7900 (-0.0140)
 Sample #3 = 0.7270 (0.0200)
 Sample #4 = 0.7660 (0.0000)
 Avg % Abs = 0.7617 (0.0020)
 STD DEV = 0.0320 (0.0171)
 REL STD DEV = 4.198 (854.401)

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 485
 9 um H2O Adjust (mg/l*10,000) = 511

 **** AUTO CAL PASS

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 485
 9 um H2O Adjust (mg/l*10,000) = 511

 **** AUTO CAL PASS

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 485
 9 um H2O Adjust (mg/l*10,000) = 511

 **** AUTO CAL PASS

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 485
 9 um H2O Adjust (mg/l*10,000) = 511

 **** AUTO CAL PASS

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 485
 9 um H2O Adjust (mg/l*10,000) = 511

 **** AUTO CAL PASS

CAL ADJUSTMENT
 #80-006385 SR

Post Cal Adjust Stability Checks #80-006385

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006385
 09/02/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.048	10:00
Air Blank	0.000	10:01
Control Test	0.048	10:01
Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:03
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

 Operator's Signature

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006385
 09/02/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.080	09:49
Air Blank	0.000	09:49
Control Test	0.079	09:50
Air Blank	0.000	09:51
Control Test	0.079	09:51
Air Blank	0.000	09:52
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

 Operator's Signature

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006385
 09/02/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.201	09:44
Air Blank	0.000	09:45
Control Test	0.199	09:46
Air Blank	0.000	09:46
Control Test	0.199	09:47
Air Blank	0.000	09:47
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel Std Dev(%)	0.5783	

SP

 Operator's Signature

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model: 8000 SN 80-006385
 09/02/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:56
Air Blank	0.000	09:57
Control Test	0.080	09:57
Air Blank	0.000	09:58
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

SP

 Operator's Signature