



INSTRUMENT PROCESSING SHEET

Agency Flagler County SOS/N 80-006616Florida Department of
Law EnforcementDate In 11-05-2020 DI Completion Date 11-06-2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Keyboard connector broken</u>	Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>222</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.539</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-22-2022</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG011102 04-22-2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.10</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP5088</u> ☒ 34°C +/- .2 Serial #: <u>MP5089</u> ☒ 34°C +/- .2 Serial #: <u>MP5090</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5088	201905A 05-14-2021																																																
0.080	MP5089	201905B 05-14-2021																																																
0.200	MP5090	201904D 04-30-2021																																																
0.080 DGS	N/A	AG011102 04-22-2022																																																
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.09 20:53:03 -05'00'	Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge <u>1018</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20420</td> <td>09-22-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06-05-2021</td> </tr> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A</td> <td>05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B</td> <td>05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D</td> <td>04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102</td> <td>04-22-2022</td> </tr> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20420	09-22-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	08819080A1	06-05-2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	05-14-2021	0.080	MP5089	201905B	05-14-2021	0.200	MP5090	201904D	04-30-2021	0.080 DGS	N/A	AG011102	04-22-2022
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	MP5091	N/A	N/A																																															
0.040	MP5082	20060	02-10-2022																																															
0.100	MP5083	20420	09-22-2022																																															
0.200	MP5084	20160	03-18-2022																																															
0.300	MP5085	20030	01-21-2022																																															
0.080 DGS	N/A	08819080A1	06-05-2021																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	MP5088	201905A	05-14-2021																																															
0.080	MP5089	201905B	05-14-2021																																															
0.200	MP5090	201904D	04-30-2021																																															
0.080 DGS	N/A	AG011102	04-22-2022																																															
Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1019</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </table> Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	Notes/Suggested Service: <u>Pressure off more than 10% during quality checks,</u> <u>performed optical bench cal adjust.</u>																																				
Simulator	Serial Number																																																	
0.000	MP5086																																																	
Interferent	MP5087																																																	
0.050	MP5088																																																	
0.080	MP5089																																																	
0.200	MP5090																																																	
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2020.11.06 15:36:11 -05'00' <i>Michael D. Hargrave</i> 2020.11.0 9:20:52:18 <i>CR</i> Tech Review / Date Admin Review / Date																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO
Time of Inspection: 12:42

Date of Inspection: 11/06/2020

Serial Number: 80-006616
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.049	0.078	0.200	0.080

Standard Deviations	0.0003	0.0003	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

11/06/2020
Date

MX
2020.11.
09
20:51:53
-05'00"

Stability Checks

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:08
Control Test	0.049	13:09
Air Blank	0.000	13:09
Control Test	0.049	13:10
Air Blank	0.000	13:10
Control Test	0.049	13:11
Air Blank	0.000	13:12
Control Test Stats		
Average	0.0450	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:13
Control Test	0.079	13:14
Air Blank	0.000	13:14
Control Test	0.078	13:15
Air Blank	0.000	13:16
Control Test	0.078	13:16
Air Blank	0.000	13:17
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

wet

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.199	13:23
Air Blank	0.000	13:24
Control Test	0.199	13:24
Air Blank	0.000	13:25
Control Test	0.199	13:26
Air Blank	0.000	13:26
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/05/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.082	13:30
Air Blank	0.000	13:30
Control Test	0.081	13:31
Air Blank	0.000	13:31
Control Test	0.081	13:31
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

Dry

Operator's Signature

2020.11.09
20:51:33 -05'00'

SR MX



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006616, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006616	UNCERTAINTY* $\pm$
Owning Agency:	FLAGLER COUNTY SO	0.050 g/ 210 L 0.004
Calibration Date:	11/06/2020	0.080 g/ 210 L 0.005
Calibration Time:	12:42	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/06/2020

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.11.09
20:51:14
-05'00'

Optical Bench Cal. Adj

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/06/2020 08:58:20

Auto Calibration
Max Power Res Value = 38
Auto Range Res Value = 20

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12918, 9um Io = 13158

CHANNEL 1

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.0530	(-0.0150)
Sample #2 =	0.0150	(0.0170)
Sample #3 =	0.0370	(0.0370)
Sample #4 =	0.0310	(0.0540)
Avg % Abs =	0.0277	(0.0360)
STD DEV =	0.0114	(0.0185)
REL STD DEV =	41.105	(51.445)

CHANNEL 2

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1550	(0.0040)
Sample #2 =	0.1420	(0.0090)
Sample #3 =	0.1510	(0.0250)
Sample #4 =	0.1420	(0.0270)
Avg % Abs =	0.1450	(0.0203)
STD DEV =	0.0052	(0.0099)
REL STD DEV =	3.584	(48.520)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12906, 9um Io = 13150

CHANNEL 1

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.7560	(-0.0280)
Sample #2 =	0.7230	(0.0050)
Sample #3 =	0.7610	(-0.0020)
Sample #4 =	0.7680	(0.0060)
Avg % Abs =	0.7507	(0.0030)
STD DEV =	0.0242	(0.0044)
REL STD DEV =	3.226	(145.297)

CHANNEL 2

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5490	(-0.0210)
Sample #2 =	1.5350	(-0.0150)
Sample #3 =	1.5530	(-0.0150)
Sample #4 =	1.5680	(-0.0200)
Avg % Abs =	1.5520	(-0.0167)
STD DEV =	0.0165	(0.0029)
REL STD DEV =	1.065	(17.321)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12900, 9um Io = 13146

CHANNEL 1

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.7580	(0.0010)
Sample #2 =	1.7580	(0.0090)
Sample #3 =	1.8040	(0.0020)
Sample #4 =	1.7860	(0.0350)
Avg % Abs =	1.7827	(0.0153)
STD DEV =	0.0232	(0.0174)
REL STD DEV =	1.300	(113.398)

CHANNEL 2

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.5500	(-0.0130)
Sample #2 =	3.5670	(-0.0150)
Sample #3 =	3.5590	(0.0000)
Sample #4 =	3.5360	(0.0100)
Avg % Abs =	3.5540	(-0.0017)
STD DEV =	0.0161	(0.0126)
REL STD DEV =	0.453	(754.983)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12892, 9um Io = 13143

CHANNEL 1

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.4830	(-0.0080)
Sample #2 =	3.4840	(-0.0020)
Sample #3 =	3.4820	(0.0000)
Sample #4 =	3.4980	(0.0000)
Avg % Abs =	3.4880	(-0.0007)
STD DEV =	0.0087	(0.0012)
REL STD DEV =	0.250	(173.205)

CHANNEL 2

Sample	% Abs	(% Abs Ref)
Sample #1 =	6.8060	(-0.0040)
Sample #2 =	6.8200	(0.0010)
Sample #3 =	6.8010	(0.0030)
Sample #4 =	6.8190	(0.0050)
Avg % Abs =	6.8133	(0.0030)
STD DEV =	0.0107	(0.0020)
REL STD DEV =	0.157	(66.667)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12889, 9um Io = 13141

CHANNEL 1

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.1280	(-0.0080)
Sample #2 =	5.1110	(0.0090)
Sample #3 =	5.0680	(0.0370)
Sample #4 =	5.0830	(0.0370)
Avg % Abs =	5.0873	(0.0277)
STD DEV =	0.0218	(0.0162)
REL STD DEV =	0.429	(58.431)

CHANNEL 2

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.8210	(-0.0080)
Sample #2 =	9.7960	(0.0170)
Sample #3 =	9.7790	(0.0290)
Sample #4 =	9.8010	(0.0230)
Avg % Abs =	9.7920	(0.0230)
STD DEV =	0.0115	(0.0060)
REL STD DEV =	0.118	(26.087)

***** AUTO CAL DATA *****

CHANNEL 1

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.028
Std Dev = 0.01 Rel Std Dev = 41.11
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.751
Std Dev = 0.02 Rel Std Dev = 3.23
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.783
Std Dev = 0.02 Rel Std Dev = 1.30
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.488
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.087
Std Dev = 0.02 Rel Std Dev = 0.43
Zero Order Coef = -71.89
First Order Coef = 2623.91
Second Order Coef = 38.64
Standard Deviation = 23.044161

CHANNEL 2

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.145
Std Dev = 0.01 Rel Std Dev = 3.58
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.552
Std Dev = 0.02 Rel Std Dev = 1.06
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.554
Std Dev = 0.02 Rel Std Dev = 0.45
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.813
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.792
Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -184.07
First Order Coef = 1325.91
Second Order Coef = 15.35
Standard Deviation = 29.125620

2020.11.
09
20:50:52
-05'00"

SR
MX

Optical bench cal. Adj.

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	-0.0003
0.100	0.099	0.0007
0.200	0.201	-0.0006
0.300	0.300	0.0002

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	-0.0001
0.100	0.099	0.0008
0.200	0.201	-0.0008
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 3535.00
Sample #2 = 3408.00
Sample #3 = 3440.00
Sample #4 = 3478.00
Average Result = 3442.0000
STD DEV = 35.1428
REL STD DEV = 1.018

***** CHANNEL 2
Sample #1 = 3329.00
Sample #2 = 3297.00
Sample #3 = 3299.00
Sample #4 = 3296.00
Average Result = 3297.3333
STD DEV = 1.5275
REL STD DEV = 0.046

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1021
3 um H2O Adjust (mg/l*10,000) = 367
9 um H2O Adjust (mg/l*10,000) = 512
**** AUTO CAL PASS

Post Stability Checks

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.049	10:31
Air Blank	0.000	10:32
Control Test	0.049	10:32
Air Blank	0.000	10:33
Control Test	0.048	10:33
Air Blank	0.000	10:34
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.200	10:41
Air Blank	0.000	10:41
Control Test	0.199	10:42
Air Blank	0.000	10:43
Control Test	0.199	10:43
Air Blank	0.000	10:44
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.077	10:36
Air Blank	0.000	10:37
Control Test	0.078	10:37
Air Blank	0.000	10:38
Control Test	0.077	10:38
Air Blank	0.000	10:39
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.081	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

wet

Dry

2020.11.
09
20:50:32
-05'00'
MX