



# INSTRUMENT PROCESSING SHEET

Agency FHP Orlandos/N 80-006628

Florida Department of Law Enforcement

Date In 06/12/2020DI Completion Date 6/28/20 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>See note attached.</u> <u>Printer cover missing, breath tube holder bent, and case severely scratched.</u>	<b>Quality Checks</b> Performed By <u>WJ</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>218</u> <u>201905A</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>218</u> <u>201905A</u> <u>05-17-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-17-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)  <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.4</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	<u>MP5088</u>	<u>218</u> <u>201905A</u> <u>05-17-2021</u>																																																
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-17-2021</u>																																																
0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>																																																
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																																
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.07.02 15:30:39 -04'00'	<b>Calibration Adjustment</b> Performed By _____ ID # _____ Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Notes/Suggested Service: _____ _____ _____ _____ _____	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 5px;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____  <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <table style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; border-bottom: 1px solid black;"><u>Michael D. Haughey</u></td> <td style="width:50%; border-bottom: 1px solid black;"><u>Brett Kirkland</u></td> </tr> <tr> <td style="font-size: small;">2020.07.02 12:42:03 -04'00'</td> <td style="font-size: small;">2020.07.02 15:24:09 -04'00'</td> </tr> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>	<u>Michael D. Haughey</u>	<u>Brett Kirkland</u>	2020.07.02 12:42:03 -04'00'	2020.07.02 15:24:09 -04'00'																																
Simulator	Serial Number																																																	
0.000	<u>MP5086</u>																																																	
Interferent	<u>MP5087</u>																																																	
0.050	<u>MP5088</u>																																																	
0.080	<u>MP5089</u>																																																	
0.200	<u>MP5090</u>																																																	
<u>Michael D. Haughey</u>	<u>Brett Kirkland</u>																																																	
2020.07.02 12:42:03 -04'00'	2020.07.02 15:24:09 -04'00'																																																	
	Tech Review / Date	Admin Review / Date																																																



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006628, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006628</u>	UNCERTAINTY* ±
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L
Calibration Date:	<u>06/28/2020</u>	0.080 g/ 210 L
Calibration Time:	<u>09:18</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/28/2020 Date  
Shayla Platt  
SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MH  
BK 2020.07.02  
15:26:04  
-0400'

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 09:18

Date of Inspection: 06/28/2020

Serial Number: 80-006628

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080

Standard Deviations	0.0003	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

*MH*

*BK* 2020.07.0  
2 15:27:19  
-04'00"

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

06/28/2020  
Date

# Stability Checks

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
06/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.049	08:50
Air Blank	0.000	08:51
Control Test	0.050	08:51
Air Blank	0.000	08:52
Control Test	0.050	08:53
Air Blank	0.000	08:53
Control Test Stats		
Average	0.0457	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	



Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
06/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.079	09:27
Air Blank	0.000	09:27
Control Test	0.079	09:28
Air Blank	0.000	09:28
Control Test	0.079	09:29
Air Blank	0.000	09:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
06/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.200	09:31
Air Blank	0.000	09:32
Control Test	0.199	09:32
Air Blank	0.000	09:33
Control Test	0.200	09:34
Air Blank	0.000	09:34
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	



Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
06/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:36
Control Test	0.081	09:37
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38
Control Test	0.080	09:38
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

MH

BK

Dry



Operator's Signature

80-006628 was in a patrol vehicle that was involved in a traffic crash.

The printer flew off and could not be located.

The cable for the laser jet printer was ripped loose and now the connection is missing the screw/nut.

The entire Intoxilyzer was bounced around in the vehicle

MS

BK 2020.07.0  
2 15:28:05  
-04'00'





INSTRUMENT PROCESSING SHEET

Agency FHP Orlando

S/N 80-006628

Florida Department of Law Enforcement

Date In 3/3/2020

DI Completion Date 3/16/20

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u>		<b>Quality Checks</b> Performed By <u>DP</u>		<b>Flow Calibration</b> Performed By _____																																																															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Middle right foot, back right foot, and front left foot missing.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>2.08</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																															
<b>Final Release Date</b>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>201905A</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201905B</u> <u>05-17-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1021</u>	<u>201905A</u> <u>05-17-2021</u>	0.080	<u>DR1275</u>	<u>201905B</u> <u>05-17-2021</u>	0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	<u>SD1021</u>	<u>201905A</u> <u>05-17-2021</u>																																																																	
0.080	<u>DR1275</u>	<u>201905B</u> <u>05-17-2021</u>																																																																	
0.200	<u>SD1011</u>	<u>201904D</u> <u>04-30-2021</u>																																																																	
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																																																	
<b>FDLE</b>  <b>MAR 16 2020</b>  <b>Alcohol Testing</b> Program		<b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1021</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1275</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1011</u>		<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1019</u> ID # <u>30793</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G8144</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>G2403</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>G2879</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>G3709</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>G8149</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>08819080A1</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1011 SP</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G8144</u>	N/A	N/A	0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>G2879</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>G8149</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1011 SP</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>DR1275</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>SD1011</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1024</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G8144</u></td> </tr> <tr> <td>Interferent</td> <td><u>DR3855</u></td> </tr> <tr> <td>0.050</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>G8144</u>	Interferent	<u>DR3855</u>	0.050	<u>SD1021</u>	0.080	<u>DR1275</u>	0.200	<u>SD1011</u>
Simulator	Serial Number	Lot Number	Expiration																																																																
0.000	<u>G8144</u>	N/A	N/A																																																																
0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>																																																																
0.100	<u>G2879</u>	<u>19160</u>	<u>7-9-21</u>																																																																
0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>																																																																
0.300	<u>G8149</u>	<u>19010</u>	<u>1-3-21</u>																																																																
0.080 DGS	N/A	<u>08819080A1</u>	<u>6-5-21</u>																																																																
Simulator	Serial Number	Lot Number	Expiration																																																																
0.050	<u>SD1011 SP</u>	<u>201905A</u>	<u>5-14-21</u>																																																																
0.080	<u>DR1275</u>	<u>201905B</u>	<u>5-14-21</u>																																																																
0.200	<u>SD1011</u>	<u>201904D</u>	<u>4-30-21</u>																																																																
0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>																																																																
Simulator	Serial Number																																																																		
0.000	<u>G8144</u>																																																																		
Interferent	<u>DR3855</u>																																																																		
0.050	<u>SD1021</u>																																																																		
0.080	<u>DR1275</u>																																																																		
0.200	<u>SD1011</u>																																																																		
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<u>SPM 3/16/20</u> <u>Scott Kirkland 3/16/2020</u> Tech Review / Date                      Admin Review / Date																																																															
Notes/Suggested Service: _____																																																																			



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006628, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006628</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/16/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:58</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Shayla Platt*

03/16/2020

Date

SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 January 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*RAM BK 3/16/2020*

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:58

Date of Inspection: 03/16/2020

Serial Number: 80-006628

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.046	0.076	0.196	0.081
0.000	0.047	0.077	0.197	0.081
0.000	0.048	0.078	0.197	0.081
0.000	0.048	0.078	0.197	0.081
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.081
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.079	0.197	0.081

Standard Deviations	0.0007	0.0009	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

GBM  
TBK  
3/16/2020

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

03/16/2020  
Date



# Stability Checks

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/13/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:33
Control Test	0.046	10:34
Air Blank	0.000	10:34
Control Test	0.046	10:35
Air Blank	0.000	10:35
Control Test	0.047	10:36
Air Blank	0.000	10:37
Control Test Stats		
Average	0.0463	
Std Dev	0.0006	
Rel Std Dev(%)	1.2461	

*SS*

Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/13/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:40
Control Test	0.075	10:40
Air Blank	0.000	10:41
Control Test	0.076	10:42
Air Blank	0.000	10:42
Control Test	0.076	10:43
Air Blank	0.000	10:43
Control Test Stats		
Average	0.0757	
Std Dev	0.0006	
Rel Std Dev(%)	0.7630	

wet

*SS*

Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/13/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.195	10:47
Air Blank	0.000	10:47
Control Test	0.196	10:48
Air Blank	0.000	10:49
Control Test	0.196	10:49
Air Blank	0.000	10:50
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

*SS*

Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/13/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.078	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:53
Control Test	0.078	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Dry

*SS*

Operator's Signature

*SSAM*  
*3/16/2020*

02/27/2020

80-006628 – Intoxilyzer 8000

Please return to:

Sergeant D.K. Hawkins

133 S. Semoran Boulevard, Suite A

Orlando, Florida 32807

Repairs Needed:

Missing feet on the bottom of the instrument

PKM  
BK  
3/16/2020

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = -0.004  
Std Dev = 0.02 Rel Std Dev = 610.04  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.721  
Std Dev = 0.02 Rel Std Dev = 2.68  
% Abs = 1.779  
Std Dev = 0.02 Rel Std Dev = 0.96  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.482  
Std Dev = 0.01 Rel Std Dev = 0.39  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.103  
Std Dev = 0.01 Rel Std Dev = 0.17  
Zero Order Coef = 11.38  
First Order Coef = 2598.15  
Second Order Coef = 38.92  
Standard Deviation = 4.633235

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.107  
Std Dev = 0.01 Rel Std Dev = 11.27  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.537  
Std Dev = 0.00 Rel Std Dev = 0.23  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.591  
Std Dev = 0.01 Rel Std Dev = 0.35  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.819  
Std Dev = 0.02 Rel Std Dev = 0.24  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.873  
Std Dev = 0.02 Rel Std Dev = 0.24  
Zero Order Coef = -146.71  
First Order Coef = 1315.08  
Second Order Coef = 14.90  
Standard Deviation = 7.226373

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.7860 (-0.0050)  
Sample #2 = 6.8060 (0.0010)  
Sample #3 = 6.8140 (0.0090)  
Sample #4 = 6.8370 (-0.0070)  
Avg % Abs = 6.8190 (0.0010)  
STD DEV = 0.0161 (0.0080)  
REL STD DEV = 0.236 (800.000)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12555, Sum Io = 12783  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7830 (-0.0310)  
Sample #2 = 1.7980 (-0.0140)  
Sample #3 = 1.7720 (0.0120)  
Sample #4 = 1.7660 (0.0230)  
Avg % Abs = 1.7878 (0.0070)  
STD DEV = 0.0170 (0.0190)  
REL STD DEV = 0.956 (271.429)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12551, Sum Io = 12780  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4280 (0.0050)  
Sample #2 = 3.4530 (0.0040)  
Sample #3 = 3.4670 (0.0140)  
Sample #4 = 3.4860 (0.0270)  
Avg % Abs = 3.4820 (0.0150)  
STD DEV = 0.0135 (0.0115)  
REL STD DEV = 0.386 (76.864)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12560, Sum Io = 12786  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.6800 (0.0040)  
Sample #2 = 0.7380 (-0.0060)  
Sample #3 = 0.7000 (0.0160)  
Sample #4 = 0.7250 (0.0070)  
Avg % Abs = 0.7210 (0.0057)  
STD DEV = 0.0193 (0.0111)  
REL STD DEV = 2.679 (195.184)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = -0.004  
Std Dev = 0.02 Rel Std Dev = 610.04  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.721  
Std Dev = 0.02 Rel Std Dev = 2.68  
% Abs = 1.779  
Std Dev = 0.02 Rel Std Dev = 0.96  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.482  
Std Dev = 0.01 Rel Std Dev = 0.39  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.103  
Std Dev = 0.01 Rel Std Dev = 0.17  
Zero Order Coef = 11.38  
First Order Coef = 2598.15  
Second Order Coef = 38.92  
Standard Deviation = 4.633235

<<<<< CHANNEL 2 >>>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.107  
Std Dev = 0.01 Rel Std Dev = 11.27  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.537  
Std Dev = 0.00 Rel Std Dev = 0.23  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.591  
Std Dev = 0.01 Rel Std Dev = 0.35  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.819  
Std Dev = 0.02 Rel Std Dev = 0.24  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.873  
Std Dev = 0.02 Rel Std Dev = 0.24  
Zero Order Coef = -146.71  
First Order Coef = 1315.08  
Second Order Coef = 14.90  
Standard Deviation = 7.226373

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.7860 (-0.0050)  
Sample #2 = 6.8060 (0.0010)  
Sample #3 = 6.8140 (0.0090)  
Sample #4 = 6.8370 (-0.0070)  
Avg % Abs = 6.8190 (0.0010)  
STD DEV = 0.0161 (0.0080)  
REL STD DEV = 0.236 (800.000)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12555, Sum Io = 12783  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7830 (-0.0310)  
Sample #2 = 1.7980 (-0.0140)  
Sample #3 = 1.7720 (0.0120)  
Sample #4 = 1.7660 (0.0230)  
Avg % Abs = 1.7878 (0.0070)  
STD DEV = 0.0170 (0.0190)  
REL STD DEV = 0.956 (271.429)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12551, Sum Io = 12780  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4280 (0.0050)  
Sample #2 = 3.4530 (0.0040)  
Sample #3 = 3.4670 (0.0140)  
Sample #4 = 3.4860 (0.0270)  
Avg % Abs = 3.4820 (0.0150)  
STD DEV = 0.0135 (0.0115)  
REL STD DEV = 0.386 (76.864)

<<<<< CHANNEL 2 >>>>>  
Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12560, Sum Io = 12786  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.6800 (0.0040)  
Sample #2 = 0.7380 (-0.0060)  
Sample #3 = 0.7000 (0.0160)  
Sample #4 = 0.7250 (0.0070)  
Avg % Abs = 0.7210 (0.0057)  
STD DEV = 0.0193 (0.0111)  
REL STD DEV = 2.679 (195.184)

Solution Stats Quadratic Fit Chan 1 :  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.000 -0.0000  
0.040 0.040 -0.0000  
0.100 0.100 0.0001  
0.200 0.200 -0.0001  
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2 :  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0001  
0.040 0.040 -0.0001  
0.100 0.100 -0.0001  
0.200 0.200 0.0002  
0.300 0.300 -0.0001

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1  
Sample #1 = 3459.00  
Sample #2 = 3504.00  
Sample #3 = 3376.00  
Sample #4 = 3427.00  
Average Result = 3435.6667  
STD DEV = 64.4386  
REL STD DEV = 1.876  
\*\*\*\*\* CHANNEL 2  
Sample #1 = 3278.00  
Sample #2 = 3287.00  
Sample #3 = 3261.00  
Sample #4 = 3267.00  
Average Result = 3271.6667  
STD DEV = 13.6137  
REL STD DEV = 0.416  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1019  
3 um H2O Adjust (mg/l\*10,000) = 374  
9 um H2O Adjust (mg/l\*10,000) = 538  
\*\*\*\* AUTO CAL PASS

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1  
Sample #1 = 3459.00  
Sample #2 = 3504.00  
Sample #3 = 3376.00  
Sample #4 = 3427.00  
Average Result = 3435.6667  
STD DEV = 64.4386  
REL STD DEV = 1.876  
\*\*\*\*\* CHANNEL 2  
Sample #1 = 3278.00  
Sample #2 = 3287.00  
Sample #3 = 3261.00  
Sample #4 = 3267.00  
Average Result = 3271.6667  
STD DEV = 13.6137  
REL STD DEV = 0.416  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1019  
3 um H2O Adjust (mg/l\*10,000) = 374  
9 um H2O Adjust (mg/l\*10,000) = 538  
\*\*\*\* AUTO CAL PASS

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12560, Sum Io = 12786  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.6800 (0.0040)  
Sample #2 = 0.7380 (-0.0060)  
Sample #3 = 0.7000 (0.0160)  
Sample #4 = 0.7250 (0.0070)  
Avg % Abs = 0.7210 (0.0057)  
STD DEV = 0.0193 (0.0111)  
REL STD DEV = 2.679 (195.184)

Bobm TSK 3/16/2020

# Post Cal Adjust Stability Checks # 80-006628

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/16/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.048	09:59
Air Blank	0.000	10:00
Control Test	0.048	10:01
Air Blank	0.000	10:01
Control Test	0.049	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP  
-----  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/16/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.078	10:04
Air Blank	0.000	10:05
Control Test	0.078	10:05
Air Blank	0.000	10:06
Control Test	0.078	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP  
-----  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/16/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.195	10:11
Air Blank	0.000	10:11
Control Test	0.196	10:12
Air Blank	0.000	10:12
Control Test	0.197	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

SP  
-----  
Operator's Signature

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006628  
03/16/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.081	09:46
Air Blank	0.000	09:47
Control Test	0.081	09:47
Air Blank	0.000	09:48
Control Test	0.080	09:48
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

SP  
-----  
Operator's Signature

OPM TSK 3/16/2020