



INSTRUMENT PROCESSING SHEET

Agency FHP Troop GS/N 80-006764Florida Department of
Law EnforcementDate In 11-06-2020 DI Completion Date 11-06-2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____		Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>137</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.535</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.17 13:56:15 -05'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG011102 04-20-2022	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.30</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/-2 Serial #: <u>MP5088</u> ☒ 34°C +/-2 Serial #: <u>MP5089</u> ☒ 34°C +/-2 Serial #: <u>MP5090</u>																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5088	201905A 05-14-2021																																																															
0.080	MP5089	201905B 05-14-2021																																																															
0.200	MP5090	201904D 04-30-2021																																																															
0.080 DGS	N/A	AG011102 04-20-2022																																																															
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number	Lot Number	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP5086																																																																
Interferent	MP5087																																																																
0.050	MP5088																																																																
0.080	MP5089																																																																
0.200	MP5090																																																																
Notes/Suggested Service: <u>Ran extra 0.080 wet stability to check seal on simulator.</u>		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																															
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place Into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2020.11.16 <u>Michael D. Haggerty</u> 12:14:50 -05'00' Tech Review / Date																																																															
		2020.11.17 <u>SR</u> 13:54:25 Admin Review / Date																																																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP G

Time of Inspection: 16:38

Date of Inspection: 11/06/2020

Serial Number: 80-006764

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#: 201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#: 201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG011102 Exp: 04/20/2022
0.000	0.047	0.078	0.195	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.077	0.196	0.077
0.000	0.048	0.077	0.196	0.077
0.000	0.048	0.078	0.195	0.077
0.000	0.047	0.078	0.195	0.077
0.000	0.047	0.077	0.195	0.077
0.000	0.047	0.077	0.196	0.077
0.000	0.047	0.077	0.195	0.077
0.000	0.048	0.077	0.195	0.077
Standard Deviations	0.0005	0.0005	0.0005	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

Signature and Printed Name

ISRAEL SOTO

11/06/2020
Date

MX
SR
2020.11.
17
13:54:44
-05'00"

Stability Checks

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006764
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:26
Control Test	0.047	14:27
Air Blank	0.000	14:27
Control Test	0.048	14:28
Air Blank	0.000	14:28
Control Test	0.047	14:29
Air Blank	0.000	14:30
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006764
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:32
Control Test	0.078	14:33
Air Blank	0.000	14:33
Control Test	0.078	14:34
Air Blank	0.000	14:34
Control Test	0.078	14:35
Air Blank	0.000	14:36
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet 1


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006764
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.195	14:38
Air Blank	0.000	14:38
Control Test	0.195	14:39
Air Blank	0.000	14:39
Control Test	0.195	14:40
Air Blank	0.000	14:41
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006764
11/06/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:42
Control Test	0.077	14:42
Air Blank	0.000	14:43
Control Test	0.077	14:43
Air Blank	0.000	14:43
Control Test	0.078	14:44
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Dry


Operator's Signature

FHP TROOP G
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006764
11/06/2020
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	15:05
Control Test	0.077	15:05
Air Blank	0.000	15:06
Control Test	0.078	15:06
Air Blank	0.000	15:07
Control Test	0.077	15:08
Air Blank	0.000	15:08
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

wet 2


Operator's Signature

MX

2020.11.
17
13:55:04
-05'00"



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006764, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006764</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP G</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/06/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>16:38</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/06/2020

Date

Israel SOTO
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.11.
17
13:55:24
-05'00"