



# INSTRUMENT PROCESSING SHEET

Agency Hillsborough County Sheriff's OfficeS/N 80-007158

Florida Department of Law Enforcement

Date In 09/14/2020DI Completion Date 09-22-2020 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>IS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>216</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>IS</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.84</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5090</u>																																		
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5088	201905A 05-14-2021																																																
0.080		201905B 05-14-2021																																																
0.200	MP5090	201904D 04-30-2021																																																
0.080 DGS		N/A	AG931603 11-12-2021																																															
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.09.25 08:35:21 -04'00'	<b>Calibration Adjustment</b> Performed By <u>IS</u> Barometric Pressure Gauge <u>1020</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03519080A4</td> <td>04-05-2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A</td> <td>05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B</td> <td>05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D</td> <td>04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603</td> <td>11-12-2021</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	03519080A4	04-05-2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP5088	201905A	05-14-2021	0.080	MP5089	201905B	05-14-2021	0.200	MP5090	201904D	04-30-2021	0.080 DGS	N/A	AG931603	11-12-2021
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	MP5091	N/A	N/A																																															
0.040	MP5082	20060	02-10-2022																																															
0.100	MP5083	20190	04-06-2022																																															
0.200	MP5084	20160	03-18-2022																																															
0.300	MP5085	20030	01-21-2022																																															
0.080 DGS	N/A	03519080A4	04-05-2021																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	MP5088	201905A	05-14-2021																																															
0.080	MP5089	201905B	05-14-2021																																															
0.200	MP5090	201904D	04-30-2021																																															
0.080 DGS	N/A	AG931603	11-12-2021																																															
Notes/Suggested Service: _____ _____ _____ _____ _____	<b>Department Inspection</b> Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"> <tr> <td style="width:50%; text-align: center;"> <u>Michael D. Haughey</u> 2020.09.24 10:58:59 -04'00'         </td> <td style="width:50%; text-align: center;">  2020.09.25 08:34:40 -04'00'         </td> </tr> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	<u>Michael D. Haughey</u> 2020.09.24 10:58:59 -04'00'	2020.09.25 08:34:40 -04'00'																																		
Simulator	Serial Number																																																	
0.000	MP5086																																																	
Interferent	MP5087																																																	
0.050	MP5088																																																	
0.080	MP5089																																																	
0.200	MP5090																																																	
<u>Michael D. Haughey</u> 2020.09.24 10:58:59 -04'00'	2020.09.25 08:34:40 -04'00'																																																	
	Tech Review / Date	Admin Review / Date																																																

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 12:04

Date of Inspection: 09/22/2020

Serial Number: 80-007158  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.048	0.079	0.200	0.000 / 0.080
0.000	0.048	0.078	0.200	0.000 / 0.080
0.000	0.048	0.079	0.200	0.000 / 0.080
0.000	0.049	0.078	0.200	0.000 / 0.080
0.000	0.048	0.079	0.200	0.000 / 0.080
0.000	0.049	0.078	0.200	0.000 / 0.080
0.000	0.048	0.079	0.200	0.000 / 0.079
0.000	0.049	0.079	0.200	0.000 / 0.080
0.000	0.048	0.079	0.200	0.000 / 0.080
0.000	0.048	0.079	0.200	0.000 / 0.080
Standard Deviations	0.0004	0.0004	0.0000	0.0000 / 0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance DRY GAS INLET NOT PLUGGED IN .

*MX*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*SP* 2020.09.  
25  
08:34:01  
-04'00'

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

09/22/2020  
Date

# Stability Checks

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007158  
09/15/2020  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007158  
09/15/2020  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007158  
09/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:58
Control Test	0.047	11:59
Air Blank	0.000	11:59
Control Test	0.046	12:00
Air Blank	0.000	12:01
Control Test	0.047	12:01
Air Blank	0.000	12:02
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.077	12:04
Air Blank	0.000	12:04
Control Test	0.077	12:05
Air Blank	0.000	12:05
Control Test	0.078	12:06
Air Blank	0.000	12:07
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Test	g/210L	Time
Air Blank	0.000	12:07
Control Test	0.198	12:08
Air Blank	0.000	12:09
Control Test	0.197	12:09
Air Blank	0.000	12:10
Control Test	0.197	12:11
Air Blank	0.000	12:11
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	


wet



Operator's Signature



Operator's Signature




Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007158  
09/15/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.079	12:22
Air Blank	0.000	12:22
Control Test	0.080	12:23
Air Blank	0.000	12:23
Control Test	0.079	12:23
Air Blank	0.000	12:24
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Dry



Operator's Signature

MX

2020.09.  
25  
08:33:41  
-04'00"



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007158, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007158</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>09/22/2020</u>	0.080 g/210 L	0.005
Calibration Time:	<u>12:04</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date 09/22/2020  
  
ISRAEL SOTO,  
Department Inspector

2020.09.  
25  
08:33:19  
-04'00'

# Optical Bench Cal Adjust

HILLSBOROUGH CO SD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007158  
09/22/2020 08:24:05

Auto Calibration  
Max Power Res Value = 73  
Auto Range Res Value = 47

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12352, Sum Io = 13282

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1590 (-0.0060)  
Sample #2 = 0.1710 (0.0310)  
Sample #3 = 0.1260 (0.0890)  
Sample #4 = 0.1430 (0.1260)  
Avg % Abs = 0.1467 (0.0820)  
STD DEV = 0.0227 (0.0479)  
REL STD DEV = 15.493 (58.397)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1180 (-0.0110)  
Sample #2 = 0.1420 (-0.0100)  
Sample #3 = 0.1130 (0.0270)  
Sample #4 = 0.1210 (0.0410)  
Avg % Abs = 0.1253 (0.0193)  
STD DEV = 0.0150 (0.0264)  
REL STD DEV = 11.950 (136.294)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12330, Sum Io = 13273

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8850 (-0.0010)  
Sample #2 = 0.9230 (0.0170)  
Sample #3 = 0.9010 (0.0470)  
Sample #4 = 0.9290 (0.0680)  
Avg % Abs = 0.9177 (0.0440)  
STD DEV = 0.0147 (0.0256)  
REL STD DEV = 1.606 (58.255)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.4490 (0.0020)  
Sample #2 = 1.4630 (0.0030)  
Sample #3 = 1.4540 (0.0090)  
Sample #4 = 1.4510 (0.0270)  
Avg % Abs = 1.4560 (0.0130)  
STD DEV = 0.0062 (0.0125)  
REL STD DEV = 0.429 (96.077)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12316, Sum Io = 13267

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 2.0700 (-0.0190)  
Sample #2 = 1.9980 (0.0580)  
Sample #3 = 2.0500 (0.0580)  
Sample #4 = 2.0530 (0.0820)  
Avg % Abs = 2.0337 (0.0660)  
STD DEV = 0.0309 (0.0139)  
REL STD DEV = 1.521 (20.995)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4350 (-0.0060)  
Sample #2 = 3.3870 (0.0270)  
Sample #3 = 3.4310 (0.0130)  
Sample #4 = 3.3940 (0.0450)  
Avg % Abs = 3.4040 (0.0283)  
STD DEV = 0.0236 (0.0160)  
REL STD DEV = 0.695 (56.617)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12300, Sum Io = 13260

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.8480 (-0.0110)  
Sample #2 = 3.8810 (0.0130)  
Sample #3 = 3.8200 (0.0520)  
Sample #4 = 3.8810 (0.0370)  
Avg % Abs = 3.8607 (0.0340)  
STD DEV = 0.0352 (0.0197)  
REL STD DEV = 0.912 (57.860)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.5380 (-0.0030)  
Sample #2 = 6.5290 (0.0140)  
Sample #3 = 6.5110 (0.0270)  
Sample #4 = 6.5340 (0.0340)  
Avg % Abs = 6.5247 (0.0250)  
STD DEV = 0.0121 (0.0101)  
REL STD DEV = 0.185 (40.596)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12291, Sum Io = 13253

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.6100 (-0.0130)  
Sample #2 = 5.5850 (0.0300)  
Sample #3 = 5.5700 (0.0650)  
Sample #4 = 5.5830 (0.0570)  
Avg % Abs = 5.5793 (0.0507)  
STD DEV = 0.0081 (0.0183)  
REL STD DEV = 0.146 (36.786)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 9.4030 (-0.0020)  
Sample #2 = 9.4090 (0.0070)  
Sample #3 = 9.3690 (0.0440)  
Sample #4 = 9.3920 (0.0290)  
Avg % Abs = 9.3900 (0.0267)  
STD DEV = 0.0201 (0.0186)  
REL STD DEV = 0.214 (69.788)

MX

# Optical bench Cal. Adj.

```

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.147
Std Dev = 0.02 Rel Std Dev = 15.49
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.918
Std Dev = 0.01 Rel Std Dev = 1.61
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.034
Std Dev = 0.03 Rel Std Dev = 1.52
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.861
Std Dev = 0.04 Rel Std Dev = 0.91
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.579
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -352.89
First Order Coef = 2434.31
Second Order Coef = 33.68
Standard Deviation = 17.896711

```

```

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.125
Std Dev = 0.01 Rel Std Dev = 11.95
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.456
Std Dev = 0.01 Rel Std Dev = 0.43
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.404
Std Dev = 0.02 Rel Std Dev = 0.69
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.525
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.390
Std Dev = 0.02 Rel Std Dev = 0.21
Zero Order Coef = -155.33
First Order Coef = 1380.61
Second Order Coef = 16.59
Standard Deviation = 25.619497

```

```

-----
; Solution Stats Quadratic Fit Chan 1
; Act      Fit      Residual
; g/210L   g/210L   g/210L
; 0.000    0.000    -0.0001
; 0.040    0.040    -0.0001
; 0.100    0.099    0.0005
; 0.200    0.200    -0.0005
; 0.300    0.300    0.0002
-----

```

```

-----
; Solution Stats Quadratic Fit Chan 2
; Act      Fit      Residual
; g/210L   g/210L   g/210L
; 0.000    0.000    -0.0004
; 0.040    0.040    0.0003
; 0.100    0.099    0.0005
; 0.200    0.201    -0.0007
; 0.300    0.300    0.0003
-----

```

```

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 2937.00
Sample #2 = 2839.00
Sample #3 = 2891.00
Sample #4 = 2832.00
Average Result = 2854.0000
STD DEV = 32.2335
REL STD DEV = 1.129

```

```


***** CHANNEL 2
Sample #1 = 3324.00
Sample #2 = 3308.00
Sample #3 = 3342.00
Sample #4 = 3304.00
Average Result = 3318.0000
STD DEV = 20.8806
REL STD DEV = 0.629

```

```

*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1020
3 um H2O Adjust (mg/lx10,000) = 955
9 um H2O Adjust (mg/lx10,000) = 491
***** AUTO CAL PASS

```


 2020.09.  
 25  
 08:32:31  
 -04'00'



# Post Cal Adjust Stability Checks

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007158  
 09/22/2020  
 Software: 8100.27

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007158  
 09/22/2020  
 Software: 8100.27

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007158  
 09/22/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	0.048	09:30
Air Blank	0.000	09:30
Control Test	0.048	09:31
Air Blank	0.000	09:32
Control Test	0.049	09:32
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.079	09:35
Air Blank	0.000	09:35
Control Test	0.079	09:36
Air Blank	0.000	09:36
Control Test	0.078	09:37
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.199	09:39
Air Blank	0.000	09:40
Control Test	0.199	09:40
Air Blank	0.000	09:41
Control Test	0.199	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature



Operator's Signature



Operator's Signature

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007158  
 09/22/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Control Test	0.080	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry



Operator's Signature

2020.09.  
 25  
 08:32:07  
 -04'00"

