



INSTRUMENT PROCESSING SHEET

Agency Florida Wildlife Conservation Commission S/N 80-007168Florida Department of
Law EnforcementDate In 11/3/2020 DI Completion Date 11/4/2020 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.05 10:40:31 -05'00'	Quality Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>233</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.158</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 1/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 1/30/2022	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>23.19C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3969</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201905A 05/14/2021																																																											
0.080	SD3968	201905B 05/14/2021																																																											
0.200	SD3969	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG003005 1/30/2022																																																											
Calibration Adjustment Performed By <u>DERR</u> Barometric Pressure Gauge <u>1020</u> ID # <u>68639</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>20190</td> <td>04/06/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005</td> <td>01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	20060	02/10/2022	0.100	MP5099	20190	04/06/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	SD3967	201905A	05/14/2021	0.080	SD3968	201905B	05/14/2021	0.200	SD3969	201904D	04/30/2021	0.080 DGS	N/A	AG003005	01/30/2022	Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use _____ 2020.11.0 <i>Denial Soto</i> 11-04-2020 <i>SA</i> 5 10:39:01 Tech Review / Date Admin Review / Date	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	MP5095	N/A	N/A																																																										
0.040	MP5098	20060	02/10/2022																																																										
0.100	MP5099	20190	04/06/2022																																																										
0.200	MP5100	20160	03/18/2022																																																										
0.300	MP5101	20030	01/21/2022																																																										
0.080 DGS	N/A	08819080A1	06/05/2021																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	SD3967	201905A	05/14/2021																																																										
0.080	SD3968	201905B	05/14/2021																																																										
0.200	SD3969	201904D	04/30/2021																																																										
0.080 DGS	N/A	AG003005	01/30/2022																																																										
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3968																																																												
0.200	SD3969																																																												
Notes/Suggested Service: <u>Optical calibration to bring</u> <u>values closer to nominal, first attempt quadratic</u> <u>fit on channel 1 showed a 0.0026 for the 0.040</u> <u>so another calibration was required.</u> _____ _____ _____																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007168, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007168</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/04/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>00:08</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/04/2020

Date

David E Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.11.05 10:38:29 -05'00'

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC

Time of Inspection: 00:08

Date of Inspection: 11/04/2020

Serial Number: 80-007168

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079

Standard Deviations	0.0005	0.0005	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

DS

CR 2020.11.
05
10:37:58
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

11/04/2020
Date

0.8

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities 2	80-007168	Florida Wildlife Conservation Commission	11/3/2020	DERR <i>Full</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>FWCC Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-007168</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>21:28</td></tr><tr><td>Control Test</td><td>0.049</td><td>21:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:29</td></tr><tr><td>Control Test</td><td>0.049</td><td>21:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:30</td></tr><tr><td>Control Test</td><td>0.049</td><td>21:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	21:28	Control Test	0.049	21:28	Air Blank	0.000	21:29	Control Test	0.049	21:30	Air Blank	0.000	21:30	Control Test	0.049	21:31	Air Blank	0.000	21:31	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>FWCC Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-007168</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>21:32</td></tr><tr><td>Control Test</td><td>0.079</td><td>21:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:34</td></tr><tr><td>Control Test</td><td>0.079</td><td>21:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:35</td></tr><tr><td>Control Test</td><td>0.079</td><td>21:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:36</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	21:32	Control Test	0.079	21:33	Air Blank	0.000	21:34	Control Test	0.079	21:34	Air Blank	0.000	21:35	Control Test	0.079	21:35	Air Blank	0.000	21:36	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>FWCC Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-007168</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>21:37</td></tr><tr><td>Control Test</td><td>0.198</td><td>21:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:38</td></tr><tr><td>Control Test</td><td>0.198</td><td>21:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:40</td></tr><tr><td>Control Test</td><td>0.199</td><td>21:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	21:37	Control Test	0.198	21:38	Air Blank	0.000	21:38	Control Test	0.198	21:39	Air Blank	0.000	21:40	Control Test	0.199	21:40	Air Blank	0.000	21:41	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		<div>FWCC Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-007168</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>21:42</td></tr><tr><td>Control Test</td><td>0.079</td><td>21:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>21:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>21:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	21:42	Control Test	0.079	21:43	Air Blank	0.000	21:43	Control Test	0.080	21:43	Air Blank	0.000	21:44	Control Test	0.080	21:44	Air Blank	0.000	21:45	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	21:28																																																																																																																																																	
Control Test	0.049	21:28																																																																																																																																																	
Air Blank	0.000	21:29																																																																																																																																																	
Control Test	0.049	21:30																																																																																																																																																	
Air Blank	0.000	21:30																																																																																																																																																	
Control Test	0.049	21:31																																																																																																																																																	
Air Blank	0.000	21:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	21:32																																																																																																																																																	
Control Test	0.079	21:33																																																																																																																																																	
Air Blank	0.000	21:34																																																																																																																																																	
Control Test	0.079	21:34																																																																																																																																																	
Air Blank	0.000	21:35																																																																																																																																																	
Control Test	0.079	21:35																																																																																																																																																	
Air Blank	0.000	21:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	21:37																																																																																																																																																	
Control Test	0.198	21:38																																																																																																																																																	
Air Blank	0.000	21:38																																																																																																																																																	
Control Test	0.198	21:39																																																																																																																																																	
Air Blank	0.000	21:40																																																																																																																																																	
Control Test	0.199	21:40																																																																																																																																																	
Air Blank	0.000	21:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	21:42																																																																																																																																																	
Control Test	0.079	21:43																																																																																																																																																	
Air Blank	0.000	21:43																																																																																																																																																	
Control Test	0.080	21:43																																																																																																																																																	
Air Blank	0.000	21:44																																																																																																																																																	
Control Test	0.080	21:44																																																																																																																																																	
Air Blank	0.000	21:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
<div>Operator's Signature</div>	<div>Operator's Signature</div>	<div>Operator's Signature</div>	<div>Operator's Signature</div>																																																																																																																																																

2020.11.05
10:36:58
-05'00'

```

<<<<< CHANNEL 2 >>>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.4560   (-0.0090)
Sample #2 = 1.4670   (-0.0130)
Sample #3 = 1.4690   (-0.0070)
Sample #4 = 1.4630   (0.0010)
Avg % Abs = 1.4663   (-0.0063)
STD DEV = 0.0031 (0.0070)
REL STD DEV = 0.208 (110.902)

FWCC
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-007168
11/03/2020 20:35:35

```

Auto Calibration
Max Power Res Value = 74
Auto Range Res Value = 53

```

Sol Value = 0.000 g/210L xxx
Fit value = 0.0000 mg/l xxxx
Samples Taken = 4, Discarded = 1
Sum Io = 12726, Sum Io = 13066
<<<< CHANNEL 1 >>>>

```

```
REL STD DEV = 24.037 (43.594)

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 0.1470 (-0.0020)
Sample #2 = 0.1410 (-0.0120)
Sample #3 = 0.1300 (-0.0230)
Sample #4 = 0.1170 (-0.0360)
Avg % Abs = 0.1293 (-0.0237)
STD DEV = 0.0120 (-0.0120)
REL STD DEV = 9.289 (50.763)
```

```
Sol Value = 0.040 g/210L xxx
Fit value = 0.1905 mg/l zzzz
Samples Taken = 4, Discarded = 1
Sum lo = 12708, Sum lo = 13061
<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 0.7870      (-0.0080)
Sample #2 = 0.7900      (-0.0020)
Sample #3 = 0.8490      (-0.0150)
Sample #4 = 0.8280      (-0.0100)
Avg % Abs = 0.8223      (-0.0090)
STD DEV = 0.0299      (0.0066)
REL STD DEV = 3.637      (72.86)
```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 6.6280    (0.0110)
Sample #2 = 6.6050    (0.0260)
Sample #3 = 6.6110    (0.0280)
Sample #4 = 6.6170    (0.0220)
Avg % Abs = 6.6110   (0.0253)
STD DEV = 0.0060 (0.0031)
REL STD DEV = 0.091 (12.059)

```

```
Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %Z%
Samples Taken = 4, Discarded = 1
Sum lo = 12698, Sum lo = 13055
<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Rel
Sample #1 = 5.3310 (-0.0080)
Sample #2 = 5.3710 (-0.0160)
Sample #3 = 5.3210 (0.0100)
Sample #4 = 5.3770 (-0.0140)
Avg % Abs = 5.3563 (-0.0067)
STD DEV = 0.0307 (0.0145)
REL STD DEV = 0.574 (217.025)
```

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.5360	(0.0000)
Sample #2 =	9.5600	(-0.0160)
Sample #3 =	9.5370	(-0.0010)
Sample #4 =	9.3550	(-0.0050)
Avg % Abs =	9.5440	(-0.0073)
STD DEV =	0.0139	(0.0078)
REL STD DEV =	0.146	(0.05,920)

Optical Calibration 2
SN: 80-007168
Agency: FWCC
Date: 11/3 /2020
Quadratic Fit: +/- 0.002g/210L
By: DERR <i>WCC</i>

2020.11.0
5 10:36:23
-05'00'

```

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.03 Rel Std Dev = 24.04

Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.822
Std Dev = 0.03 Rel Std Dev = 3.64

Sol Val = 0.4752 mg/l or 0.100 g/210L

```

```

Std Dev = 0.03 Rel Std Dev = 1.4E
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Rbs = 3.678
Std Dev = 0.02 Rel Std Dev = 0.63
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Rbs = 5.356
Std Dev = 0.03 Rel Std Dev = 0.57
Zero Order Coef = -245.02
First Order Coef = 2547.49
Second Order Coef = 30.61
Standard Deviation = 22.359835
-----
<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 a/210L

```

Sol. Val = 0.1905 mg/l or 0.040 g/210L
% Rbs = 1.466
Std Dev = 0.00 Rel Std Dev = 0.21
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Rbs = 3.456
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Rbs = 6.611
Std Dev = 0.01 Rel Std Dev = 0.09
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Rbs = 9.544
Std Dev = 0.01 Rel Std Dev = 0.15
Zero Order Coef = -159.00
First Order Coef = 1368.67
Second Order Coef = 15.05
Standard Deviation = 20.766731

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0005
0.040	0.039	0.0007
0.100	0.100	-0.0001
0.200	0.200	-0.0003
0.300	0.300	0.0002

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.040	0.0005
0.100	0.100	0.0002
0.200	0.201	-0.0005
0.300	0.300	0.0002

```
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3112.00
Sample #2 = -225.00
Sample #3 = -212.00
Sample #4 = -207.00
Average Result = -214.6667
STD DEV = 9.2916
REL STD DEV = 4.328
*****
```

```
***** CHANNEL 2
Sample #1 = 3354.00
Sample #2 = -149.00
Sample #3 = -133.00
Sample #4 = -143.00
Average Result = -141.6667
STD DEV = 8.0829
REL STD DEV = 5.706
```

```

*****
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l  %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3136.00
Sample #2 = 3128.00
Sample #3 = 3133.00
Sample #4 = 3071.00

```

```
Average Result = 3110.6667
STD DEV = 34.4432
REL STD DEV = 1.107

*****
***** CHANNEL 2
Sample #1 = 3362.00
Sample #2 = 3362.00
Sample #3 = 3380.00
Sample #4 = 3346.00
```

```
Average Result = 3362.6667
STD DEV = 17.0098
REL STD DEV = 0.506

*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1020
3 um H2O Adjust (mg/lx10,000)
9 um H2O Adjust (mg/lx10,000)
```


FWCC

Intoxilyzer - Alcohol Analyzer

Model 8000 SN 80-007168

11/03/2020 19:39:28

Auto Calibration

Max Power Res Value = 74

Auto Range Res Value = 52

Sol Value = 0.000 g/210L xxx

Fit value = 0.0000 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

Sum Io = 12571, Sum Io = 12389

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1510 (0.0000)

Sample #2 = 0.1190 (0.0400)

Sample #3 = 0.1500 (0.0560)

Sample #4 = 0.0840 (0.1060)

Avg % Abs = 0.1177 (0.0673)

STD DEV = 0.0330 (0.0344)

REL STD DEV = 28.062 (51.132)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1850 (0.0070)

Sample #2 = 0.1580 (0.0210)

Sample #3 = 0.1770 (0.0150)

Sample #4 = 0.1170 (0.0440)

Avg % Abs = 0.1507 (0.0267)

STD DEV = 0.0307 (0.0153)

REL STD DEV = 20.353 (57.405)

Sol Value = 0.040 g/210L xxx

Fit value = 0.1905 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

Sum Io = 12556, Sum Io = 12384

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8480 (-0.0120)

Sample #2 = 0.7770 (0.0270)

Sample #3 = 0.7860 (0.0180)

Sample #4 = 0.7890 (0.0490)

Avg % Abs = 0.7840 (0.0313)

STD DEV = 0.0062 (0.0153)

REL STD DEV = 0.797 (50.897)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.4490 (-0.0020)

Sample #2 = 1.4270 (0.0070)

Sample #3 = 1.4100 (0.0070)

Sample #4 = 1.3560 (0.0280)

Avg % Abs = 1.4110 (0.0140)

STD DEV = 0.0155 (0.0121)

REL STD DEV = 1.100 (86.603)

Sol Value = 0.100 g/210L xxx

Fit value = 0.4762 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

Sum Io = 12550, Sum Io = 12380

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 1.9550 (-0.0030)

Sample #2 = 1.9420 (0.0020)

Sample #3 = 1.9160 (0.0200)

Sample #4 = 1.9630 (0.0050)

Avg % Abs = 1.9403 (0.0090)

STD DEV = 0.0235 (0.0096)

REL STD DEV = 1.213 (107.152)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.4920 (-0.0160)

Sample #2 = 3.4740 (0.0070)

Sample #3 = 3.4470 (0.0120)

Sample #4 = 3.4580 (-0.0010)

Avg % Abs = 3.4730 (0.0060)

STD DEV = 0.0255 (0.0066)

REL STD DEV = 0.735 (109.291)

Sol Value = 0.200 g/210L xxx

Fit value = 0.9524 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

Sum Io = 12547, Sum Io = 12379

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.7240 (-0.0030)

Sample #2 = 3.7170 (0.0040)

Sample #3 = 3.6580 (0.0400)

Sample #4 = 3.7060 (0.0230)

Avg % Abs = 3.6937 (0.0223)

STD DEV = 0.0314 (0.0190)

REL STD DEV = 0.849 (80.638)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 6.6630 (-0.0170)

Sample #2 = 6.6370 (0.0040)

Sample #3 = 6.6180 (0.0230)

Sample #4 = 6.6630 (0.0010)

Avg % Abs = 6.6393 (0.0093)

STD DEV = 0.0226 (0.0119)

REL STD DEV = 0.340 (127.825)

Sol Value = 0.300 g/210L xxx

Fit value = 1.4286 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

Sum Io = 12544, Sum Io = 12377

<<<< CHANNEL 1 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 5.3980 (-0.0170)

Sample #2 = 5.3810 (0.0050)

Sample #3 = 5.3740 (0.0230)

Sample #4 = 5.3760 (0.0210)

Avg % Abs = 5.3770 (0.0163)

STD DEV = 0.0036 (0.0099)

REL STD DEV = 0.067 (60.403)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)

Sample #1 = 9.5690 (-0.0120)

Sample #2 = 9.5290 (0.0120)

Sample #3 = 9.5470 (0.0080)

Sample #4 = 9.5560 (0.0070)

Avg % Abs = 9.5440 (0.0090)

STD DEV = 0.0137 (0.0026)

REL STD DEV = 0.144 (29.397)

Optical Calibration

SN: 80-007168

Agency: FWCC

Date: 11/3 /2020

Quadratic Fit: +/- 0.002g/210L

By: DERR *ML*

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.118

Std Dev = 0.03 Rel Std Dev = 28.06

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.784

Std Dev = 0.01 Rel Std Dev = 0.80

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.940

Std Dev = 0.02 Rel Std Dev = 1.21

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.694

Std Dev = 0.03 Rel Std Dev = 0.85

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.377

Std Dev = 0.00 Rel Std Dev = 0.07

Zero Order Coef = -222.90

First Order Coef = 2530.77

Second Order Coef = 30.72

Standard Deviation = 76.631073

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.151

Std Dev = 0.03 Rel Std Dev = 21.35

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.411

Std Dev = 0.02 Rel Std Dev = 1.10

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.473

Std Dev = 0.03 Rel Std Dev = 0.73

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.639

Std Dev = 0.02 Rel Std Dev = 0.34

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.544

Std Dev = 0.01 Rel Std Dev = 0.14

Zero Order Coef = -138.21

First Order Coef = 1356.54

Second Order Coef = 15.97

Standard Deviation = 64.657196

***** CHANNEL 1 *****

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 0.002 -0.0016

0.040 0.037 0.0026

0.100 0.101 -0.0009

0.200 0.200 -0.0004

0.300 0.300 0.0003

***** CHANNEL 2 *****

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 0.001 -0.0014

0.040 0.038 0.0020

0.100 0.100 -0.0001

0.200 0.201 -0.0010

0.300 0.300 0.0005

Sol Value = 0.080 g/210L xxx

Fit value = 0.3810 mg/l %ZZZ

Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****

Sample #1 = 3115.00

Sample #2 = 3115.00

Sample #3 = 3126.00

Sample #4 = 3033.00

Average Result = 3091.3333

STD DEV = 50.8167

REL STD DEV = 1.644

***** CHANNEL 2 *****

Sample #1 = 3346.00

Sample #2 = 3343.00

Sample #3 = 3343.00

Sample #4 = 3336.00

Average Result = 3340.6667

STD DEV = 4.0415

REL STD DEV = 0.121

***** CHANNEL 1 *****

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1020

3 um H2O Adjust (mg/l x 10,000) = 718

9 um H2O Adjust (mg/l x 10,000) = 469

***** CHANNEL 2 *****

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1020

3 um H2O Adjust (mg/l x 10,000) = 718

9 um H2O Adjust (mg/l x 10,000) = 469

2020.11

.05

10:36:00

0-05'00'

2020.11.05
 10:35:40
 Perf By  DERR 

Type of Test	Serial Number	Agency	Date
Stabilities	80-007168	Florida Wildlife Conservation Commission	11/3/2020

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>FMCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007168 11/03/2020 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>19:12</td></tr><tr><td>Control Test</td><td>0.046</td><td>19:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:13</td></tr><tr><td>Control Test</td><td>0.047</td><td>19:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:14</td></tr><tr><td>Control Test</td><td>0.047</td><td>19:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:15</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0467</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2372</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	19:12	Control Test	0.046	19:12	Air Blank	0.000	19:13	Control Test	0.047	19:13	Air Blank	0.000	19:14	Control Test	0.047	19:15	Air Blank	0.000	19:15	Control Test Stats			Average	0.0467		Std Dev	0.0006		Rel Std Dev(%)	1.2372		<div>FMCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007168 11/03/2020 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>19:17</td></tr><tr><td>Control Test</td><td>0.078</td><td>19:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:18</td></tr><tr><td>Control Test</td><td>0.077</td><td>19:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:19</td></tr><tr><td>Control Test</td><td>0.077</td><td>19:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	19:17	Control Test	0.078	19:18	Air Blank	0.000	19:18	Control Test	0.077	19:19	Air Blank	0.000	19:19	Control Test	0.077	19:20	Air Blank	0.000	19:21	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466		<div>FMCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007168 11/03/2020 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>19:23</td></tr><tr><td>Control Test</td><td>0.195</td><td>19:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:24</td></tr><tr><td>Control Test</td><td>0.196</td><td>19:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:25</td></tr><tr><td>Control Test</td><td>0.196</td><td>19:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:26</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	19:23	Control Test	0.195	19:23	Air Blank	0.000	19:24	Control Test	0.196	19:25	Air Blank	0.000	19:25	Control Test	0.196	19:26	Air Blank	0.000	19:26	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		<div>FMCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007168 11/03/2020 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>19:29</td></tr><tr><td>Control Test</td><td>0.077</td><td>19:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:30</td></tr><tr><td>Control Test</td><td>0.077</td><td>19:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:31</td></tr><tr><td>Control Test</td><td>0.078</td><td>19:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	19:29	Control Test	0.077	19:29	Air Blank	0.000	19:30	Control Test	0.077	19:30	Air Blank	0.000	19:31	Control Test	0.078	19:31	Air Blank	0.000	19:31	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:12																																																																																																																																																	
Control Test	0.046	19:12																																																																																																																																																	
Air Blank	0.000	19:13																																																																																																																																																	
Control Test	0.047	19:13																																																																																																																																																	
Air Blank	0.000	19:14																																																																																																																																																	
Control Test	0.047	19:15																																																																																																																																																	
Air Blank	0.000	19:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0467																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2372																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:17																																																																																																																																																	
Control Test	0.078	19:18																																																																																																																																																	
Air Blank	0.000	19:18																																																																																																																																																	
Control Test	0.077	19:19																																																																																																																																																	
Air Blank	0.000	19:19																																																																																																																																																	
Control Test	0.077	19:20																																																																																																																																																	
Air Blank	0.000	19:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:23																																																																																																																																																	
Control Test	0.195	19:23																																																																																																																																																	
Air Blank	0.000	19:24																																																																																																																																																	
Control Test	0.196	19:25																																																																																																																																																	
Air Blank	0.000	19:25																																																																																																																																																	
Control Test	0.196	19:26																																																																																																																																																	
Air Blank	0.000	19:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:29																																																																																																																																																	
Control Test	0.077	19:29																																																																																																																																																	
Air Blank	0.000	19:30																																																																																																																																																	
Control Test	0.077	19:30																																																																																																																																																	
Air Blank	0.000	19:31																																																																																																																																																	
Control Test	0.078	19:31																																																																																																																																																	
Air Blank	0.000	19:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		
<div>Signature</div> <div>Operator's Signature</div>	<div>Signature</div> <div>Operator's Signature</div>	<div>Signature</div> <div>Operator's Signature</div>	<div>Signature</div> <div>Operator's Signature</div>																																																																																																																																																