



INSTRUMENT PROCESSING SHEET

Agency Suwannee County Sheriff's Office s/N 80-007382

Florida Department of Law Enforcement Date In 09/17/2020 DI Completion Date 09-22-2020 Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input checked="" type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>IS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>239</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP5088	201905A 05-14-2021																																																													
0.080	MP5089	201905B 05-14-2021																																																													
0.200	MP5090	201904D 04-30-2021																																																													
0.080 DGS	N/A	AG931603 11-12-2021																																																													
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.09.23 15:37:21 -04'00'		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																													
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1019</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5086																																																														
Interferent	MP5087																																																														
0.050	MP5088																																																														
0.080	MP5089																																																														
0.200	MP5090																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Form 47</u>																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
David Eliezer Reyes Rivera Digitally signed by David Eliezer Reyes Rivera Date: 2020.09.22 14:51:03 -04'00'		 2020.09.23 15:36:21 -04'00'																																																													
Tech Review / Date		Admin Review / Date																																																													

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO  
Time of Inspection: 14:18

Date of Inspection: 09/22/2020

Serial Number: 80-007382  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.080

Standard Deviations	0.0000	0.0004	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

09/22/2020  
Date

SR


2020.09.2  
3 15:35:23  
-04'00'

**DERR** Digitally signed  
by DERR  
Date: 2020.09.22  
14:50:35 -04'00'

# Stability Checks

SUWANNEE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007382  
 09/22/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.049	09:34
Air Blank	0.000	09:35
Control Test	0.049	09:35
Air Blank	0.000	09:36
Control Test	0.048	09:37
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

SUWANNEE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007382  
 09/22/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Control Test	0.080	09:40
Air Blank	0.000	09:40
Control Test	0.080	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature

SUWANNEE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007382  
 09/22/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.201	09:43
Air Blank	0.000	09:44
Control Test	0.201	09:44
Air Blank	0.000	09:45
Control Test	0.200	09:46
Air Blank	0.000	09:46
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	



Operator's Signature

SUWANNEE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-007382  
 09/22/2020  
 Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.082	09:53
Air Blank	0.000	09:54
Control Test	0.081	09:54
Air Blank	0.000	09:55
Control Test	0.082	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Dry



Operator's Signature

**DERR** Digitally signed  
 by DERR  
 Date: 2020.09.22  
 14:50:10 -04'00'

 2020.09.  
 23  
 15:34:23  
 -04'00'



**ANAB**  
ANSI National Accreditation Board  
ACCREDITED  
ISO/IEC 17025  
FORENSIC CALIBRATION  
LABORATORY

# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007382, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007382</u>	UNCERTAINTY* ±
Owning Agency:	<u>SUWANNEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/22/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>14:18</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/22/2020 \_\_\_\_\_  
Date Israel Soto,  
Department Inspector

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Digitally signed  
by DERR  
Date: 2020.09.22  
14:49:47 -04'00'

2020.09.22  
3 15:33:17  
-04'00'





# REQUEST FOR REGISTRATION

Florida Department of  
Law Enforcement

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007382

OWNING AGENCY: Suwannee County Sheriff

DATE OF DEPARTMENT INSPECTION: 09-22-2020

AGENCY INSPECTOR: David Falgout

ADDRESS: 305 SW Pine Ave

CITY, STATE, ZIP: Live Oak FL 32064

TELEPHONE NUMBER: 386-590-3127

FAX NUMBER: 386-364-3506

EMAIL ADDRESS (if available): david.falgout@suwanneesherriff.com

***For Program Office Use Only:***

Registration Issued *BK* 2020.09.24  
08:11:32 -04'00'

Instrument Added to Evidentiary Instrument Database

Instrument Added to Monthly Statistics Database

Contact Information Added to Instrument Database