

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Sumter County
 Time of Inspection:

Date of Inspection: 06/1/2021

Serial Number: 80-000223
 Software: 8100.27

| CHECK OR TEST | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| Diagnostic Check (Pre-Inspection): OK | | |
| Alcohol Free Subject Test: 0.000 | | |
| Mouth Alcohol Test: Slope Not Met | | |
| Interferent Detect Test: Interferent Detect | | |
| Diagnostic Check (Post-Inspection): OK | | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.20g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____ |
|----------------------------|---|---|---|--|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: _____

Remarks: Agency Inspection could not be conducted. Instrument would not warm up and come out of Not Ready Mode. Per Department Inspector Taylor Gutschow, the records have been uploaded into COBRA and the instrument will be returned to the Department

The above instrument complies (____) does not comply (____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. Terence L. Hollis 615

Signature and Printed Name

06/11/2021

Date