

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 00:47

Date of Inspection: 08/07/2021

Serial Number: 80-000224

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202101C Exp: 08/07/2021 <u>01/12/2023</u> | 0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021 | 0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023 |
|----------------------------|---|---|---|---|
| 0.000 | 0.046 | 0.077 | 0.193 | 0.080 |
| 0.000 | 0.049 | 0.080 | 0.200 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.080 |

Number of Simulators Used: 5

Remarks:

MA LOT # 2021-B, ACETONE LOT #202-A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JONATHAN R HURT

Signature and Printed Name

08/07/2021
Date