

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Suwannee Sheriff Office

Serial Number: 80.000780

Time of Inspection: N/A

Date of Inspection: N/A

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks: Inspection was not completed do to the fact I was out of town at the end of the month for a family illness

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

 Date