

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO  
Time of Inspection: 12:09

Date of Inspection: 01/30/2021

Serial Number: 80-000819  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:05620080A2 Exp: 04/05/2022
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.078

Number of Simulators Used: 3

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MARY E BOWERMASTER

Signature and Printed Name

01/30/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO

Serial Number: 80-000819

Time of Inspection: 13:53

Date of Inspection: 02/26/2021

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 05620080A2 Exp: 04/05/2022
0.000	0.048	0.080	0.200	0.077
0.000	0.048	0.080	0.200	0.077
0.000	0.048	0.079	0.200	0.077

Number of Simulators Used: 4

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MARY E BOWERMASTER

Signature and Printed Name

02/26/2021

Date



To: Whom it May Concern

From: Agency Inspector Mary Bowermaster

Date: 03/30/2021

Re: Instrument 80-000819

---

The above noted instrument was not in evidentiary use during March 2021 related to printer issues. The instrument will be sent back to the FDLE ATP for inspection.



To: Whom it May Concern

From: Agency Inspector Mary Bowermaster

Date: 04/30/2021

Re: Instrument 80-000819

---

The above noted instrument was not in evidentiary use during April 2021 related to printer issues. The instrument will be sent back to the FDLE ATP for inspection.



To: Whom it May Concern

From: Agency Inspector Mary Bowermaster

Date: 05/30/2021

Re: Instrument 80-000819

---

The above noted instrument was not in evidentiary use during May 2021 related to printer issues. The instrument has been sent back to the FDLE ATP for inspection.

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO  
Time of Inspection: 16:13

Date of Inspection: 07/31/2021

Serial Number: 80-000819  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:05620080A2 Exp: 04/06/2022
0.000	0.048	0.079	0.198	0.076
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.079	0.198	0.076

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

MARY E BOWERMASTER

07/31/2021  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO  
Time of Inspection: 00:42

Date of Inspection: 09/01/2021

Serial Number: 80-000819  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:05620080A2 Exp: 04/05/2022
0.000	0.050	0.078	0.199	0.077
0.000	0.050	0.078	0.199	0.077
0.000	0.049	0.079	0.200	0.077

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Mary E Bowermaster*  
\_\_\_\_\_  
Signature and Printed Name

Inspector began 8/31/2021 09/01/2021  
Date

Delayed completion related to electrical supply to room - (initials)

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CITRUS COUNTY SO  
Time of Inspection: 00:42

Date of Inspection: 09/01/2021

Serial Number: 80-000819  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:05620080A2 Exp: 04/05/2022
0.000	0.050	0.078	0.199	0.077
0.000	0.050	0.078	0.199	0.077
0.000	0.049	0.079	0.200	0.077

Number of Simulators Used: 4

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MARY E BOWERMASTER

Signature and Printed Name

09/01/2021  
Date