

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL GABLES PD
Time of Inspection: 14:18

Date of Inspection: 05/24/2021

Serial Number: 80-000871
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
0.000				
0.000				
0.000				

Number of Simulators Used: 4


Remarks:

Int Det, NO INTERFERANT DETECTED Non-compliance: NO INT. DETECTED.

SECOND TEST CONDUCTED BECAUSE OF BROKEN SIMULATOR BOTTLE PROHIBITING PROPER SEAL OF INTERFERENT TEST

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


#8433

 CRISTINO PEREZ
 Signature and Printed Name

05/24/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL GABLES PD
Time of Inspection: 15:21

Date of Inspection: 05/24/2021

Serial Number: 80-000871
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:989917 Exp: 07/25/2021
0.000	0.049	0.079	0.000 / 0.197	0.080
0.000	0.049	0.079	0.000 / 0.198	0.080
0.000	0.049	0.079	0.000 / 0.198	0.080

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance.

Improper Seal on 20 jar.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature] #8433

CRISTINO PEREZ

Signature and Printed Name

05/24/2021
Date