

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 12:58

Date of Inspection: 08/05/2021

Serial Number: 80-000906
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023 | 0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022 | 0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24119080A1 Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.044 / 0.048 | 0.076 | 0.196 | 0.082 |
| 0.000 | 0.047 / 0.049 | 0.078 | 0.199 | 0.081 |
| 0.000 | 0.048 / 0.050 | 0.079 | 0.201 | 0.081 |

Number of Simulators Used: 5


Remarks:

05: Control Outside Tolerance.

Unknown reason for failure of first 0.05 test. Retested and passed. Informed department inspector. by #3223

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



YVETTE R GONZALES

Signature and Printed Name

08/05/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 11:11

Date of Inspection: 10/20/2021

Serial Number: 80-000906
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023 | 0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022 | 0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24119080A1 Exp: 11/05/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.199 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.078 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.079 |

Number of Simulators Used: 5


Remarks:

~~0.05 / 0.08 DID NOT GET SLOPE NOT MET.~~

I did not receive the results of Slope Not Met on the first try of Mouth Alcohol test. After re-rinsing my mouth w/ the mouth alcohol solution, I did receive Slope Not Met. *Y #3223*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



YVETTE R. GONZALES

Signature and Printed Name

10/20/2021
Date