

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 10:19

Date of Inspection: 11/16/2021

Serial Number: 80-000907
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.042 / 0.041			
0.000	0.041 / 0.042			
0.000	0.042 / 0.043			

Number of Simulators Used: 5

Remarks:


Time-Date changed. 05: Control Outside Tolerance, Control OutOUTSIDE TOLERANCENon-compliance: OUTSIDE TOLERANCE.

05 Solution was tested on another instrument and the solution was found to be within acceptable ranges.

Discussed with FDLE and FDLE advised they will inspect the instrument further. Sent to FDLE (FT Myers)

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM D COLEMAN

Signature and Printed Name

11/16/2021
Date