

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 10:21


Date of Inspection: 07/28/2021

Serial Number: 80-000939
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) <i>201908B</i> Lot#:201910D Exp: 10/22/2021 <i>08/07/2021</i>	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31420080A2 Exp: 01/05/2023
0.000	0.048	0.078	0.195	0.077
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks: *wrong lot # & Expiration entered during AI, updated info ABOVE* 
7/26/2021

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROMAN SERRANO

Signature and Printed Name

07/28/2021
Date