

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 13:29

Date of Inspection: 04/28/2021

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1094981 Exp: 02/22/2022
0.000	0.055	0.085	0.211 / 0.209	0.078
0.000	0.055	0.085	0.210 / 0.208	0.078
0.000	0.055	0.085	0.209 / 0.208	0.078


Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. *FIRST INSPECTION SINCE SIMULATORS HAD THE YEARLY MAINTENANCE WHICH WAS DONE IN APRIL 2021. UNKNOWN WHY THE NUMBERS WERE READING HIGH.* (X)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #17979

 Signature and Printed Name
 KYLE DUFFEY

 04/28/2021
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 13:34

Date of Inspection: 10/27/2021

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1094981 Exp: 02/22/2022
0.000	0.040 / 0.050	0.078	0.200	0.079
0.000	0.041 / 0.049	0.078	0.201	0.079
0.000	0.041 / 0.050	0.079	0.201	0.079

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance UNKNOWN GSKT & JAR INTACT.
POSSIBLE AIR LEAK. TOOK APART, EXAMINED AND RE-ASSEMBLED. FOUND NO DEBRIS OR DAMAGED GLASS. FUNCTIONED SECOND TIME.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]
17979

KYLE DUFFEY

Signature and Printed Name

10/27/2021
Date