

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD

Serial Number: 80-001063

Time of Inspection: 11:15

Date of Inspection: 04/07/2021

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1031121 Exp: 03/03/2021
0.000	0.048	0.079	0.198	0.076
0.000	0.048	0.079	0.198	0.076
0.000	0.048	0.079	0.199	0.077

Number of Simulators Used: 5

Remarks:

BSO FRAZIER WAS PRESENT.

It was brought to my attention by Dept. Insp. David Reyes that the Dry Gas std. Test had an incorrect exp. date. The correct exp. date is 10/12/2021.

Rafael Deriet # 2736

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Rafael Deriet # 2736 RAFAEL DERIET
Signature and Printed Name

04/07/2021
Date