

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 09:36

Date of Inspection: 08/27/2021

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22620080A2 Exp: 10/05/2022
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:
MONTHLY INSPECTION


**RECEIVED
FDLE**

SEP 01 2021

Alcohol Testing Program

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICHARD L PATRONE

Signature and Printed Name

08/27/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 10:35

Date of Inspection: 12/27/2021

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22620080A2 Exp: 10/05/2022
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.081	0.200	0.080
0.000	0.050	0.080	0.201	0.080

Number of Simulators Used: 5


Remarks:

A F / M A: RFI Detect.

Note 0.08 Test Lot was 202007A exp. 7.21.2022. Entered 0.05 Lot information for 0.08 Test by accident.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 MATTHEW P JOHNSON
 Signature and Printed Name

12/27/2021
 Date